



**Rajasthan Medical Services Corporation Limited, Jaipur**  
**Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005**

Ph. No. 0141-2223887, Fax : 0141-2228065

E-mail : edepmrmisc-ri@nic.in

CIN:U24232RJ2011SGC035067

Website: http://rmisc.health.rajasthan.gov.in

No. F-8() RMSC/EPM/M-2/18-19/NIB-266/ 1163

Dated: 20.03.18

**Rate Contract and Price Details**

S.N.	Item	Detail
1	Name of Equipment / Item	<b>LABOUR TABLE</b>
2	Name & Address of Firm	M/s A K Surgicals, 25/11 Ist Floor, Kaveri Path, New Sanganer Road, Mansarovar, Jaipur-302020 E-mail: aksurgicals@yahoo.co.in
3	Brand/ Model / Make	AKS 3008
4	Packing Unit	Each
5	R/C No. & Validity	R.C. No. 430 (Valid Upto 31.03.2020)
6	<b>Rates for supply to RMSC from Suppliers</b>	
	(i) New Basic Rate (in Rs. )	22013.00
	(ii) IGST @ 18% (in Rs. )	3962.34
	(iii) CGST (in Rs. )	0.00
	(iv) SGST (in Rs. )	0.00
	(v) <b>Approved Rate From Supplier To RMSC {(i)+(ii)+(iii)+(iv) (in Rs. )</b>	<b>25975.34</b>
7	<b>Rate For Supply to Consignee From RMSC</b>	
	(i) New Basic Rate (in Rs.)	22013.00
	(ii) RMSC Surcharge@5% On (i) (in Rs.)	1100.65
	(iii) CGST@ 9 % { On (i)+(ii) } (in Rs.)	2080.23
	(iv) SGST@ 9 % { On (i)+(ii) } (in Rs.)	2080.23
	(v) <b>Approved Rate For Consignee From RMSC {(i)+(ii)+(iii)+(iv) (in Rs.)</b>	<b>27274.11</b>
8	<b><u>Technical Specifications of Labour Table</u></b>	
	<ol style="list-style-type: none"> <li>1. Smoothly finished complete stainless steel labour table (Stainless Steel grade 304)</li> <li>2. Should have three section top with rexine covered minimum 25mm thick sectional mattress. The density of mattress foam should not be less than 40.</li> </ol>	

*de*



3. Should have a dimension of L 180 X W 90 X H 80 cm (Tolerance  $\pm 5\%$ )
4. Should be able to hold minimum of 150 Kg.
5. Back Rest should be adjustable on ratchet.
6. Removable SS Bowls under the U cut, size proportionate to the U cut size provided with SS head end size riling made from solid SS rods, SS had grips.
7. Leg section should be telescoped under the mainframe completely for lithotomy positions.
8. Should have S.S. head & side riling steel hand grips, a pair of rubber padded stainless steel lithotomy rods and lithotomy crutches with 2 pairs of rexine ankle straps.
9. Should provide a detachable saline stand having a provision for fixing four positions (Head end and middle position) on the table 2 Nos. of saline rods of SS 304 grade should be provided.
10. Base should be mounted on four heavy duty PVC/ rubber shoes.
11. Guarantee: Three years from the date of installation.
12. Product should be CE Certified. (Valid documentation should be submitted in technical bid).
13. Certificate of stainless steel grade 304 should be submitted in technical bid, also at the time of demo issued from authorized metal testing laboratory.
14. Final technical approval after demonstration.
15. Company should have service network in Rajasthan.
16. Installation will be done by supplier free of cost.
17. The company should submit technical compliance sheet as per technical specifications mentioning the make & model of quoted item along with catalogue in the technical bid.

*sd*

Executive Director (EPM)  
RMSC, Jaipur

Dated: 20.03.18

No. F-8() RMSC/EPM/M-2/18-19/NIB-266/ 1163

Copy forwarded to the following for information & necessary action:

1. P.S. to Mission Director NHM Jaipur.
2. Director, (PH)/(Aids)/(FW)/(RCH)/(IEC)/(ESI).
3. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar & Udaipur.
4. Superintendent Jaipuriya Hospital Jaipur (Associated with RUHS College of Medical Scinces, Jaipur).
5. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
6. All P.M.O.'s.
7. All CM&HO's.

*ds*



8. Executive Director (Logistics/Finance)
9. AGM (IT), RMSC, Jaipur
10. All DPC's DDW Rajasthan.
11. BME (R&M), RMSC, Jaipur.
12. M/s A K Surgicals, 25/11 Ist Floor, Kaveri Path, New Sanganer Road, Mansarovar, Jaipur-302020.
13. Guard file.

  
20.3.2018  
Executive Director (EPM)  
RMSC, Jaipur

