

Rajasthan Medical Services Corporation Limited, Jaipur

'D'-block, swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur

Phone no. 0141-2223887, Fax no. 0141-2228065

E-mail: rmssc@nic.in

CIN : U24232RJ2011SGC035067

Website : www.rmssc.nic.in

No. F-8() RMSC/EPM/M-1/RC -2018-19/ERW/ NIB-261/ 2018/ 1523

Dated: 3-04-18

All Principal, Medical Colleges,
All the Superintendent of Medical College & Attached Hospitals.
Director (PH/HA/IEC/AIDS/RCH)

Sub:- Approval/Purchase of Item, 'Elisa Reader and Washer', for the Rate Contract and supply, Ending on 30.04.2020.

Please find enclosed herewith approved firm and rates etc. in respect of Item 'Elisa Reader and Washer', for the Rate Contract and supply, ending on 30.04.2020. The Rate Contract and its terms & conditions are also available on Rajasthan Medical Services Corporation Ltd., Website www.rmssc.nic.in.

In reference to above Rate Contract No. F-8() RMSC/EPM/M-1/RC -2018-19/ERW/ NIB-261/ 2018/ 1523 Dated 3.4.18..... the supplier has executed the agreement with RMSC, as enclose here for your information & necessary action. This bears an approval of M.D., RMSCL, Jaipur.

3-4-2018


Executive Director (EPM)
Rajasthan Medical Services Corporation
Ltd, Jaipur

No. F-8() RMSC/EPM/M-1/RC -2018-19/ERW/ NIB-261/ 2018/

Dated:

Copy forwarded to the following for information & necessary action:

1. P.S. to Hon'ble Health Minister, Medical & Health Deptt. Govt. of Rajasthan.
2. P.S. to PHS, Medical Health & FW Deptt. Rajasthan, Jaipur.
3. P.S. to Mission Director NHM Jaipur.
4. Director, (PH)/(Aids)/(FW)/(RCH)/(IEC)/(ESI).
5. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar & Udaipur.
6. Principal R.U.H.S. College of Dental Science Jaipur.
7. Principal, RUHS College of Medical Sciences, Jaipur.
8. Superintendent Japuriya Hospital Jaipur (Associated with RUHS College of Medical Sciences, Jaipur).
9. Superintendent, SMS Hospital, Zenana Hospital, Mahila Chikitsalya, Mental Hospital, J.K. Lon Hospital, T.B & Chest Hospital, Jaipur.
10. Superintendent Associated Group of Hospital-Ajmer, M.G. Hospital, Jodhpur, Ummaid Hospital-Jodhpur, PBM Hospital-Bikaner, MBS Hospital-Kota & A.G. Hospital-Udaipur.
11. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
12. Addl. Director, Medical & Health Services (PW)/ (RH)/(Hospital Admn.)
13. Joint Director, Medical & Health Services, (Zone) Jaipur, Ajmer, Udaipur, Jodhpur, Bikaner, Kota & Bharatpur.
14. All P.M.O.'s.
15. All CM&HO's.
16. All Dy. CM&HO's.



17. AGM (IT), RMSC, Jaipur.
18. M/s Meril Diagnostics Pvt. Ltd., Survey No. 135/2/B , Bilakhia House
Muktanand Marg, Chala, Gujrat .
19. Guard file.

dy
3-4-2018

Executive Director (EPM)
Rajasthan Medical Services Corporation
Ltd, Jaipur



Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065
CIN:U24232RJ2011SGC035067

E-mail : edepmmsc-ri@nic.in
Website: <http://rpsc.health.rajasthan.gov.in>

No. F-8 () RMSC/EP/M-1/RC-2018-19/ERW/NIB-261/2018/ 1824

Dated: 3-04-18

Rate Contract and Price Details

S.N.	Item	Detail	
1	Name of Equipment / Item	Elisa Reader and Washer	
2	Name & Address of Firm	M/s Meril Diagnostics Pvt. Ltd, Survey No. 135/2/B, Bilakhia House Muktanand Marg, Chala, Gujrat E-mail: prakesh.teckchandani@merillife.com	
3	Brand/ Model / Make	Meril/Merilyzer ELAQuant &	
4	Packing Unit	Each	
5	R/C No. & Validity	R.C. No. 443 (Valid Upto 30.04.2020)	
6		Rates for supply to RMSC from Suppliers	
	(i) New Basic Rate (in Rs.)		218000.00
	(ii) I GST @18%(in Rs.)		39240.00
	(iii) CGST (in Rs.)		0.00
	(iv) SGST (in Rs.)		0.00
	(v) Approved Rate From Supplier To RMSC {(i)+(ii)+(iii)+(iv)} (in Rs.)		257240.00
7		Rate For Supply to Consignee From RMSC	
	(i) New Basic Rate (in Rs.)		218000.00
	(ii) RMSC Surcharge@5% On (i) (in Rs.)		10900.00
	(iii) CGST@ 9 % { On (i)+(ii)} (in Rs.)		20601.00
	(iv) SGST@ 9 % { On (i)+(ii)} (in Rs.)		20601.00
	(v) Approved Rate For Consignee From RMSC {(i)+(ii)+(iii)+(iv)} (in Rs.)		270102.00
8		Technical Specifications of Elisa Reader and Washer	
	1. Filter based ELISA reader with halogen lamp as light source and silicon photo detectors.		

km



2. Should have a linear measurement range of 0 to 4 Abs or more
3. Should be compatible to wavelength range from 400 to 700 nm.
4. Should have a photometric accuracy of $\pm 2\%$ or better
5. Should have a resolution of 0.001Abs
6. Should have variable speed plate shaking capability
7. Should have easy access 8 position filter wheel
8. Machine should be supplied with 4 standard filters of User choice
9. Should have automatic filter selection option
10. Should have automatic calibration before each reading
11. Should be able to read 96 well format in 6 seconds
12. Capable of doing multi standard tests and controls
13. Should be capable of reading U, V and flat type wells.
14. The instrument should run in stand- alone mode OR with computer & software controlled.
15. System should have USB port for external printer connectivity option for data printout.
16. Software CD should have unlimited user licence to analyze the data in different user/ computer systems.
17. Instrument should be either CE mark, IVD/ FDA or Europe regulations
18. Demonstration of equipment is must for technical approval.

Data Analysis Software :

- Analysis software should be supplied with the instrument and has unlimited user system license. Software should have inbuilt calculation, such as Blank Subtraction, Quantitative Curve Fit, Quantitative Classification , Spectral Analysis and Kinetic Calculation , as well as a comprehensive reporting tool.
- Both measured and calculated data can also easily be exported to other systems like Microsoft Excel for further data handling
- User should be able to load the software in to multiple computer systems for data analysis
- Software should have option for area selection i. e. different protocols at different area of the same plate
- Database based software to run backups of all data, restore backup data (in case of hardware failure of original computer)
- Supplier should quote for the following accessories : UPS, Computer with Printer
- Firm to give 3 years guarantee and rates for 5 years CMC.

Technical specification of item Washer

- Semi- automatic Micro plate ELISA washer with co-axial wash heads for excellence washing & can be switched between 8 to 12 way wash heads.
- Should be supplied with at least 4 standard program cards (1-4 washing cycles) & one programmable card for user defined program
- Dispensing precision should be at least 5 %

