



Rajasthan Medical Services Corporation Limited, Jaipur

'D'-block, swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur

Phone no. 0141-2223887, Fax no. 0141-2228065

CIN : U24232RJ2011SGC035067

E-mail: rmsc@nic.in

Website : www.rmsc.nic.in

No. F-8() RMSC/EPM/M-1/ RC 1-19/NIB-298/ Six body Mort./ 2018/ 1855 Dated: 5-04-18

All Principal, Medical Colleges,
All the Superintendent of Medical College & Attached Hospitals.
Director (PH/HA/IEC/AIDS/RCH)

Sub:- Approval/Purchase of Item, 'Six Body Mortuary Cabinet', for the Rate Contract and supply, Ending on 30.04.2020.

Please find enclosed herewith approved firm and rates etc. in respect of Item 'Six Body Mortuary Cabinet', for the Rate Contract and supply, ending on 30.04.2020. The Rate Contract and its terms & conditions are also available on Rajasthan Medical Services Corporation Ltd., Website www.rmsc.nic.in.

In reference to above Rate Contract No. F-8() RMSC/EPM/M-1/ RC 1-19/NIB-298/ Six body Mort./ 2018/ 1855 Dated 5.4.18 the supplier has executed the agreement with RMSC, as enclose here for your information & necessary action. This bears an approval of M.D., RMSCL, Jaipur.

Executive Director (EPM)
Rajasthan Medical Services Corporation
Ltd, Jaipur

No. F-8() RMSC/EPM/M-1/ RC 1-19/NIB-298/ Six body Mort./ 2018/

Dated:

Copy forwarded to the following for information & necessary action:

1. P.S. to Hon'ble Health Minister, Medical & Health Deptt. Govt. of Rajasthan.
2. P.S. to PHS, Medical Health & FW Deptt. Rajasthan, Jaipur.
3. P.S. to Mission Director NHM Jaipur.
4. Director, (PH)/(Aids)/(FW)/(RCH)/(IEC)/(ESI).
5. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar & Udaipur.
6. Principal R.U.H.S. College of Dental Science Jaipur.
7. Principal, RUHS College of Medical Sciences, Jaipur.
8. Superintendent Japuriya Hospital Jaipur (Associated with RUHS College of Medical Sciences, Jaipur).
9. Superintendent, SMS Hospital, Zenana Hospital, Mahila Chikitsalya, Mental Hospital, J.K. Lon Hospital, T.B & Chest Hospital, Jaipur.
10. Superintendent Associated Group of Hospital-Ajmer, M.G. Hospital, Jodhpur, Ummaid Hospital-Jodhpur, PBM Hospital-Bikaner, MBS Hospital-Kota & A.G. Hospital-Udaipur.
11. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
12. Addl. Director, Medical & Health Services (PW)/ (RH)/(Hospital Admn.)
13. Joint Director, Medical & Health Services, (Zone) Jaipur, Ajmer, Udaipur, Jodhpur, Bikaner, Kota & Bharatpur.
14. All P.M.O.'s.
15. All CM&HO's.
16. All Dy. CM&HO's.

Executive Director (EPM)
Rajasthan Medical Services Corporation
Ltd, Jaipur

[Handwritten Signature]
4.4.2018

- 17. AGM (IT), RMSC, Jaipur.
- 18. M/s Authentic Instruments 31, RIICO industrial area, Jhotwara Jaipur.
- 19. Guard file.





Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax :0141-2228065

CIN:U24232RJ2011SGC035067

E-mail : edep@rmsc-ri@nic.in

Website: <http://rmsc.health.rajasthan.gov.in>

No. F-80) RMSC/EPM/M-I/RC 1-19/NIB-298/ Six body Mort/ 2018/ 1856

Dated: 5-04-18

Rate Contract and Price Details

S.N.	Item	Detail	
1	Name of Equipment / Item	Six Body Mortuary Cabinet	
2	Name & Address of Firm	M/s Authentic Instruments 31, RICO industrial area, Jhotwara Jaipur 302018 Email : support@authenticinstruments.com Authentic/MCR-6	
3	Brand/ Model / Make	Each	
4	Packing Unit		
5	R/C No. & Validity	R.C. No. 446 (Valid Upto 30.04.2020)	
6		Rates for supply to RMSC from Suppliers	
	(i) New Basic Rate (in Rs.)		395000.00
	(ii) IGST (in Rs.)		0.00
	(iii) CGST @ 9% (in Rs.)		35550.00
	(iv) SGST @ 9% (in Rs.)		35550.00
	(v) Approved Rate From Supplier To RMSC {(i)+(ii)+(iii)+(iv)} (in Rs.)		466100.00
7		Rate For Supply to Consignee From RMSC	
	(i) New Basic Rate (in Rs.)		395000.00
	(ii) RMSC Surcharge@5% On (i) (in Rs.)		19750.00
	(iii) CGST @ 9 % { On (i)+(ii) } (in Rs.)		37328.00
	(iv) SGST @ 9 % { On (i)+(ii) } (in Rs.)		37328.00
	(v) Approved Rate For Consignee From RMSC {(i)+(ii)+(iii)+(iv)} (in Rs.)		489406.00
8		1. Technical Specifications of Six Body Mortuary Cabinet	
	1. Mortuary cabinet should be single compartment and have 6 cadaver storage facilities and should have hinged door with locking		



	arrangement in 2 cadavers / column x 3 column. (Minimum outer dimension L or D x W x H as 7.5 feet x 9 feet x 5 feet \pm 2cm)
2.	Inner surface & Outer surface made of Stainless steel-304 panel with insulation, in between inner & outer panels, made of PUF (CFC Free) Density 40 kg/m ³ .
3.	Front door with stainless steel SS 304 Grade with provision for numbering & tag on each door.
4.	Temperature should be maintained between 2°C to 5°C
5.	Refrigeration of entire compartment should be controlled and Temperature can be adjustable, LED/ LCD Display. It should have double Cooling unit with individual Temperature controller cum indicator unit, if main unit fails refrigeration control shift to stand by unit. Automatic time cycling between main unit and standby unit has to be provided by the bidder. Compressor should be ISI Mark (certificate should be supplied with each mortuary cabinet at the time of supply.)
6.	Should have stabilizer backup to control voltage fluctuation.
7.	Capacity: 10000 (BTU/H) per unit with standby.
8.	Power Supply: 230V/50 Hz supply
9.	Foamed Panels Insulation thickness of panel not less than 60mm.
10.	One piece stainless steel tray for keeping cadaver.
11.	The carriage consists of three piece assembly which consists of stationary frame a lower Carriage and an upper carriage. The lower and upper carriage assembly is riding on wheels and tracks that allow easy telescopic action. The complete assembly is locked when returned to closed position.
12.	Defrosting Automatic type.
13.	Should be hermetically sealed with forced Air Circulation.
14.	Should have Alarm system & Internal Lighting.
15.	Installation will be done by supplier free of cost including all parts like wires, tubes/ pipe, joints & attachments, small fixtures etc. Suitable 220 V,50 Hz electrical supply will be provided by the consignee at the site.
16.	Firm should mention all the pre-installation requirements in technical bid.
17.	Manufacturer should be ISO-9001 certified for quality standards. (Certificate to be submitted in technical Bid).
18.	Equipment model should be CE certified. (Certificate to be submitted in technical Bid).
19.	Guarantee: Three years on equipment, accessories and all peripherals / sub systems from the date of installation, which includes 4 visits per year and unlimited breakdown calls by service/application support engineer during the Guarantee period.
20.	CMC: CMC will be given @5% (of net rate-inclusive of Excise Duty & exclusive of VAT/CST etc.) plus service tax (as applicable). The CMC may be awarded for three years (on yearly basis) after completion of guarantee period of three years.
21.	Final Approval after Technical Demonstration.
22.	The Supply Order of this item "Six Body Mortuary Cabinet" under this purchase order is subjected to pre-dispatch inspection by RMSCL team at dispatch of each lot.



Executive Director (EPM)
RMSC, Jaipur
Dated:

[Signature]
4.4.2018

No. F-8) RMSC/EPM/M-1/ RC 1-19/NIB-298/ Six body Mort/ 2018/

Copy forwarded to the following for information & necessary action:

1. P.S. to Mission Director NHM Jaipur.
2. Director, (PH)/(Aids)/(FW)/(RCH)/(IEC)/(ESD).
3. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar & Udaipur.
4. Superintendent Jaipuria Hospital Jaipur (Associated with RUHS College of Medical Sciences, Jaipur).
5. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
6. All P.M.O.'s.
7. All CM&HO's.
8. Executive Director (Logistics/Finance)
9. AGM (TT), RMSC, Jaipur
10. All DPC's DDW Rajasthan.
11. BME (R&M), RMSC, Jaipur,
12. M/s Authentic Instruments 31, RICO industrial area, Jhotwara Jaipur
13. Guard file.

[Signature]
4.4.2018
Executive Director (EPM)
RMSC, Jaipur

