

Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tikak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065

E-mail : edep@rmsc-n@nic.in
 Website: www.health.rajasthan.gov.in

CIN:U24232RJ2011SCC035067

No. F-8(40) RMSC/HP/M-4/18-19/NIB-225/ 1569

Dated: 5.04.18

Order for Rate Contract


S. No.	Item	Detail
1.	Name of Equipment / Item	Saxlet Extraction Unit
2.	Name & Address of Firm	M/S Sudarshan Chemicals, Sudarshan Building, Jhalawar Road, Kota-324007 sudarshanchemicals@gmail.com
3.	Brand/ Model / Make	ALPINA
4.	Packing Unit	Each
5.	R/C No. & Validity	(R.C. No 449 /Valid upto 30.04.2020)
6.	Net Rate Per packing Unit (Rs.)	30920.00 (Rs. Thirty Thousand Nine Hundred Twenty Only) HSN Code 8421
8.	RMSC Surcharge @5%	Rs. 1546.00
9.	Rates for Supply to RMSC from Suppliers	
	(i) New Basic Rate (In Rs.)	30920.00
	(ii) I GST @ 0 % (In Rs.)	0.00
	(iii) CGST @ 9 % (In Rs.)	2782.80
	(iv) SGST @ 9 % (In Rs.)	2782.80
	Approved Rate for Suppliers From RMSC {(i)+(ii)+(iii)+(iv)} (In Rs.)	36486.00

Rate for Supply To Consignee From RMSC

(i)	New Basic Rate (In Rs.)	30920.00
(ii)	RMSC Surcharge @ 5% on (In Rs.)	1546.00
(iii)	CGST @9% {On (i)+(ii)} (In Rs.)	2921.94
(iv)	SGST @ 9% {On (i)+(ii)} (In Rs.)	2921.94
(v)	Approved Rate for Consignee From RMSC {(i)+(ii)+(iii)+(iv)} (In Rs.)	38310.00

11. Technical Specification of Soxhlet Extraction Unit

1. The complete Soxhlet Extraction Apparatus consists of solvent reservoir extraction assembly with 6 extraction heaters to suit 250 ml flasks.
 2. Outer casing is made of S.S.- 304 Steel.
 3. Gap between casing & heater is filled with high grade glass wool insulation.
 4. Extraction Heaters made of Fiber glass hand knitted netting having Nichrome heating elements.
 5. Heating Elements having individual Regulators & Pilot Lamps to indicate working of the heaters.
 6. Control box is away but junctioned with the heating console.
 7. Complete with Two retort rods, Cross bar, Clamp holder & Boss heads.
 8. Operates on 230 Volts AC Single phase 50 Hz.
 9. Should be supplied with heating element and 250ml extraction units of borocil, suitable connection pipe 10 Meter.
 10. Equipment should be CE Certified.
 11. Guarantee: Two years on equipment from the date of installation on all Parts of equipment including metal fitting / tubes / Glass parts etc. against manufacturing defects
- CMC: CMC shall be given @ 5 % of net rate excluding GST (as applicable) and Yearly escalation of 5% on last year CMC Price. The CMC may be awarded for five years (on Yearly Basis) after completion of Guarantee Period of two years.



 Executive Director (EPM)
 Rajasthan Medical Services Corporation Ltd.,
 Jaipur

F-8(40) RMSC/EPM/M-4/16-17/NIB-225/

Copy forwarded to the following for information & necessary action:

1. P.S. to Mission Director NHM Jaipur.
2. Director, (PHD)/(Aids)/(FW)/(RCH)/(IEC)/(ESI).
3. Principal Medical College, Aimer, Bikaner; Jaipur, Todipur Kotla, Jhalawar & Udaipur.
4. Superintendent Japuriya Hospital Jaipur (Associated with RUHS College of Medical Sciences, Jaipur).
5. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj, Jaipur.
6. All P.M.O.s.
7. All CM&HO's.
8. Executive Director (Logistics/Finance).
9. AGM (TD), RMSC, Jaipur.
10. All DPCs DDW Rajasthan.
11. M/s Sudarshan Chemicals, Sudarshan Building, Jhalawar Road, Kota-324007 sudarshanchemicals@gmail.com
12. Guard file.

Dated:


Executive Director (EPM)
Rajasthan Medical Services Corporation Ltd,
Jaipur

पुस्तकालय विभाग