



Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065

CIN:U24232RJ2011SGC035067

No. F-80) RMSC/EPM/M-2/18-19/NIB-263/1630

E-mail : edepmrm-sc-ri@nic.in
Website: <http://rmsc.health.rajasthan.gov.in>

Dated: 10-04-18

Rate Contract and Price Details

S.N.	Item	Detail	
1	Name of Equipment / Item	<i>Cardiac Monitor with Defibrillator</i>	
2	Name & Address of Firm	M/s Surbhi Meditek, 138, First Floor, Metropolis Tower, Nirmohi Nagar, Near Purani Chungi, Ajmer Road, Jaipur-302019 E-Mail: surbhi.meditek@gmail.com MINDRAY	
3	Brand/ Model / Make	MINDRAY	
4	Packing Unit	Each	
5	R/C No. & Validity	R.C. No. 451 (Valid Upto 30.04.2020)	
6		Rates for supply to RMSC from Suppliers	
	(i) New Basic Rate (in Rs.)		1,62,000.00
	(ii) IGST @ 12% (in Rs.)		19,440.00
	(iii) CGST (in Rs.)		
	(iv) SGST (in Rs.)		
	(v) Approved Rate From Supplier To RMSC {(i)+(ii)+(iii)+(iv)} (in Rs.)		1,81,440.00
7		Rate For Supply to Consignee From RMSC	
	(i) New Basic Rate (in Rs.)		1,62,000.00
	(ii) RMSC Surcharge@5% On (i) (in Rs.)		8100.00
	(iii) CGST@ 6 % { On (i)+(ii)} (in Rs.)		10206.00
	(iv) SGST@ 6 % { On (i)+(ii)} (in Rs.)		10206.00
	(v) Approved Rate For Consignee From RMSC {(i)+(ii)+(iii)+(iv)} (in Rs.)		190512.00



Technical Specifications of *Cardiac Monitor with Defibrillator*

1. Defibrillator should be Bi-Phasic
2. Defibrillator should be ready for AED facility and non invasive pacing.
3. Should have a high resolution colour TFT display of minimum 6 inch or more.
4. Should have direct trim knob and direct function keys for mute and freeze.
5. Should have energy levels for defibrillation up to 200 joules or more
6. Should be mains and battery operated. Internal battery should provide backup operation up to 2 -3 hours in monitoring or at least 90 defibrillation shocks should be delivered from fully charged battery.
7. Integrated external re-usable adult and paediatric paddles for defibrillation.
8. Should have non synchronised and synchronised cardio version.
9. Facilities of ECG pickup from paddles in case of ECG electrodes are not connected to the Defibrillator.
10. The charging time should be less than 7 seconds for charging up to 200 joules.
11. Should have integrated Printer.
12. Should be European CE/USFDA approved.
13. **Guarantee:** Three years on equipment from the date of installation and 06 months on accessories.
14. **CMC:** CMC shall be given @ 5% of net rate (excluding GST as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after Guarantee period of three years.
15. Equipment should be supplied with compatible 3/5 lead ECG cable (2 Nos.) & Operating / user Manual (1 No.)
16. **Rates of Consumable:** Rates of consumable like Battery, 3/5 lead ECG cable, Disposable AED pad, thermal paper pack should be provided by bidder in BOQ in e-proc. Portal.
17. Installation will be done by supplier free of cost.
18. Firm should have service engineer network in Rajasthan.
19. Demonstration of equipment is must for technical approval.
20. The firm should submit technical compliance sheet along with catalogue as per technical specifications mentioning the make & model of quoted item in the Technical bid.

Executive Director (EPM)
RMSC, Jaipur

No. F-8() RMSC/EPM/M-2/18-19/NIB-263/ 1630

Dated: 10-04-18

Copy forwarded to the following for information & necessary action:

1. P.S. to Mission Director NHM Jaipur.
2. Director, (PH)/(Aids)/(FW)/(RCH)/(EC)/(ESD).
3. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar & Udaipur.
4. Superintendent Jaipuriya Hospital Jaipur (Associated with RUHS College of Medical Sciences, Jaipur).
5. Director, Medical & Health Services (ESD), Mobile Surgical Unit, Raj. Jaipur.
6. All P.M.O.'s.
7. All CM&HO's.
8. Executive Director (Logistics/Finance)
9. AGM (IT), RMSC, Jaipur
10. All DPC's DDW Rajasthan.
11. BME (R&M), RMSC, Jaipur.
12. M/s Surbhi Meditek, 138, First Floor, Metropolis Tower, Nirmohi Nagar, Near Purani Chungi, Ajmer Road, Jaipur-302019
13. Guard file.

Executive Director (EPM)
RMSC, Jaipur

