



Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065

E-mail : edepmrmisc-rl@nic.in

CIN:U24232RJ2011SGC035067

Website: www.rmssc.health.rajasthan.gov.in

No. F-8(313)RMSC/EPM/M-3/2017-18/NIB-273/1907 Dated:- 25-04-18

Rate Contract

S.N.	Item	Detail
1.	Name of Equipment / Item	Deep Freezer-Horizontal
2.	Name & Address of Firm	M/s Authentic Instruments, 31, RIICI Indl. Area, Jhotwara, Kamani Choraha, Jaipur-302012 Email:- support@authenticinstruments.com , calibration@authenticinstruments.com
3.	Brand/ Model / Make	Authentic
4.	HSN Code	
5.	Packing Unit	Each
6.	R/C No. & Validity	(R.C. No-463) valid up to 30.4.2020
7.	Net Rate Per packing Unit (Rs.)	105000.00
8.	Taxes	GST As applicable
9.	RMSC Surcharges @ 5%	As applicable
10.	Rates for Supply to RMSC from Suppliers	
	I. New Basic Rate (In Rs.)	105000.00
	II. IGST (In Rs.) 0 %	0.00
	III. CGST @ % 14 (In Rs.)	14700.00
	IV. SGST @ % 14 (In Rs.)	14700.00
	V. Approved Rate For Supplier To RMSC {(i)+(ii)+(iii)+(iv) (In Rs.)	134400.00
11.	Rate For Supply to Consignee From RMSC	
	I. New Basic Rate (In Rs.)	105000.00
	II. RMSC Surcharge @ 5 % On (i) (In Rs.)	5250.00
	III. CGST @ 14% { On (i)+(ii) } (In Rs.)	15435.00
	IV. SGST @ 14% { On (i)+(ii) } (In Rs.)	15435.00
	V. Approved Rate For Consignee from RMSC {(i)+(ii)+(iii)+(iv) (In Rs.)	141120.00

**Technical Specifications of Deep Freezer (Horizontal)**

1. Storage capacity 200 litres & Provides temperature of -20°C with accuracy +1°C
2. Outer chamber is made of Mild Steel duly pre treated & finished with powder coated paint for lasting finish.
3. Inner chamber made of STAINLESS STEEL (304 Grade) and should be 100% leak Proof.
4. Outer front double walled door is insulated and is fitted with magnetic tape for air tight closing for no temperature loss, provided with lock and key arrangement.
5. Door operated illumination lamp is fitted inside the chamber for easy visibility.
6. A full view inner plexi-glass door enables inspection and monitoring of inner chamber milk bottles without disturbing the process temperature.
7. High grade PUF INSULATION between outer and inner chamber for minimal thermal losses to maintain -20°C for at least 1 hr in case of power-cut.
8. Direct cooling system for stable temperature control by ISI marked Compressor/Condensing Units CFC free & R-134 a eco friendly refrigerant.
9. Control panel having all controls and circuitry are housed at the top fitted with Temperature controller cum Indicator Microprocessor based with LED display for Set Value (SV) & Process Value (PV) to select any desired temperature.
10. Caster wheel mounted for easy portability.
11. Metal mesh (Jali) for storage of milk containers 30ml ± 5ml.
12. Suitable for operation on 220 to 240 volt, single phase, 50 Hz, A.C. power supply
13. Refrigerant used be non toxic, non hazardous, CFC/HCFC free.
14. Should be provided with Voltage stabilizer or Inbuilt Stabilizer.
15. Should be provided with paper/digital chart recorder.

Terms & Conditions:

1. Manufacturer should have ISO 13485 Certificate.
2. Equipment should be ISI/CE certified.
3. **Guarantee:** Two years on equipment from the date of installation.
4. **CMC:** CMC will be given at 5% plus GST (as applicable) and yearly escalation of 5% on last year's CMC price. The CMC may be awarded for three years (on yearly basis) after completion of Guarantee period of two years.
5. Installation & Training will be done by supplier free of cost.
6. The service engineer should be based in Rajasthan.
7. The company should mention the make & model name / number of the quoted equipment and submit the technical brochure of the quoted model in the Technical bid along with compliance sheet as per technical specifications.

[Signature]
23.4.2018
Executive Director (EPM)
RMSC, Jaipur

No. F-8(342)RMSC EPM/M-3/2017-18/NIB-258.

Dated:

Copy forwarded to the following for information & necessary action:

1. P.S. to Mission Director NHM Jaipur.
2. Director, (PH)/(Aids)/(FW)/(RCH)/(IEC)/(ESI).
3. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar & Udaipur.
4. Superintendent Jaipuriya Hospital Jaipur (Associated with RUHS College of Medical Sciences, Jaipur).
5. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
6. All P.M.O.'s.
7. All CM&HO's.
8. Executive Director (Logistics/Finance).
9. AGM (IT), RMSC, Jaipur.
10. All DPC's DDW Rajasthan.
11. M/s Authentic Instruments, 31, RIICI Indl. Area, Jhotwara, Kamani Choraha, Jaipur-302012.
Email:-support@authenticinstruments.com, calibration@authenticinstruments.com
12. Guard file.

[Signature]
23.4.2018
Executive Director (EPM)
RMSC, Jaipur