



Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065

CIN:U74232RJ2011SGC035067

E-mail : edepmmrc-ri@nic.in

Website: <http://rmsc.health.rajasthan.gov.in>

No. F-8() RMSC/EPM/M-2/18-19/NIB-262/ 1827

Dated: 23-04-18

Rate Contract and Price Details

S.N.	Item	Detail
1	Name of Equipment / Item	Multi Para Monitor
2	Name & Address of Firm	M/s Schiller Healthcare India Pvt. Ltd., Advance House, Makwana Road, Off. Andheri Kurla Road, Marol Naka Station, Andheri (E), Mumbai- 400 059 E-Mail:-mail@schillerindia.com
3	Brand/ Model / Make	Truscope Ultra Q7
4	Packing Unit	Each
5	R/C No. & Validity	R.C. No. 466 (Valid Upto 30.04.2020)
6		Rates for supply to RMSC from Suppliers
	(i) New Basic Rate (in Rs.)	1,65,000.00
	(ii) IGST @ 12% (in Rs.)	19800.00
	(iii) CGST (in Rs.)	
	(iv) SGST (in Rs.)	
	(v) Approved Rate From Supplier To RMSC {(i)+(ii)+(iii)+(iv)} (in Rs.)	184800.00
7		Rate For Supply to Consignee From RMSC
	(i) New Basic Rate (in Rs.)	165000.00
	(ii) RMSC Surcharge@5% On (i) (in Rs.)	8250.00
	(iii) CGST@ 6 % { On (i)+(ii)} (in Rs.)	10395.00
	(iv) SGST@ 6 % { On (i)+(ii)} (in Rs.)	10395.00
	(v) Approved Rate For Consignee From RMSC {(i)+(ii)+(iii)+(iv)} (in Rs.)	194040.00

ds



Technical Specifications of Multi Para Monitor

1. Should have ECG, SpO₂, NIBP, Respiration & Temperature as standard parameters.
2. Should have facility to display 12 lead ECG, RR, HR, SpO₂, NIBP, and Temperature.
3. Monitor should have option for modular upgradable for CO (Cardiac Output), EtCO₂ (Capnography) & 2 IBP.
4. Should display at least 10 waveforms of selected parameters simultaneously.
5. Should have inbuilt battery backup at least one hour.
6. **Display:** Color TFT touch screen display size of 15" or more.
7. Should have dual temperature monitoring either in °Celsius or °Fahrenheit.
8. Should have facility for displaying multi - screen configurations.
9. Should be able to store & display at least 24 hours of graphical trends of all parameters.
10. Should be suitable for monitoring adult & pediatric & neonate patients.
11. The SpO₂ technology should also sense in hypotension, shivering & motion like Masimo.
12. Should have oscillometric Technology for measurement of NIBP with Auto, STAT and Manual modes.
13. Should have different patient type selection.
14. The pulse rate should be displayed either with ECG or SpO₂.
15. The respiration rate should be calculated through Impedance method.
16. Should be able to analyze arrhythmias & ST segment changes.
17. Memory should not wipe off when the power is turned off.
18. Should be able to give visual & audible alarms with three levels of volume adjustment.
19. Should have connectivity to Central station through Ethernet card or Wireless Connectivity.
20. Should be European CE/USFDA approved.
21. **Guarantee:** Three years on equipment from the date of installation and 06 months on accessories.
22. **CMC:** CMC shall be given @ 5% of net rate (excluding GST as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after Guarantee period of three years.
23. Equipment should be supplied with compatible 5 lead ECG cable (1 Nos.), 10 lead ECG cable (1 Nos.), SpO₂ (Adult) probe with extension cable (2 Nos.), NIBP Cuff with extension cable of Large Adult, Adult & Pediatric (2 each), Temperature probe (2 Nos.) & Operating / user Manual (1 No.)
24. **Rates of Accessories and Consumable:** Rates of accessories and consumable like Battery, 5 lead ECG cable, 10 lead ECG cable, SpO₂ (Adult) probe, SpO₂ extension cable, NIBP extension cable, NIBP Cuff (Large Adult), NIBP Cuff (Adult), NIBP Cuff (Pediatric) & Temperature probe should be provided by bidder in BOQ in e-proc. Portal.
25. Cost of optional/ upgradeable parameters should be mentioned separately in BOQ, so that if required at later stage, we can upgrade the same. However this cost will not be added for the price consideration of the monitor.

de



26. Installation will be done by supplier free of cost.
27. Firm should have service engineer network in Rajasthan.
28. Demonstration of equipment is must for technical approval.
29. The firm should submit technical compliance sheet along with catalogue as per technical specifications mentioning the make & model of quoted item in the Technical bid.


No. F-8) RMSC/EPM/M-2/18-19/NIB-262/

Copy forwarded to the following for information & necessary action:

1. P.S. to Mission Director NHM Jaipur.
2. Director, (PH)/(Aids)/(FW)/(RCH)/(IEC)/(ESD), Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar & Udaipur.
3. Superintendent Jaipuria Hospital Jaipur (Associated with RUHS College of Medical Sciences, Jaipur).
4. Director, Medical & Health Services (ESD), Mobile Surgical Unit, Raj. Jaipur.
6. All P.M.O.'s.
7. All CM&HO's.
8. Executive Director (Logistics/Finance)
9. AGM (IT), RMSC, Jaipur
10. All DPC's DDW Rajasthan.
11. BME (R&M), RMSC, Jaipur.
12. M/s Schiller Healthcare India Pvt. Ltd., Advance House, Makwana Road, Off. Andheri Kuria Road, Marol Naka Station, Andheri (E), Mumbai- 400 059
13. Guard file.


Executive Director (EPM)
RMSC, Jaipur

Dated:


Executive Director (EPM)
RMSC, Jaipur

