



Rajasthan Medical Services Corporation Limited, Jaipur
Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065
CIN:U24232RJ2011SGC035067

E-mail : edepmmsc-rj@nic.in
Website: <http://rmsc.health.rajasthan.gov.in>

No. F-8() RMSC/EPM/M-2/18-19/NIB-364/ 4128

Dated: 1-11-18

Rate Contract and Price Details

S.N.	Item	Detail
1	Name of Equipment / Item	BOD Incubator:-Digital (S.S. of 304)
2	Name & Address of Firm	M/s Remi Elektrotechnik Limited, Remi House, 11 Cama Indl. Estate, Goregaon (E), Mumbai-400 063, India E-Mail:- bids@remilabworld.com
3	Brand/ Model / Make	Brand: -- CI 6S+VS02 REMI
4	Packing Unit	Each
5	R/C No. & Validity	R.C. No. 525 (Valid Upto 31.10.2020)
6	Rates for supply to RMSC from Suppliers	
	(i) New Basic Rate (in Rs.)	95000.00
	(ii) IGST @ % (in Rs.)	11400.00
	(iii) CGST@ 6% { On (i)+(ii) } (in Rs.)	0.00
	(iv) SGST@ 6% { On (i)+(ii) } (in Rs.)	0.00
	(v) Approved Rate From Supplier To RMSC {(i)+(ii)+(iii)+(iv)} (in Rs.)	106400.00
7	Rate For Supply to Consignee From RMSC	
	(i) New Basic Rate (in Rs.)	95000.00
	(ii) RMSC Surcharge@5% On (i) (in Rs.)	4750.00
	(iii) CGST@ 6% { On (i)+(ii) } (in Rs.)	5985.00
	(iv) SGST@ 6% { On (i)+(ii) } (in Rs.)	5985.00
	(v) Approved Rate For Consignee From RMSC {(i)+(ii)+(iii)+(iv)} (in Rs.)	111720.00

(Signature)



8

Technical specifications of BOD Incubator:-Digital (S.S of 304)

1. Double walled double door Construction. Temp range 0° to 80°C with accuracy $\pm 0.1^\circ\text{C}$ high quality, environment friendly refrigerant, with voltage stabilizer of sufficient capacity .
2. Independent tem. Measuring through PT-100 sensor with indicator LCD display.
3. Inner Acrylic Door: Full Size 8 mm thick, Interior lighting facility, PID Control : Microprocessor based PID Controller.
4. Capacity of inner chamber approximately 150 \pm 5 litre
5. Recovery time short, precise regulation of temp and acoustic alarm.
6. Digital safety thermostat.
7. Adjustable ventilation rate 10 to 100% thin form air circulation.
8. The Unit Shall be capable of Operating Continuously in ambient temp of 10-45°C
9. temp. Temp. should be equally distributed in whole incubator, Synthetic rubber gasket at the door.
9. Whole system should be **CE certified**.

Terms & Conditions:

1. **Guarantee:** Three years on equipment from the date of installation.
2. **CMC:** CMC shall be given @ 5% of net rate (excluding GST as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after Guarantee period of three years.
3. Installation & Training will be done by supplier free of cost.
4. The service engineer should be based in Rajasthan.
5. The company should mention the make & model name / number of the quoted equipment and submit the technical brochure of the quoted model in the Technical bid along with compliance sheet as per technical specifications.

Executive Director (EPM)
RMSC, Jaipur



No. F-8() RMSC/EPM/M-2/18-19/NIB-364/ 4128

Dated: 1-11-18

Copy forwarded to the following for information & necessary action:

1. P.S. to Mission Director NHM Jaipur.
2. Director, (PH)/(Aids)/(FW)/(RCH)/(IEC)/(ESI).
3. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar & Udaipur.
4. Superintendent Jaipuriya Hospital Jaipur (Associated with RUHS College of Medical Sciences, Jaipur).
5. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
6. All P.M.O.'s.
7. All CM&HO's.
8. Executive Director (Logistics/Finance)
9. AGM (IT), RMSC, Jaipur
10. All DPC's DDW Rajasthan.
11. BME (R&M), RMSC, Jaipur.
12. M/s Remi Elektrotechnik Limited, Remi House, 11 Cama Indl. Estate, Goregaon (E), Mumbai-400 063, India
13. Guard File.

13.11.2018

Executive Director (EPM)
RMSC, Jaipur