

Rajasthan Medical Services Corporation Limited, Jaipur

'D'-block, swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur

Phone no. 0141-2223887, Fax no. 0141-2228065

CIN : U24232RJ2011SGC035067

E-mail: rmisc@nic.in

Website : www.rmisc.nic.in

No. F--8() RMSC/EPM/M-1/ RC 18-19/-40deg/ NIB-359/2018/4436

Dated: 6.12.18

All Principal, Medical Colleges,
All the Superintendent of Medical College & Attached Hospitals.
Director (PH/HA/IEC/AIDS/RCH)

Sub:- Approval/Purchase of Item, '(-) 40 degree C Deep freezer (Cap. 300 Plasma Bags)', for the Rate Contract and supply, Ending on 30.11.2020.

Please find enclosed herewith approved firm and rates etc. in respect of Item '(-) 40 degree C Deep freezer (Cap. 300 Plasma Bags)', for the Rate Contract and supply, ending on 30.11.2020. The Rate Contract and its terms & conditions are also available on Rajasthan Medical Services Corporation Ltd., Website www.rmisc.nic.in.

In reference to above Rate Contract No. F-8() RMSC/EPM/M-1/ RC 18-19/-40deg / NIB-359/2018/Dated the supplier has executed the agreement with RMSC, as enclose here for your information & necessary action. This bears an approval of M.D., RMSCL, Jaipur.



Executive Director (EPM)
Rajasthan Medical Services Corporation
Ltd, Jaipur

No. F--8() RMSC/EPM/M-1/ RC 18-19/-40deg / NIB-359/2018/

Dated:

Copy forwarded to the following for information & necessary action:

1. P.S. to Hon'ble Health Minister, Medical & Health Deptt. Govt. of Rajasthan.
2. P.S. to PHS, Medical Health & FW Deptt. Rajasthan, Jaipur.
3. P.S. to Mission Director NHM Jaipur.
4. Director, (PH)/(Aids)/(FW)/(RCH)/(IEC)/(ESI).
5. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar & Udaipur.
6. Principal R.U.H.S. College of Dental Science Jaipur.
7. Principal, RUHS College of Medical Scinces, Jaipur.
8. Superintendent Japuriya Hospital Jaipur (Associated with RUHS College of Medical Scinces, Jaipur).
9. Superintendent, SMS Hospital, Zenana Hospital, Mahila Chikitsalya, Mental Hospital, J.K. Lon Hospital, T.B & Chest Hospital, Jaipur.
10. Superintendent Associated Group of Hospital-Ajmer, M.G. Hospital, Jodhpur, Ummaid Hospital-Jodhpur, PBM Hospital-Bikaner, MBS Hospital-Kota & A.G. Hospital-Udaipur.
11. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
12. Addl. Director, Medical & Health Services (PW)/(RH)/(Hospital Admn.)
13. Joint Director, Medical & Health Services, (Zone) Jaipur, Ajmer, Udaipur, Jodhpur, Bikaner, Kota & Bharatpur.
14. All P.M.O.'s.
15. All CM&HO's.
16. All Dy. CM&HO's.



17. AGM (IT), RMSC, Jaipur.
18. M/s Remi Elektrotechnik Ltd., Remi House, Mumbai
19. Guard file.

[Handwritten signature]
6-12-2018

**Executive Director (EPM)
Rajasthan Medical Services Corporation
Ltd, Jaipur**



Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065
CIN:U24232RJ2011SGC035067

E-mail : edepmrmssc-ri@nic.in
Website: <http://rmssc.health.rajasthan.gov.in>

No. F-8() RMSC/EPM/M-1/ RC 18-19/-40 deg/ NIB-359/2018/ 4437

Dated: 6-12-18

Rate Contract and Price Details

S.N.	Item	Detail
1	Name of Equipment / Item	(-) 40 degree C Deep freezer (Cap. 300 Plasma Bags)
2	Name & Address of Firm	M/s Remi Elektrotechnik Ltd., Remi House, 11 CAMA INDL Estate, Goregaon (E), Mumbai-400063 Ph. 91-22-4058 9888 Fax: 91-22-4058 9890 E-mail : sales@remilabworld.com Remi RPF 224 ultra with VS-03
3	Brand/ Model / Make	
4	Packing Unit	Each
5	R/C No. & Validity	R.C. No. 542 (Valid Upto 30.11.2020)
6		Rates for supply to RMSC from Suppliers
	(i) New Basic Rate (in Rs.)	259000.00
	(ii) IGST @ 18% (in Rs.)	46620.00
	(iii) CGST (in Rs.)	0.00
	(iv) SGST (in Rs.)	0.00
	(v) Approved Rate From Supplier To RMSC {(i)+(ii)+(iii)+(iv)} (in Rs.)	305620.00
7		Rate For Supply to Consignee From RMSC
	(i) New Basic Rate (in Rs.)	259000.00
	(ii) RMSC Surcharge@5% On (i) (in Rs.)	12950.00
	(iii) CGST@ 9 % { On (i)+(ii)} (in Rs.)	24476.00
	(iv) SGST@ 9 % { On (i)+(ii)} (in Rs.)	24476.00
	(v) Approved Rate For Consignee From RMSC {(i)+(ii)+(iii)+(iv)} (in Rs.)	320902.00

Handwritten signature and date: 6-12-2018

Technical Specifications of (-) 40 degree C Deep freezer (Cap. 300 Plasma Bags)

Specifications of item

1. Purpose of Equipment
 - a) To Freeze or store Plasma.
 - b) Must be designed specifically for blood bank use. Commercial or modified commercial freezers for other purpose are not acceptable.
2. Type of Equipment
 - a) Microprocessor controlled Temperature controlling. Digital LCD /LED display at eye Level.
 - b) With audio alarm for temperature deviation, power failure, door open etc.
 - c) Should have 24 hour battery backup for uninterrupted 24X7 electronic temperature circular chart recorder /Inkless graphic Temperature recorder with range of 0°C to -50°C with data logger with supply of free charts for a period of guaranty and controller unit keeps the temperature monitoring uninterrupted.
 - d) Approved standard electrical Blood Bank plasma freezer that uses a compressor circulating CFC-free refrigerant / Green gas.
 - e) Upright type
3. Quality Standard
 - a) Manufacturing should be compliant with ISO 13485 (Valid documentation should be submitted in technical bid).
 - b) Should be compliant with European CE (Notified body) or US FDA (Valid documentation should be submitted in technical bid).
 - c) Equipment must meet electrical safety specifications of IEC 60601-1
4. Capacity
 - a) At least 300 standard plasma bags
5. Construction
 - a) Outside C. R. (Corrosion Resistant) Sheet at least 1 mm thick
 - b) Inside stainless steel of at least 22 G.
 - c) Insulation polyurethane foam >80mm thick, foaming agent CFC free

Handwritten signature and date: 6-12-2018





- d) Should be mounted on lockable caster wheels
- 6. Drawers
 - a) At least four or more in number made up of SS 304 grade stainless steel
- 7. Door
 - a) Separate inner doors to prevent cold loss
 - b) Automatic/Magnetic/Latch closing of at least inner doors.
 - c) Heating device/ vacuum release valve in front to avoid condensation
 - d) Opening angle limited (eg <135°)
 - e) Door open/ajar audio and visual alarm
 - f) Door lock should be available
- 8. Electrical characteristics
 - a) Compatible with Input 240V 50 Hz Single phase AC
 - b) Should have an integrated voltage stabilizer or external stabilizer of appropriate ratings meeting ISI Specifications (Input 160-260 V and output 220-240 V and 50 Hz).
 - c) Minimum compressor starting voltage should be 22% below normal voltage
- 9. Internal Temperature
 - a) Should be able to maintain internal temperature not warmer than -30°C Audible / visual over temperature Alarm.
 - b) Whatever the load, setting accuracy less than or equal to 1°C
 - c) For the purpose of defrosting (either Automatic or manual system), temperature should not go outside safe range
- 10. External Ambient Temperature:
 - a) Can perfectly maintain internal temperature as above at full load in an ambient temperature of +10 to at least +40 °C
- 11. Hold-Over Time
 - a) 2 Hrs at ambient temp.
- 12. Cooling down time
 - A full load of plasma packs at +25 °C takes a maximum of 5 hr for all packs to reach below -5°C.



- A full load of plasma packs at +25 °C takes a maximum 30 hrs for all packs to reach below - 20 °C
 - 13. Additional requirements:
 - Complete with comprehensive set of spare parts and a suitable capacity stabilizer and one -1.5 Ton split A.C of reputed make (can be o-general/mistubishi/daikain) should be supplied with the system free of cost. (Guarantee card with bill have to be submit with item to claim guarantee)
 - 14. SMS alerts to user defined recipients in case of any deviation in temperature, break down and power failure with provision to connect with central (Temperature) monitoring system / facility of wireless data transfer facility & storage of temp. data on cloud, with continuous accessibility of data on PC or laptop and on smart phones with internet facility
 - 15. Noise factor should not exceed 60 decibel.
 - 16. Guarantee: Three years on equipment from the date of installation.
 - 17. CMC: CMC will be given @ 5 % (of net rate -Rate exclusive of GST (as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after completion of Guarantee period of three years
 - 18. The company should submit technical compliance sheet as per technical specifications mentioning the make & model of quoted item along with catalogue in the Technical bid.
 - 19. Installation will be done by supplier free of cost.
 - 20. The service engineer should be based in Rajasthan.
 - 21. Four Preventive Maintenance services annually during Guarantee and CMC period are essential.
 - 22. Demonstration of equipment is must for final technical approval
- Comprehensive Maintenance Contract Charges (per year/per machine)

Rates (Rs.) Per Unit	First Year	Second Year	Third Year	Fourth Year	Fifth Year
Comprehensive Maintenance Agreement (CMA)	CMC: CMC will be given @ 5 % (of net rate -Rate exclusive of GST (as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after completion of Guarantee period of three years				



[Handwritten Signature]
6-12-2018

Executive Director (EPM)
RMSC, Jaipur

Dated:

No.F-8() RMSC/EPMM-1/ RC 18-19/-40deg/ NIB-359/2018/

Copy forwarded to the following for information & necessary action:

1. P.S. to Mission Director NHM Jaipur.
2. Director, (PH)/(Aids)/(FW)/(RCH)/(TEC)/(ESI).
3. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar & Udaipur.
4. Superintendent Jaipuriya Hospital Jaipur (Associated with RUHS College of Medical Sciences, Jaipur).
5. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
6. All P.M.O.'s.
7. All CM&HO's.
8. Executive Director (Logistics/Finance)
9. AGM (IT), RMSC, Jaipur
10. All DPC's DDW Rajasthan.
11. BME (R&M), RMSC, Jaipur.
12. M/s Remi Elektrotechnik Ltd., Remi House, Mumbai
13. Guard file.

[Handwritten Signature]
6-12-2018

Executive Director (EPM)
RMSC, Jaipur