755

Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax: 0141-2228065

CIN:U24232RJ2011SGC035067

E-mail: edepmrmsc-ri@nic.in
Website: http://rmsc.health.rajasthan.gov.in

No.:F-8() RMSC/EPM/M-3/NIB-450/2020/ 4389

Dated: 07/07/2020

Rate Contract and Price Details

S.N.	Item	Detail			
1	Name of Equipment / Item	Pulse Oxymeter			
2	Name & Address of Firm	M/s India Medtronics Pvt. Ltd. 1241 Solitaire Corporate Park, Bldg no. 12,4th Floor Andheri- Ghatkopal Link Road Andheri (East),Mumbai Email:-ajay.k.vashista@medtronic.com, rs.indiatender@medtronic.com			
3	Indicative Qty	500			
4	Status of Bidder	Importer			
5	Performance Security	Received Bank Guarantee No. LG/6216/GT015751/20 issue dated 01-07- 2020 for Rs. 13,72,000.00 through Bank of America			
6	Brand/ Make	Medtronic			
	Model	Nellcor bed side Spo2 Monitor			
7	Packing Unit	Each			
8	HSN Code	90181990			
9	R/C No. & Validity	R.C. No. 601 (Valid Upto 30.06.2022)			
10	Rates for supply to RMSC from Suppliers				
	(i) New Basic Rate (in Rs.)	49000.00			
	(ii) IGST @ 12% (in Rs.)	5880.00			
	(iii) CGST (in Rs.)	0.00			
	(iv) SGST (in Rs.)	0.00			
	(v) Approved Rate From Supplier To RMSC {(i)+(ii)+(iii)+(iv) (in Rs.)	54880.00			
11	Rate For Supply to Consignee From RMSC				
	(i) Basic Rate (in Rs.)	49000.00			
	(ii) RMSC Surcharge@5% On (i) (in Rs.)	2450.00			
	(iii) CGST@ 6% { On (i)+(ii)} (in Rs.)	3087.00			
	(iv) SGST@ 6% { On (i)+(ii)} (in Rs.)	3087.00			
	(v) Approved Rate for Consignee from RMSC {(i)+(ii)+(iii)+(iv) (in Rs.)	57624.00			



12


Rate For Consumables items

S. No	Name of approved item(s) with full specification		Make	Brand/ Model	Packing Unit	Approved Rate Per Unit (Rs.)	
1.		SPO2 Probe (Adult)	Medtronic	DS100A	Each	3700 Plus 12 %	GST
2.	Consumab	SPO2 Probe (Pediatrics)	Medtronic	D-YS	Each	3700 Plus 12 %	GST
3.	les details	SPO2 Probe (neonatal)	Medtronic	D-YS	Each	3700 Plus 12 %	GST
4.		Battery	Medtronic	10005948 5HR Battery	Each	7500 Plus 12 %	GST

13 Technical Specification of item Pulse Oxymeter

- 1. Suitable for all types of patients- adult, pediatric, neonatal.
- 2. Compact Table Top Pulse Oxymeter with colour monitor at least 3.5" TFT Screen.
- 3. SpO2 measurement range at least 40-70 and 70 to 99%, minimum gradation 1%.
- 4. Accuracy of SpO2 better than $\pm 1\%$ fro range 40-70 and better than $\pm 3\%$ for range 70-99.
- 5. Pulse rate range at least 30 to 240bpm, minimum gradation 1 bpm.
- 6. Accuracy of pulse rate better than ± 5 bpm.
- 7. Audiovisual alarms required: high and low SpO2 and pulse rate (operator variable settings), sensor disconnected, sensor failure, low battery.
- 8. Plethysmograph (may be in form of bar) display is mandatory.
- 9. Should have minimum 24 hrs trend memory for SpO2 & PR.
- 10. Display shows SpO₂(%), HR (bpm) and signal strength bar.
- 11. Spo2 technology should also sense in hypotention, shivering and motion.
- 12. Large display readable from distance, display covers of durable plastic.
- 13. User preset of high/low alarms on SpO2 and pulse rate monitoring.
- 14. Silencing feature for audio alarm.
- 15. Carrying handle should be provide for easy mobility.
- 16. Display reports system errors, probe failure, probe connection and built-in battery status.
- 17. Automatic switch from mains to batteries in case of power failure.
- 18. Power requirements: 220 V/50 Hz and internal re-chargeable battery (at least 4 hrs back up ,automatic recharge)
- 19. Manufacturer should have ISO 13485 Certificate.
- 20. Quoted model should be European CE(Notified Body)/USFDA approved.
- 21. **Guarantee:** Three years on equipment from the date of installation.
- 22. CMC: CMC will be given @ 5 % (of net rate exclusive of GST) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after completion of Guarantee period of three years.





- a. Reusable SpO₂ sensors for Adult including connection cable
- b. Reusable SpO₂ sensors for Pediatrics, Clip on type including connection cable
- c. Reusable SpO2 sensors for neonate, Wrap around type
- d. Forehead sensor
- 24. Installation will be done by supplier free of cost.
- 25. The service engineer should be based in Rajasthan.
- 26. Demonstration of equipment is must for technical approval.
- 27. The company should mention the make & model name / number of the quoted equipment and submit the technical brochure of the quoted model in the Technical bid along with compliance sheet as per technical specifications.

Other terms and condition to be covered during guarantee period by the supplier:-

- 1. Response time- < 48 Hours after logged the complain in complain portal.
- 2. Service hours- As per health facility schedule.
- 3. Preventive Maintenance Schedule (PM)-As per OEM (Mention PM Schedule)
- 4. Calibration Schedule As per OEM (Mention Schedule)
- 5. Up time- 95% (346 Days)
- 6. Technical & Application Support, Demonstrations & Trainings As & when required
- 7. Toll Free No. (24*7)-
- 8. Life of the equipment- As per OEM
- 9. Service engineer detail in Rajasthan-
- The agreement signed between RMSCL and the firm forms integral part of the rate contract.

Executive Director (EPM)
RMSCL, Jaipur

Dated: 07/07/2020

No.:F-8() RMSC/EPM/M-3/NIB-450/2020/ 4389

Copy forwarded to the following for information & necessary action:

1. P.S. to Mission Director NHM Jaipur.

2. Director, (PH)/(Aids)/(FW)/(RCH)/(IEC)/(ESI).

3. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar & Udaipur.

4. Superintendent Jaipuriya Hospital Jaipur (Associated with RUHS College of Medical Scinces, Jaipur).

5. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.

- 6. All P.M.O.'s.
- 7. All CM&HO's.
- 8. Executive Director (Logistics/Finance)
- 9. AGM (IT), RMSC, Jaipur
- 10. All DPC's DDW Rajasthan.
- 11. BME (R&M), RMSC, Jaipur.
- 12. M/s India Medtronics Pvt. Ltd , Mumbai
- 13. Guard file.

Executive Director (EPM)
RMSCL, Jaipur