

# Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065

CIN:U24232RJ2011SGC035067

E-mail : [edepmrmisc-rj@nic.in](mailto:edepmrmisc-rj@nic.in)

Website: [www.health.rajasthan.gov.in](http://www.health.rajasthan.gov.in)

No. F-8(133) RMSC/EPM/M-4/18-19/NIB-390/

Dated:

All Principal, Medical Colleges,  
All the Superintendent of Medical College & Attached Hospitals.  
Director (PH/HA/IEC/AIDS/RCH)

## Sub:- The Rate Contract for the Maternity Cot.

Please find enclosed herewith approved firm and rates etc. in respect of Item **Maternity Cot** for the rate contract ending on 31-03-2021. The Rate Contract and its terms & conditions are also available on Rajasthan Medical Services Corporation Ltd., Website [www.rmssc.nic.in](http://www.rmssc.nic.in).

In reference to above Rate Contract No. F-8(133) RMSC/EPM/M-4/18-19/NIB-390/413 Dated: 11.2.19 the supplier has executed the agreement with RMSC, as enclose here for your information & necessary action.

Executive Director (EPM)

Rajasthan Medical Services Corporation Ltd, Jaipur

No. F-8(133) RMSC/EPM/M-4/18-19/NIB-390/680

Dated: 6.3.19

Copy forwarded to the following for information & necessary action:

1. PS to Hon'ble Health Minister, Medical & Health Deptt. Govt. of Rajasthan.
2. PS to PHS, Medical Health & FW Deptt. Rajasthan, Jaipur.
3. PA to Director, (PH)/(Aids)/(FW)/(RCH)/(IEC).
4. Project Director, NRHM.
5. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar & Udaipur.
6. Superintendent, SMS Hospital, Zenana Hospital, Mahila Chikitsalya, Mental Hospital, J.K. Lon Hospital, T.B & Chest Hospital, Jaipur.
7. Superintendent Associated Group of Hospital-Ajmer, M.G. Hospital, Jodhpur, Ummaid Hospital- Jodhpur, PBM Hospital-Bikaner, MBS Hospital-Kota & A.G. Hospital-Udaipur.
8. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
9. Addl. Director, Medical & Health Services (PW)/(RH)/(Hospital Admn.)
10. Joint Director, Medical & Health Services, (Zone) Jaipur, Ajmer, Udaipur, Jodhpur, Bikaner, Kota & Bharatpur.
11. All Dy. CM&HO's.
12. AGM (IT), RMSC, Jaipur.
13. M/s United Surgical Industries 445, F.I.E. Patparganj Industrial Area Delhi Email: [info@unisurg.com](mailto:info@unisurg.com)
14. Guard file.

Executive Director (EPM)

Rajasthan Medical Services Corporation Ltd, Jaipur

**Rajasthan Medical Services Corporation Limited, Jaipur**  
Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065  
CIN:U24232RJ2011SGC035067

E-mail : [edepmrmsc-rj@nic.in](mailto:edepmrmsc-rj@nic.in)  
Website: [www.health.rajasthan.gov.in](http://www.health.rajasthan.gov.in)

No. F-8(133) RMSC/EPM/M-4/18-19/NIB-390/

Dated:

**Order for Rate Contract**

S. No.	Item	Detail
1.	Name of Equipment / Item	Maternity Cot
2.	Name & Address of Firm	M/s United Surgical Industries 445, F.I.E. Patparganj Industrial Area Delhi Email: <a href="mailto:info@unisurg.com">info@unisurg.com</a>
3.	Brand/ Model / Make	United USI-E7
4.	Packing Unit	Each
5.	R/C No. & Validity	(R.C. No 545 Valid upto 31.03.2021)
6.	Net Rate Per packing Unit (Rs.)	72700.00 (Seventy Two Thousand Seven Hundred Only) <b>HSN Code 9402</b>
8.	RMSC Surcharge @5%	Rs. 3635
9.	<b>Rates for Supply to RMSC from Suppliers</b>	
	(i) Basic Rate (In Rs.)	72700.00
	(ii) IGST@ 18 % (In Rs.)	13086
	(iii) CGST @ 0 % (In Rs.)	0.00
	(iv) SGST @ 0 % (In Rs.)	0.00
	(v) Approved Rate for Suppliers From RMSC {(i)+(ii)+(iii)+(iv)} (In Rs.)	85786.00
10.	<b>Rate for Supply To Consignee From RMSC</b>	
	(i) Basic Rate (In Rs.)	72700.00
	(ii) RMSC Surcharge @ 5% on (In Rs.)	3635.00
	(iii) CGST @9% {On (i)+(ii)} (In Rs.)	6870.15
	(iv) SGST @ 9% {On (i)+(ii)} (In Rs.)	6870.15
	(v) Approved Rate for Consignee From RMSC {(i)+(ii)+(iii)+(iv)} (In Rs.)	90075.3

Specifications of Maternity Cot (Birthing Bed)

1. Bed portion Size: 2155 mm (L) X 1010 mm (W) X 880 mm (H) without mattress. (all numerical parameters will have a tolerance of  $\pm 10\%$ ).
2. Three section high pressure laminated top. It shall have detachable laminated top for middle section and leg section for ease of cleaning.
3. Leg section can be telescoped under backrest for lithotomy position.
4. Backrest, trendelenburg and height adjustment position operated by Electro mechanical adjustment thru hand set; battery backup with inbuilt battery charger shall be provided. The hand set shall have indications for power on and the battery charge.
5. All electro mechanical actuators need to be compatible with class of IP 54.
6. The bed shall have work simultaneously on two actuators system for raising and lowering height. Bed frame made mainly from 60 x 30mm x 1.6mm mm thick ERW tube with proper support. This frame shall be fitted on the base frame mainly made of 60 x 30 x 1.6mm ERW tubes on various supporting links.
7. The base frame shall be mounted on 125mm dia. non-rusting castor wheels with central and directional locking mechanism and pedal operated at the head end of the bed. Castor made from high grade non floor-staining synthetic materials with integrated thread guards. Wheel centre having precision ball bearing to run smoothly.
8. The bed shall have easily detachable moulded head & foot side panels. Four corner rubber buffers.
9. Bed shall have swing down type railings; Railing shall be made from non-rusting moulded material. These shall be fitted to the mattress support sections and should be able to raise and lock through spring lock mechanism through an operating lever.
10. S.S. saline rod with 12mm dia S.S. rod shall telescope in SS socket tube approx 15.8 mm dia x 18G welded on angular base bracket of 14G SS sheet. Nylon bracket provided to prevent color damage it could be placed at two different locations.
11. Patient's bearing down handgrips should be made of 22mm dia 18 G stainless steel tube fixed with PVC handle and should have location adjustability for patient's convenience.
12. A pair of upholstery aluminum lithotomic crutches mounted on S.S. rods.
13. Waste collecting removable stainless steel tray shall be provided at perennial recess.
14. It shall have foot support for high / low chair position.
15. The backrest and middle section shall have 75 mm thickness foam mattress P.U. 40 D. with rexine cover and zip. The leg section shall have 130mm thickness for propose alignment with other adjacent sections. The PU foam shall have high resilience and radiolucent property. All MS parts should have powder coated. SS parts finished with Matt Polish.
17. This product should be US FDA/European CE/BIS approved.
18. Manufacturer should have ISO 13485 certificate for quality standards.
19. Electrical Safety confirms to the standard for electric safety IEC 60601-1.
20. Guarantee: Three Years on equipment with all accessories from the date of installation.
21. CMC: CMC will be given @ 5 % plus GST (as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after guarantee period of three years.
22. Installation will be done by supplier free of cost. Supplier to perform installation, safety and operation check before handover. Local clinical staff to affirm completion of installation.
23. Firm should mention all the pre-installation requirements in technical bid.
24. Training of users in operation and basic maintenance shall be provided. Advance maintenance tasks required shall be documented.
25. Service and Operation manuals to be provided.
26. Final technical approval after demonstration.
27. Pre dispatch inspection will be done on each order.
28. The company should mention the make & model name/ number of the quoted equipment and submit the technical brochure of the quoted model in the Technical bid along with compliance sheet as per technical specifications.


Executive Director (EPM)

Rajasthan Medical Services Corporation Ltd, Jaipur

No. F-8(133) RMSC/EPM/M-4/18-19/NIB-390/

Copy forwarded to the following for information & necessary action:

1. P.S. to Mission Director NHM Jaipur.
2. Director, (PHD)/(Aids)/(FW)/(RCH)/(IEC)/(EST).
3. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar & Udaipur.
4. Superintendent Japuniya Hospital Jaipur (Associated with RUHS College of Medical Sciences, Jaipur).
5. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
6. All P.M.O.'s.
7. All CM&HO's.
8. Executive Director (Logistics/Finance)
9. AGM (IT), RMSC, Jaipur
10. All DPC's DDW Rajasthan.
11. **M/s United Surgical Industries 445, F.I.E. Patparganj Industrial Area Delhi Email: info@unisurg.com**
12. Guard file.

  
5-3-2019

Executive Director (EPM)  
Rajasthan Medical Services Corporation Ltd.,  
Jaipur