



Rajasthan Medical Services Corporation Limited, Jaipur

'D'-block, swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur

Phone no. 0141-2223887, Fax no. 0141-2228065

E-mail: edepmrmsc-rj@nic.in

CIN : U24232RJ2011SGC035067

Website: <http://rmsc.health.rajasthan.gov.in>

No. F-8() RMSC/EPM/M-1/NIB-760/RC-821/2023-24/ 698

Dated: 11/11/23

All Principals Medical Colleges,
All the Superintendents of Medical Colleges & Attached Hospitals.
Director (PH/HA/IEC/AIDS/RCH)

Sub:- Approval/Purchase of Item, 'Dialysis Machine and R.O. System on Turnkey Basis', for the Rate Contract and supply, ending on 31.10.2025.

Please find enclosed herewith approved firm and 'Dialysis Machine and R.O. System on Turnkey Basis', for the rate contract and supply ending on 31.10.2025. The rate contract and its terms & conditions are also available on Rajasthan Medical Services Corporation Ltd., Website www.rmsc.health.rajasthan.gov.in

In reference to the above rate contract No. F-8() RMSC/EPM/M-1/NIB-760/RC-821/2023-24/Dated..... the supplier has executed the agreement with RMSC, as enclosed herewith for your information & necessary action. This bears an approval of M.D., RMSCL, Jaipur.

Executive Director (EPM)

RMSCL, Jaipur

Dated: 11/11/23

No. F-8() RMSC/EPM/M-1/NIB-760/RC-821/2023-24/ 698

Copy forwarded to the following for information & necessary action:

1. P.S. to Hon'ble Health Minister, Medical & Health Deptt. Govt. of Rajasthan.
2. P.S. to ACS, Medical Health & FW Deptt. Rajasthan, Jaipur.
3. PS to Pr. Secy/Secy, Medical Education Department, Rajasthan, Jaipur
4. P.S. to Mission Director NHM Jaipur.
5. PA to Managing Director, RMSCL
6. Director, (PH)/(Aids)/(FW)/(RCH)/(IEC)/(ESI).
7. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar & Udaipur.
8. Principal R.U.H.S. College of Dental Science Jaipur.
9. Principal, RUHS College of Medical Sciences, Jaipur.
10. Superintendent Japuriya Hospital Jaipur (Associated with RUHS College of Medical Sciences, Jaipur).
11. Superintendent, SMS Hospital, Zenana Hospital, MahilaChikitsalya, Mental Hospital, J.K. Lon Hospital, T.B & Chest Hospital, Jaipur.
12. Superintendent Associated Group of Hospital-Ajmer, M.G. Hospital, Jodhpur, Ummaid Hospital- Jodhpur, PBM Hospital- Bikaner, MBS Hospital-Kota & A.G. Hospital-Udaipur.
13. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
14. Addl. Director, Medical & Health Services (PW)/(RH)/(Hospital Admn.)
15. Joint Director, Medical & Health Services, (Zone) Jaipur, Ajmer, Udaipur, Jodhpur, Bikaner, Kota & Bharatpur.
16. All P.M.O.'s.
17. All CM&HO's.
18. All Dy. CM&HO's.
19. AGM (IT), RMSC, Jaipur
20. All BME, RMSC
21. M/s Aprameya Engineering Limited, Ahmedabaad.
22. Guard file.

Executive Director (EPM)

RMSCL, Jaipur



Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, SwasthyaBhawan, TilakMarg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065
CIN:U24232RJ2011SGC035067

E-mail : edepmrmisc-rj@nic.in
Website: <http://rmisc.health.rajasthan.gov.in>

No. F-8() RMSC/EPM/M-1/NIB-760/RC-821/2023-24/ 698

Dated: 11/11/23

Rate Contract and Price Details

| S.N. | Item | Detail |
|------|--|---|
| 1. | Name of Equipment / Item | Dialysis Machine and R.O. System on Turnkey Basis |
| 2. | Name & Address of Firm | M/s Aprameya Engineering Limited No. 908, 9Th Floor, Venus Atlantis Corporate Park, Prahlanagar Road, Nr. Petrol Pump, Ahmedabad-380015, Gujarat, India Email:- aprameyaengg@gmail.com GSTIN:24AAWCA3138N1Z6 |
| 3. | Bidder Status | Authorized Distributor |
| 4. | Performance Security | Rs. 19287600.00 vide BG No. 01701ILG004823 dated 16.10.2023 valid up to 31.03.2029 (Punjab National Bank) |
| 5. | Brand/ Make | NIPRO (Dialysis) and Amvin Aqua (R.O. System) |
| | Model | Surdial 55 Plus (Dialysis) and 250 LPH reverse Osmosis Plant (R.O. System) |
| 6. | Packing Unit | Each |
| 7. | R/C No. & Validity | R.C. No. 821 (Valid Up to 31.10.2025) |
| 8. | Indicative Quantity | Qty-182 Units (364 Dialysis Machine for 182 units & 182 R.O system for 182 units) |
| 9. | Rates per Unit (2 Dialysis & 1 R.O.) for supply to RMSC from Suppliers | |
| | (i) New Basic Rate (in Rs.) | 1796200.00 |
| | (ii) IGST @18% (in Rs.) | 323316.00 |
| | (iii) CGST (in Rs.) | 0.00 |
| | (iv) SGST (in Rs.) | 0.00 |
| | (v) Approved Rate From Supplier To RMSC {(i)+(ii)+(iii)+(iv)} (in Rs.) | 2119516.00 |
| 10. | Rates per Unit (2 Dialysis & 1 R.O.) For Supply to Consignee From RMSC | |
| | (i) Basic Rate (in Rs.) | 1796200.00 |
| | (ii) RMSC Surcharge@5% On (i) (in Rs.) | 89810.00 |
| | (iii) CGST@ 9% { On (i)+(ii)} (in Rs.) | 169740.90 |
| | (iv) SGST@ 9% { On (i)+(ii)} (in Rs.) | 169740.90 |
| | (v) Approved Rate for Consignee from RMSC {(i)+(ii)+(iii)+(iv)} (in Rs.) | 2225491.80 |
| 11. | Technical Specification for Dialysis Machine:- <ol style="list-style-type: none"> Blood pumps to achieve a wide range of flow adaptable to standard A-V blood lines. Heparin pump with adaptability of various sizes of syringes up to 20ml with pump flow rate from 0 to 9.9 ml min and bolus infusion up to 5ml/min. It should have sequential ultra filtration facility. It should have arterial, venous & trans-membrane pressure monitoring facility. | |



4. Functional air bubble detector.
5. It should have dialy sate flow rate range from 300 to 800ml/min
6. Variable conductivity setting between 12 to 16ms/cm
7. Functional blood leak detector.
8. Dialysate temperature regulator that has a range of temperature 35°C to 39°C
9. Safety device: Functioning Alarms, Venous Blood Clamp.
10. It should have control & monitoring screen with minimum 10" coloured TFT/LCD Display (Screen). Display should have screen with high clarity & easily readable.
11. It should have hot rinsing and hot chemical disinfection facility (Temp. up to 80°C) with recirculation system.
12. It should have in-built battery backup of minimum 15 minutes for complete extracorporeal blood system during power failure.
13. Non-Invasive patient blood pressure monitoring.

General Terms & Conditions:

1. The product should be USFDA/European CE (Notified Body) certified. (Valid documentation should be submitted in technical bid).
2. Demonstration of the machine is must for final technical approval.
3. List of consumables, accessories & spare parts as per requirement of the equipment should be provided by supplier in the technical bid.
4. The company have excellent service backup with resident engineer based in major cities of Rajasthan,
5. Installation and training for medical/technical staff to getting acquaint with machine at consignee will be done by the supplier free of cost. Operating manual/training material (1set) will be supplied by the supplier free of cost at the time of installation.
6. 3 year warrantee on complete product from the date of installation.
7. CMC: CMC will be given @ 5% exclusive GST (as applicable) and yearly escalation of 5% on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after completion of warrantee period of three years.
8. Four Preventive Maintenance services during warrantee and CMC period are essential. The company should submit technical compliance sheet as per technical specifications mentioning the make & model of quoted product along with catalogue in the Technical bid.

Technical specifications of RO Water System (250 litre/Hour) for Heamodialysis Machine

1. It should have capacity to produce 250 Liter Hour post RO water.
2. It should have Raw water tank, multi grade Filter, Iron removal, Charcoal (twin) Filter, softener, Reverse Osmosis, Ultra Violet Sterilizer, RO water storage tank of 1000 liters.
3. It should have Rinse and Flush facility for the membranes in R.O. Unit.
4. It should have conductivity meter, Pressure Gauge, Flow indicator.
5. The chemical, contaminants, endotoxin and microbial tests of post R.D water should pass AAMI standards.
6. Technical evaluation of Water Treatment Plant RO System shall be done with technical documents, if the bidder has an installed RO System at any Government institution than bidder has required to submit its satisfactory installation report. Otherwise, The bidder shall make arrangements for physical inspection, where the RO system of the bidder is installed.



The bidder, who gets the purchase order for installation of RO System, will install RO System at one consignee institute and only after getting the report of satisfactory installation and performance from that consignee Institution, will install RO system in the rest of the consignee Institutes.

7. Guarantee for Water Treatment Plant R.O. System shall be taken for two years and CMC for Five years.
8. Firm shall provide a 1000 Liters Water Storage Tank with each Water Treatment Plant (WTP) R.O. System
9. Firm shall install the pipeline from Raw Water Storage Tank or Hospital Campus Water Source to Water Treatment Plant (WTP) R.O. System (with commissioning)
10. All work related to installation of R.O. system and connection of R.O. water to Dialysis Machine and drainage will be completed by the bidder.

- The agreement signed between RMSCL and the firm forms integral part of the rate contract.

Executive Director (EPM)
RMSCL, Jaipur

No. F-8() RMSC/EPM/M-1/NIB-760/RC-821/2023-24/ 698

Dated: 11/11/23

Copy forwarded to the following for information & necessary action:

1. P.S. to Hon'ble Health Minister, Medical & Health Deptt. Govt. of Rajasthan.
2. P.S. to ACS, Medical Health & FW Deptt. Rajasthan, Jaipur.
3. PS to Pr. Secy/Secy, Medical Education Department, Rajasthan, Jaipur
4. P.S. to Mission Director NHM Jaipur.
5. PA to Managing Director, RMSCL
6. Director, (PH)/(Aids)/(FW)/(RCH)/(IEC)/(ESI).
7. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar & Udaipur.
8. Principal R.U.H.S. College of Dental Science Jaipur.
9. Principal, RUHS College of Medical Sciences, Jaipur.
10. Superintendent Japuriya Hospital Jaipur (Associated with RUHS College of Medical Sciences, Jaipur).
11. Superintendent, SMS Hospital, Zenana Hospital, MahilaChikitsalya, Mental Hospital, J.K. Lon Hospital, T.B & Chest Hospital, Jaipur.
12. Superintendent Associated Group of Hospital-Ajmer, M.G. Hospital, Jodhpur, Ummaid Hospital- Jodhpur, PBM Hospital-Bikaner, MBS Hospital-Kota & A.G. Hospital-Udaipur.
13. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
14. Addl. Director, Medical & Health Services (PW)/ (RH)/(Hospital Admn.)
15. Joint Director, Medical & Health Services, (Zone) Jaipur, Ajmer, Udaipur, Jodhpur, Bikaner, Kota & Bharatpur.
16. All P.M.O.'s.
17. All CM&HO's.
18. All Dy. CM&HO's.
19. AGM (IT), RMSC, Jaipur
20. All BME, RMSC
21. M/s Aprameya Engineering Limited, Ahmedabad.
22. Guard file.

Executive Director (EPM)
RMSCL, Jaipur