

# Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, SwasthyaBhawan, TilakMarg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax: 0141-2228065

CIN:U24232RJ2011SGC035067

E-mail: edepmrmsc-rj@nic.in

Website: http://rmsc.health.rajasthan.gov.in

Dated: 22/11/2021

No. F-8() RMSC/EPM/M-3/2021-22/NIB-608/ 9 14

### **Rate Contract and Price Details**

S.N.	Item	Detail			
1.	Name of Equipment / Item	Nebulizer			
2.	Name & Address of Firm	M/s B S Imaging,			
		303/7, Aditya Complex, Preet Vihar			
		Delhi- 110092			
2	Bidder Status	Authorised Distributor			
3.					
4.	Packing Unit	Each			
5.	Brand Make/Model	OMRON/NE-C106			
6.	R/C No. & Validity	R.C. No. 712 (Valid Upto 30.09.2023)			
7.	Performance Security	Amount Rs. 293500.00 through BG No No.			
		0037NDDG00038522 dated 20.10.2021 ICICI Bank,			
		Jaipur			
8.	Indicative Quantity	7000			
9.	Rates for supply to RMSC from Suppliers				
	(i) New Basic Rate (in Rs.)	1495	.00		
	(ii) IGST @ 12%(in Rs. )	. 179	.40		
	(iii) CGST @ 6% (in Rs. )	. 0	0.00		
	(iv) SGST@ 6% (in Rs. )	0	0.00		
	(v) Approved Rate From Suppl	lier To RMSC 1674	.40		
	$\{(i)+(ii)+(iii)+(iv)\}\$ (in Rs.)				
10.	Rate For Supply to Consignee From RMSC				
	(i) Basic Rate (in Rs.)	1495	5.00		
	(ii) RMSC Surcharge@5% On (i) (ii	n Rs.) 74	1.75		
	(iii) CGST@ 6% { On (i)+(ii)} (in R		185		
	(iv) SGST@ 6% { On (i)+(ii)} (in Rs		185		
	(v) Approved Rate for Consign		3.12		
	$\{(i)+(ii)+(iii)+(iv)\}\ (in Rs.)$				
11.	Technical Specifications of Nebulizer				
	General				
	1. USE				
	1.1 Clinical Purpose	Designed to generate aerosolized medication/fluids (finely dispersed airbo	orne		
		droplets in a liquid phase) intended to be inhaled by a patient with a respiral disorder	tory		
	1.2 Used by clinical department/ward	All	-		
	Technical				
	2. TECHNICAL CHARACTERSTICS				
	2.1 Technical Characteristics (specific to this	Medicine cup capacity of minimum 5ml			
	type of device)	Manual	-		
	2.2 Settings 2.3 User's Interface	Manual	-		
	2.3 User's Interface	Manual			



	HYSICAL CHARACTERSTICS	
.1	Dimensions (metric)	Should be compact
.2		<2kg
3.4	Noise (in dBA), heat dissipation	<60dBA (Relevant test report from NABL accredited laboratory should be submitted in the technical bid)
3.5	Mobility, Portability	Yes
4. E	NERGY SOURCE (ELECTRIC	ITY,UPS, SOLAR, GAS,WATER, CO2)
1.1	Power Requirements	220 V AC + 10%, 50Hz power supply;5A Plug;
1.3	Tolerance (to variations, Shutdown)	±10% of input AC
1.4	Protection	Electrical protection by resettable over current breakers or replaceable fuses fitted in both live and neutral lines
4.5	Power Consumptions	Should be compatible with other life saving equipments running parallel
5. A	CCESSORIES, SPARE PARTS	
5.1	Accessories & Spares	With necessary accessories: nebulisation mask (both adult and paediatric size). PVC tubing for nebulizer (Two pair extra); cable cord
5.2	Consumables/reagents (open, closed System)	Aerosol/medicinal solutions
6. E		ARTMENTAL CONSIDERATIONS
5.1	Atmosphere/Ambiance (air conditioning,	Capable of being stored continuously in ambient temperature of 0 to 50 deg C
	humidity, dust)	and relative humidity of 15 to 90%. Capable of operating continuously in
		ambient temperature of 10 to 40 deg C and relative humidity of 15 to 90%
6.2	User's Care, Cleaning, Disinfection & Sterility issues	The unit should be cleanable with alcohol and/or/other chemical agents.
	STANDARDS & SAFETY	FDA(US)/European CE (Notified Body)/BIS and ISO 13485:2003; ISO 27427-
7.1	Certificates (Pre-Market, Sanitary)	FDA(US)/European CE (Notified Body)/BIS and ISO 13485:2003; ISO 27427-2013; IEC-60601-1
7.1	STANDARDS & SAFETY	2013; IEC-60601-1 ON
7.1 <b>8. T</b>	Certificates (Pre-Market, Sanitary)  TRAINING AND INSTALLATION Pre-installation requirements: nature,	2013; IEC-60601-1
7.1 <b>8. T</b> 8.1	Certificates (Pre-Market, Sanitary)  FRAINING AND INSTALLATION  Pre-installation requirements: nature, values, quality, tolerance	2013; IEC-60601-1  ON  Supplier to perform installation, safety and operation checks before handover  Certificate of calibration and inspection from the factory
7.1 <b>8. T</b> 8.1 8.2	Certificates (Pre-Market, Sanitary)  FRAINING AND INSTALLATION  Pre-installation requirements: nature, values, quality, tolerance  Requirements for sign-off  Training of staff(Medical, Paramedical,	2013; IEC-60601-1  ON  Supplier to perform installation, safety and operation checks before handover
7.1 <b>8. T</b> 8.1  8.2  8.3	Certificates (Pre-Market, Sanitary)  FRAINING AND INSTALLATION  Pre-installation requirements: nature, values, quality, tolerance  Requirements for sign-off  Training of staff(Medical, Paramedical, Technicians)	2013; IEC-60601-1  ON  Supplier to perform installation, safety and operation checks before handover  Certificate of calibration and inspection from the factory  Training of users in operation and basic maintenance shall be provided
7.1 8. T 8.1 8.2 8.3	Certificates (Pre-Market, Sanitary)  TRAINING AND INSTALLATION Pre-installation requirements: nature, values, quality, tolerance Requirements for sign-off Training of staff(Medical, Paramedical, Technicians)  WARRANTY AND MAINTENA	2013; IEC-60601-1  ON  Supplier to perform installation, safety and operation checks before handover  Certificate of calibration and inspection from the factory  Training of users in operation and basic maintenance shall be provided
7.1 8. T 8.1 8.2 8.3 9. V	Certificates (Pre-Market, Sanitary)  TRAINING AND INSTALLATION Pre-installation requirements: nature, values, quality, tolerance Requirements for sign-off Training of staff(Medical, Paramedical, Technicians)  WARRANTY AND MAINTENA Warranty	2013; IEC-60601-1  ON  Supplier to perform installation, safety and operation checks before handover  Certificate of calibration and inspection from the factory  Training of users in operation and basic maintenance shall be provided  NCE  3 Years  Maintenance manual detailing complete maintaining schedule
7.1 8. T 8.1 8.2 8.3 9. V 9.1 9.2	Certificates (Pre-Market, Sanitary)  TRAINING AND INSTALLATION Pre-installation requirements: nature, values, quality, tolerance Requirements for sign-off Training of staff(Medical, Paramedical, Technicians)  WARRANTY AND MAINTENA Warranty Maintenance tasks	2013; IEC-60601-1  ON  Supplier to perform installation, safety and operation checks before handover  Certificate of calibration and inspection from the factory  Training of users in operation and basic maintenance shall be provided  ANCE  3 Years
7.1 8. T 8.1 8.2 8.3 9. V 9.1 9.2 9.3	Certificates (Pre-Market, Sanitary)  FRAINING AND INSTALLATION  Pre-installation requirements: nature, values, quality, tolerance  Requirements for sign-off  Training of staff(Medical, Paramedical, Technicians)  WARRANTY AND MAINTENA  Warranty  Maintenance tasks  Service contract clauses, including prices	2013; IEC-60601-1  ON  Supplier to perform installation, safety and operation checks before handover  Certificate of calibration and inspection from the factory  Training of users in operation and basic maintenance shall be provided  NCE  3 Years  Maintenance manual detailing complete maintaining schedule
7.1 8. T 8.1 8.2 8.3 9. V 9.1 9.2 9.3 10.	Certificates (Pre-Market, Sanitary)  FRAINING AND INSTALLATIO  Pre-installation requirements: nature, values, quality, tolerance Requirements for sign-off Training of staff(Medical, Paramedical, Technicians)  WARRANTY AND MAINTENA  Warranty Maintenance tasks Service contract clauses, including prices  DOCUMENTATION	2013; IEC-60601-1  ON  Supplier to perform installation, safety and operation checks before handover  Certificate of calibration and inspection from the factory  Training of users in operation and basic maintenance shall be provided  NCE  3 Years  Maintenance manual detailing complete maintaining schedule  Local clinical staff to affirm completion of installation  Advanced maintenance tasks required shall be documented
7.1 8. T 8.1 8.2 8.3 9. V 9.1 9.2 9.3 10.	Certificates (Pre-Market, Sanitary)  FRAINING AND INSTALLATIO  Pre-installation requirements: nature, values, quality, tolerance Requirements for sign-off Training of staff(Medical, Paramedical, Technicians)  WARRANTY AND MAINTENA  Warranty Maintenance tasks Service contract clauses, including prices  DOCUMENTATION  Operating manuals, service manuals,	2013; IEC-60601-1  ON  Supplier to perform installation, safety and operation checks before handover  Certificate of calibration and inspection from the factory  Training of users in operation and basic maintenance shall be provided  NCE  3 Years  Maintenance manual detailing complete maintaining schedule  Local clinical staff to affirm completion of installation  Advanced maintenance tasks required shall be documented  User, technical and maintenance manuals to be supplied in English language
7.1 8. T 8.2 8.3 9. V 9.1 9.2 9.3 10.	Certificates (Pre-Market, Sanitary)  FRAINING AND INSTALLATIO  Pre-installation requirements: nature, values, quality, tolerance Requirements for sign-off Training of staff(Medical, Paramedical, Technicians)  WARRANTY AND MAINTENA  Warranty Maintenance tasks Service contract clauses, including prices  DOCUMENTATION	2013; IEC-60601-1  ON  Supplier to perform installation, safety and operation checks before handover  Certificate of calibration and inspection from the factory  Training of users in operation and basic maintenance shall be provided  NCE  3 Years  Maintenance manual detailing complete maintaining schedule  Local clinical staff to affirm completion of installation  Advanced maintenance tasks required shall be documented  User, technical and maintenance manuals to be supplied in English language
7.1 8. T 8.1 8.2 8.3 9. V 9.1 9.2 9.3 10.	Certificates (Pre-Market, Sanitary)  FRAINING AND INSTALLATIO  Pre-installation requirements: nature, values, quality, tolerance Requirements for sign-off Training of staff(Medical, Paramedical, Technicians)  WARRANTY AND MAINTENA  Warranty Maintenance tasks Service contract clauses, including prices  DOCUMENTATION  Operating manuals, service manuals,	2013; IEC-60601-1  ON  Supplier to perform installation, safety and operation checks before handover  Certificate of calibration and inspection from the factory  Training of users in operation and basic maintenance shall be provided  NCE  3 Years  Maintenance manual detailing complete maintaining schedule  Local clinical staff to affirm completion of installation  Advanced maintenance tasks required shall be documented  User, technical and maintenance manuals to be supplied in English language  List to be provided of equipment and procedures required for local calibration and routine maintenance
7.1 8. T 8.1 8.2 8.3 9. V 9.1 9.2 9.3 10.1	Certificates (Pre-Market, Sanitary)  FRAINING AND INSTALLATION  Pre-installation requirements: nature, values, quality, tolerance Requirements for sign-off Training of staff(Medical, Paramedical, Technicians)  WARRANTY AND MAINTENA  Warranty Maintenance tasks Service contract clauses, including prices  DOCUMENTATION  Operating manuals, service manuals, other manuals	2013; IEC-60601-1  ON  Supplier to perform installation, safety and operation checks before handover  Certificate of calibration and inspection from the factory  Training of users in operation and basic maintenance shall be provided  NCE  3 Years  Maintenance manual detailing complete maintaining schedule  Local clinical staff to affirm completion of installation  Advanced maintenance tasks required shall be documented  User, technical and maintenance manuals to be supplied in English language  List to be provided of equipment and procedures required for local calibratiand routine maintenance  List to be provided of important spares and accessories with their part numbers.
7.1 8. T 8.2 8.3 9. V 9.1 9.2 9.3 10.1	Certificates (Pre-Market, Sanitary)  FRAINING AND INSTALLATION  Pre-installation requirements: nature, values, quality, tolerance Requirements for sign-off Training of staff(Medical, Paramedical, Technicians)  WARRANTY AND MAINTENA  Warranty Maintenance tasks Service contract clauses, including prices  DOCUMENTATION  Operating manuals, service manuals, other manuals  Other accompanying documents	2013; IEC-60601-1  ON  Supplier to perform installation, safety and operation checks before handover  Certificate of calibration and inspection from the factory  Training of users in operation and basic maintenance shall be provided  NCE  3 Years  Maintenance manual detailing complete maintaining schedule  Local clinical staff to affirm completion of installation  Advanced maintenance tasks required shall be documented  User, technical and maintenance manuals to be supplied in English language  List to be provided of equipment and procedures required for local calibration and routine maintenance
7.1 8. T 8.1 8.2 8.3 9. V 9.1 9.2 9.3 10.1	Certificates (Pre-Market, Sanitary)  FRAINING AND INSTALLATION  Pre-installation requirements: nature, values, quality, tolerance Requirements for sign-off Training of staff(Medical, Paramedical, Technicians)  WARRANTY AND MAINTENA  Warranty  Maintenance tasks Service contract clauses, including prices  DOCUMENTATION  Operating manuals, service manuals, other manuals  Other accompanying documents	2013; IEC-60601-1  ON  Supplier to perform installation, safety and operation checks before handover  Certificate of calibration and inspection from the factory  Training of users in operation and basic maintenance shall be provided  NCE  3 Years  Maintenance manual detailing complete maintaining schedule  Local clinical staff to affirm completion of installation  Advanced maintenance tasks required shall be documented  User, technical and maintenance manuals to be supplied in English language  List to be provided of equipment and procedures required for local calibratiand routine maintenance  List to be provided of important spares and accessories with their part number and cost. Certificate of calibration and inspection to be provided
7.1 8. T 8.1 8.2 8.3 9. V 9.1 9.2 9.3 10.1	Certificates (Pre-Market, Sanitary)  FRAINING AND INSTALLATIO  Pre-installation requirements: nature, values, quality, tolerance  Requirements for sign-off  Training of staff(Medical, Paramedical, Technicians)  WARRANTY AND MAINTENA  Warranty  Maintenance tasks  Service contract clauses, including prices  DOCUMENTATION  Operating manuals, service manuals, other manuals  Other accompanying documents  NOTES  Service Support Contact details (Hierchy Wise; including a toll free/ landline	2013; IEC-60601-1  ON  Supplier to perform installation, safety and operation checks before handover  Certificate of calibration and inspection from the factory  Training of users in operation and basic maintenance shall be provided  NCE  3 Years  Maintenance manual detailing complete maintaining schedule  Local clinical staff to affirm completion of installation  Advanced maintenance tasks required shall be documented  User, technical and maintenance manuals to be supplied in English language  List to be provided of equipment and procedures required for local calibratiand routine maintenance  List to be provided of important spares and accessories with their part number and cost. Certificate of calibration and inspection to be provided  Contact details of manufacturer, supplier and local service agent to be provided
7.1 8. T 8.1 8.2 8.3 9. V 9.1 9.2 9.3 10.1	Certificates (Pre-Market, Sanitary)  FRAINING AND INSTALLATIO  Pre-installation requirements: nature, values, quality, tolerance Requirements for sign-off Training of staff(Medical, Paramedical, Technicians)  WARRANTY AND MAINTENA  Warranty Maintenance tasks Service contract clauses, including prices  DOCUMENTATION  Operating manuals, service manuals, other manuals  Other accompanying documents  NOTES  Service Support Contact details (Hierchy Wise; including a toll free/ landline number)	2013; IEC-60601-1  ON  Supplier to perform installation, safety and operation checks before handover  Certificate of calibration and inspection from the factory  Training of users in operation and basic maintenance shall be provided  NCE  3 Years  Maintenance manual detailing complete maintaining schedule  Local clinical staff to affirm completion of installation  Advanced maintenance tasks required shall be documented  User, technical and maintenance manuals to be supplied in English language  List to be provided of equipment and procedures required for local calibratic and routine maintenance  List to be provided of important spares and accessories with their part number and cost. Certificate of calibration and inspection to be provided

The agreement signed between RMSCL and the firm forms integral part of the rate contract.

Executive Director (EPM)
RMSC, Jaipur

Dated: 22/11/2024

## ,. F-8() RMSC/EPM/M-3/2021-22/NIB-608/ 9114

Copy forwarded to the following for information & necessary action:

- 1. P.S. to Hon'ble Health Minister, Medical & Health Deptt. Govt. of Rajasthan.
- 2. P.S. to ACS/Pr. Secy/Secy, Medical Health & FW Deptt. Rajasthan, Jaipur.
- 3. PS to Pr. Secy/Secy, Medical Education Department, Rajasthan, Jaipur
- 4. P.S. to Mission Director NHM Jaipur.
- 5. PA to Managing Director, RMSCL
- 6. Director, (PH)/(Aids)/(FW)/(RCH)/(IEC)/(ESI).
- 7. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar& Udaipur.
- 8. Principal R.U.H.S. College of Dental Science Jaipur.
- 9. Principal, RUHS College of Medical Scinces, Jaipur.
- 10. Superintendent Japuriya Hospital Jaipur (Associated with RUHS College of Medical Scinces, Jaipur).
- 11. Superintendent, SMS Hospital, Zenana Hospital, MahilaChikitsalya, Mental Hospital, J.K. Lon Hospital, T.B & Chest Hospital, Jaipur.
- 12. Superintendent Associated Group of Hospital-Ajmer, M.G. Hospital, Jodhpur, Ummaid Hospital-Jodhpur, PBM Hospital-Bikaner, MBS Hospital-Kota & A.G. Hospital-Udaipur.
- 13. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
- 14. Addl. Director, Medical & Health Services (PW)/ (RH)/(Hospital Admn.)
- 15. Joint Director, Medical & Health Services, (Zone) Jaipur, Ajmer, Udaipur, Jodhpur, Bikaner, Kota & Bharatpur.
- 16. All P.M.O.'s.
- 17. All CM&HO's.
- 18. All Dy. CM&HO's.
- 19. AGM (IT), RMSC, Jaipur.
- 20. M/s B S Imaging, Delhi
- 21. Guard file.

**Executive Director (EPM)** 

RMSC, Jaipur

### मुख्यमंत्री निःशल्क जाँच योजन

#### Rajasthan Medical Services Corporation Limited, Jaipur

'D'-block, swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur

Phone no. 0141-2223887, Fax no. 0141-2228065

Website: http://rmsc.health.rajasthan.gov.in

CIN: U24232RJ2011SGC035067

Dated: 22/11/2021

No. F-8() RMSC/EPM/M-3/2021-22/NIB-608/ 9113

All Principals Medical Colleges,

All the Superintendents of Medical Colleges& Attached Hospitals.

Director (PH/HA/IEC/AIDS/RCH)

Sub:- Approval/Purchase of Item, "Nebulizer", for the Rate Contract and supply, ending on 30.09.2023.

Please find enclosed herewith approved firm and "Nebulizer", for the rate contract and supply ending on 30.09.2023. The rate contract and its terms & conditions are also available on Rajasthan Medical Services Corporation Ltd., Website www.rmsc.nic.in.

In reference to the above rate contract No. F-8( ) RMSC/EPM/M-3/2021-22/NIB-608/..... Dated...... the supplier has executed the agreement with RMSC, as enclosed herewith for your information & necessary action. This bears an approval of M.D., RMSCL, Jaipur.

**Executive Director (EPM)** 

Dated: 22/11/2021

RMSC, Jaipur

No. F-8() RMSC/EPM/M-3/2021-22/NIB-608/9 \\3

Copy forwarded to the following for information & necessary action:

- 1. P.S. to Hon'ble Health Minister, Medical & Health Deptt. Govt. of Rajasthan.
- 2. P.S. to ACS/Pr. Secy/Secy, Medical Health & FW Deptt. Rajasthan, Jaipur.
- 3. PS to Pr. Secy/Secy, Medical Education Department, Rajasthan, Jaipur
- 4. P.S. to Mission Director NHM Jaipur.
- 5. PA to Managing Director, RMSCL
- 6. Director, (PH)/(Aids)/(FW)/(RCH)/(IEC)/(ESI).
- 7. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar& Udaipur.
- 8. Principal R.U.H.S. College of Dental Science Jaipur.
- 9. Principal, RUHS College of Medical Scinces, Jaipur.
- 10. Superintendent Japuriya Hospital Jaipur (Associated with RUHS College of Medical Scinces, Jaipur).
- 11. Superintendent, SMS Hospital, Zenana Hospital, MahilaChikitsalya, Mental Hospital, J.K. Lon Hospital, T.B & Chest Hospital, Jaipur.
- 12. Superintendent Associated Group of Hospital-Ajmer, M.G. Hospital, Jodhpur, Ummaid Hospital- Jodhpur, PBM Hospital-Bikaner, MBS Hospital-Kota & A.G. Hospital-Udaipur.
- 13. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
- 14. Addl. Director, Medical & Health Services (PW)/ (RH)/(Hospital Admn.)
- 15. Joint Director, Medical & Health Services, (Zone) Jaipur, Ajmer, Udaipur, Jodhpur, Bikaner, Kota & Bharatpur.
- 16. All P.M.O.'s.
- 17. All CM&HO's.
- 18. All Dy. CM&HO's.
- 19. AGM (IT), RMSC, Jaipur.
- 20. M/s B S Imaging, Delhi
- 21. Guard file.

**Executive Director (EPM)** 

RMSC, Jaipur