

Rajasthan Medical Services Corporation Limited, Jaipur

'D'-block, swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur

Phone no. 0141-2223887, Fax no. 0141-2228065

E-mail: edepmrmse-rj@nic.in

Website: http://rmse.health.rajasthan.gov.in

CIN : U24232RJ2011SGC035067

No. F-8() RMSC/EPM/M-1/NIB-504/2021-22/ 8652


Dated: 01/10/2021

All Principals Medical Colleges,
All the Superintendents of Medical Colleges & Attached Hospitals.
Director (PH/HA/IEC/AIDS/RCH)

Sub:- Approval/Purchase of Item, 'Blood Collection Monitor', for the Rate Contract and supply, ending on 30.09.2023.

Please find enclosed herewith approved firm and Blood Collection Monitor, for the rate contract and supply ending on 30.09.2023. The rate contract and its terms & conditions are also available on Rajasthan Medical Services Corporation Ltd., Website www.rmse.health.rajasthan.gov.in

In reference to the above rate contract No. F-8() RMSC/EPM/M-1/2021-22/NIB-504/8652, Dated: 01/10/2021 the supplier has executed the agreement with RMSC, as enclosed herewith for your information & necessary action. This bears an approval of M.D., RMSCL, Jaipur.


Executive Director (EPM)

Rajasthan Medical Services Corporation Ltd, Jaipur

No. F-8() RMSC/EPM/M-1/NIB-504/2021-22/ 8652

Dated: 01/10/2021

Copy forwarded to the following for information & necessary action:

1. P.S. to Hon'ble Health Minister, Medical & Health Deptt. Govt. of Rajasthan.
2. P.S. to ACS/ Pr. Secy/Secy, Medical Health & FW Deptt. Rajasthan, Jaipur.
3. PS to Pr. Secy/Secy, Medical Education Department, Rajasthan, Jaipur
4. P.S. to Mission Director NHM Jaipur.
5. PA to Managing Director, RMSCL
6. Director. (PH)/(Aids)/(FW)/(RCH)/(IEC)/(ESI).
7. Principal. Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar & Udaipur.
8. Principal R.U.H.S. College of Dental Science Jaipur.
9. Principal, RUHS College of Medical Sciences, Jaipur.
10. Superintendent Japuriya Hospital Jaipur (Associated with RUHS College of Medical Sciences, Jaipur).
11. Superintendent, SMS Hospital, Zenana Hospital, MahilaChikitsalya, Mental Hospital, J.K. Lon Hospital, T.B & Chest Hospital, Jaipur.
12. Superintendent Associated Group of Hospital-Ajmer, M.G. Hospital, Jodhpur, Ummaid Hospital- Jodhpur, PBM Hospital- Bikaner, MBS Hospital-Kota & A.G. Hospital-Udaipur.
13. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
14. Addl. Director, Medical & Health Services (PW)/ (RH)/(Hospital Admn.)
15. Joint Director, Medical & Health Services, (Zone) Jaipur, Ajmer, Udaipur, Jodhpur, Bikaner, Kota & Bharatpur.
16. All P.M.O.'s.
17. All CM&HO's.
18. All Dy. CM&HO's.
19. AGM (IT), RMSC, Jaipur
20. All BME, RMSC
21. M/s Remi Elektrotechnik Limited, Mumbai
22. Guard file.


Executive Director (EPM)

Rajasthan Medical Services Corporation Ltd, Jaipur

Rajasthan Medical Services Corporation Limited, Jaipur**Gandhi Block, SwasthyaBhawan, TilakMarg, C-Scheme, Jaipur - 302005**

Ph. No. 0141-2223887, Fax : 0141-2228065

E-mail : edepmrmsc-rj@nic.in

CIN:U24232RJ2011SGC035067

Website: http://rmsc.health.rajabhawan.gov.in

No. F-8() RMSC/EPM/M-1/NIB-504/2021-22/ 8651

Dated: 01/10/2021

Rate Contract and Price Details

S.N.	Item	Detail
1.	Name of Equipment / Item	Blood Collection Monitor
2.	Name & Address of Firm	M/s Remi Elektrotechnik Limited, Remi House, 11 Cama Indl. Estate, Goregaon (E), Mumbai-400 063, India E-Mail:- bids@remilabworld.com
3.	Bidder Status	Manufacturer
4.	Performance Security	SBI BG No. 6122421BG0001214 dt 07.09.2021 Rs. 98000.00 Valid up to 30.04.2024
5.	Brand/ Make	REMI
	Model	BCM 20 Ultra
6.	Packing Unit	Each
7.	R/C No. & Validity	R.C. No. 687 (Valid Up to 30.09.2023)
8.	Indicative Quantity	50
9.	Rates for supply to RMSC from Suppliers	
	(i) New Basic Rate (in Rs.)	70000.00
	(ii) IGST @ 12% (in Rs.)	8400.00
	(iii) CGST (in Rs.)	0.00
	(iv) SGST (in Rs.)	0.00
	(v) Approved Rate From Supplier To RMSC {(i)+(ii)+(iii)+(iv)} (in Rs.)	78400.00
10.	Rate For Supply to Consignee From RMSC	
	(i) Basic Rate (in Rs.)	70000.00
	(ii) RMSC Surcharge@5% On (i) (in Rs.)	3500.00
	(iii) CGST@ 6 % { On (i)+(ii)} (in Rs.)	4410.00
	(iv) SGST@ 6 % { On (i)+(ii)} (in Rs.)	4410.00
	(v) Approved Rate for Consignee from RMSC {(i)+(ii)+(iii)+(iv)} (in Rs.)	82320.00
11.	<u>Technical Specifications of Item Blood Collection Monitor</u>	
	Bag holder tray suitable for all types of blood Bags. Magnetic removable cradle Tray for easy cleaning	
	1. Should have facility to preset total volume of blood to be collected and accordingly monitor and display amount collected. It should have facility to clamp to stop the collection of blood as soon as preset volume is collected and not allow over collection.	
	2. Battery backup should be > 8 hours (12VDC) with continuous workload(rechargeable battery)	
	3. Battery charger should be inbuilt.	
	4. Should be portable (Suitable for outdoor blood donation camps).	
	5. Should have standby / pause mode Manual clamp facility to abort collection Automatic Release of Clamp when the Bag is lifted.	
	6. Should be able to operate at Temperature of +5° to +45° C and relative humidity (RH) of 5 to 95 %	
	7. There should be continuous digital display of preset volume, Blood flow rate and total time taken at the	



- end of collection.
8. Oscillation: 12 ± 2 rpm
 9. Should mix the blood with anti – coagulant solution during Collection and ensure that only correct amount of blood is collected
 10. There Should be Visual display and audible alarm:
 - (i) when flow rate goes below 20 ml /min or high flow rate above 180 ml / min
 - (ii) at the end of collection
 - (iii) when battery low
 - (iv) during pause function
 - (v) any abnormal condition
 11. **Quality Standard**
 - a) Manufacturing should be compliant with ISO 13485 (Valid documentation should be submitted in technical bid).
 - b) Should be compliant with European CE (Notified body) /US FDA (Valid documentation should be submitted in technical bid).
 - c) Equipment must meet electrical safety specifications of IEC 60601-1 and 60601-1-2 (as relevant)
 12. Every Bio-mixer should be provided with carry box with handle
 13. Firm should supply the relevant calibration certificate for the equipment from NABL accredited Lab.
 14. User's list should be provided with satisfactory report for the last three years from three Licensed Blood Banks with contact details.
 15. Original literature of equipment should be submitted.
 16. Electrical: The equipment should be able to run on the existing Electrical provision.
 17. **Broad specifications are:**
Automatic Type Input 150-280V, Output 220 V +/- 7 %, 50 Hz /60 Hz.
Single phase, AC with automatic 2-4 sec Cut Off and 6-9 minutes
Restart delay. Quick start arrangements for bypassing the start
delay. Suitable MCB on input voltmeter and indicators on Front
Panel. Input Power Cable with 15 A Plug and six way output
Terminal strip for two outlets
 18. **Accessories:-**
 - Floor stand
 - Transport case with built-in charger
 - Suitable Calibration weight
 19. **Guarantee:** Two years on equipment from the date of installation.
 20. **CMC:** CMC will be given @ 5 % of net rate – Rate exclusive of GST (as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after completion of Guarantee period of two years.
 21. The company should submit technical compliance sheet as per technical specifications mentioning the make & model of quoted item along with catalogue in the Technical bid.
 22. Installation will be done by supplier free of cost.
 23. The service engineer should be based in Rajasthan.
 24. Two Preventive Maintenance services during Guarantee and CMC period are essential.
 25. Demonstration of equipment is must for final technical approval

- The agreement signed between RMSCL and the firm forms integral part of the rate contract.


Executive Director (EPM)
RMSC, Jaipur

Dated: 01/10/2021

No. F-8() RMSC/EPM/M-1/NIB-504/2021-22/ 8651

Copy forwarded to the following for information & necessary action:

1. P.S. to Hon'ble Health Minister, Medical & Health Deptt. Govt. of Rajasthan.
2. P.S. to ACS/ Pr. Secy/Secy, Medical Health & FW Deptt. Rajasthan, Jaipur.



3. PS to Pr. Secy/Secy, Medical Education Department, Rajasthan, Jaipur
4. P.S. to Mission Director NHM Jaipur.
5. PA to Managing Director, RMSCL
6. Director, (PH)/(Aids)/(FW)/(RCH)/(IEC)/(ESI).
7. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Bharatpur & Udaipur.
8. Principal R.U.H.S. College of Dental Science Jaipur.
9. Principal, RUHS College of Medical Sciences, Jaipur.
10. Superintendent Japuriya Hospital Jaipur (Associated with RUHS College of Medical Sciences, Jaipur).
11. Superintendent, SMS Hospital, Zenana Hospital, MahilaChikitsalya, Mental Hospital, J.K. Lon Hospital, T.B & Chest Hospital, Jaipur.
12. Superintendent Associated Group of Hospital-Ajmer, M.G. Hospital, Jodhpur, Ummaid Hospital- Jodhpur, PBM Hospital-Bikaner, MBS Hospital-Kota & A.G. Hospital-Udaipur.
13. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
14. Addl. Director, Medical & Health Services (PW)/ (RH)/(Hospital Admn.)
15. Joint Director, Medical & Health Services, (Zone) Jaipur, Ajmer, Udaipur, Jodhpur, Bikaner, Kota & Bharatpur.
16. All P.M.O.'s.
17. All CM&HO's.
18. All Dy. CM&HO's.
19. AGM (IT), RMSC, Jaipur
20. All BME, RMSC
21. M/s Remi Elektrotechnik Limited, Mumbai.
22. Guard file.


Executive Director (EPM)
RMSC, Jaipur