

मुख्यमंत्री निःशुल्क जांच

## Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, SwasthyaBhawan, TilakMarg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065

CIN:U24232RJ2011SGC035067

E-mail : edepmrmisc-ri@nic.in

Website: http://rmisc.health.rajasthan.gov.in

No.F-8(308) RMSC/EPM/M-2/NIB-633/2021-22/ 258

Dated: 1/7/22

All Principals Medical Colleges,  
All the Superintendents of Medical Colleges & Attached Hospitals.  
Director (PH/HA/IEC/AIDS/RCH)

**Sub:- Approval/Purchase of Item, 'Multi Para Monitor, for the Rate Contract and supply, ending on 30.06.2024.**

Please find enclosed herewith approved firm and Multi Para Monitor for the rate contract and supply ending on 30.06.2024. The rate contract and its terms & conditions are also available on Rajasthan Medical Services Corporation Ltd., Website www. rmisc.health.rajasthan.gov.in

In reference to the above rate contract No.F-8(308) RMSC/EPM/M-2/NIB-633/2021-22/..... Dated ..... the supplier has executed the agreement with RMSC, as enclosed herewith for your information & necessary action. This bears an approval of M.D., RMSCL, Jaipur.

*[Signature]*  
Executive Director (EPM)

Rajasthan Medical Services Corporation Ltd, Jaipur

Dated:

No.F-8(308) RMSC/EPM/M-2/NIB-633/2021-22/

Copy forwarded to the following for information & necessary action:

1. P.S. to Hon'ble Health Minister, Medical & Health Deptt. Govt. of Rajasthan.
2. P.S. to ACS/Principal Secy/Secy. Medical Health & FW Deptt. Rajasthan, Jaipur.
3. PS to Pr. Secy/Secy, Medical Education Department, Rajasthan, Jaipur
4. P.S. to Mission Director NHM Jaipur.
5. PA to Managing Director, RMSCL
6. Director, (PH)/(Aids)/(FW)/(RCH)/(IEC)/(ESI).
7. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar & Udaipur.
8. Principal R.U.H.S. College of Dental Science Jaipur.
9. Principal, RUHS College of Medical Sciences, Jaipur.
10. Superintendent Japuriya Hospital Jaipur (Associated with RUHS College of Medical Sciences, Jaipur).
11. Superintendent, SMS Hospital, Zenana Hospital, MahilaChikitsalya, Mental Hospital, J.K. Lon Hospital, T.B & Chest Hospital, Jaipur.
12. Superintendent Associated Group of Hospital-Ajmer, M.G. Hospital, Jodhpur, Ummaid Hospital- Jodhpur, PBM Hospital-Bikaner, MBS Hospital-Kota & A.G. Hospital-Udaipur.
13. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
14. Addl. Director, Medical & Health Services (PW)/ (RH)/(Hospital Admn.)
15. Joint Director, Medical & Health Services, (Zone) Jaipur, Ajmer, Udaipur, Jodhpur, Bikaner, Kota & Bharatpur.
16. All P.M.O.'s.
17. All CM&HO's.
18. All Dy. CM&HO's.
19. AGM (IT), RMSC, Jaipur
20. All Zonal BME are directed to carry out inspection of the supplied equipment as per RMSCL Letter no. 2513 dated 10-07-2018.
21. M/s Clarity Medical Pvt. Ltd., Plot No. 1687-A, JLPL Industrial Park, Sector-82, Mohali-140306 (PB) India  
Email:-contact@clarity-medical.com
22. Guard file.

*[Signature]*  
Executive Director (EPM)

Rajasthan Medical Services Corporation Ltd, Jaipur

# Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, SwasthyaBhawan, TilakMarg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065  
CIN:U24232RJ2011SGC035067

E-mail : edepmrmisc-rj@nic.in  
Website: http://rmisc.health.rajasthan.gov.in

No.F-8(308) RMSC/EPM/M-2/NIB-633/2021-22/ 258

Dated: 11/7/22

## Rate Contract and Price Details

S.N.	Item	Detail
1.	Name of Equipment / Item	Multi Para Monitor
2.	Name & Address of Firm	M/s Clarity Medical Pvt. Ltd., Plot No. 1687-A, JLPL Industrial Park, Sector-82, Mohali-140306 (PB) India Email:-contact@clarity-medical.com
3.	Bidder Status	Manufacturer (Non MSME)
4.	Performance Security	Performance Security Deposit by Bank Guarantee No. BG No. 0058NDDG00004523 Dated 24.06.2022 (ICICI Bank Ltd., SCF 21-22, Phase VII, SAS Nagar, Mohali, Punjab-160059 for Amount of Rs. 9,64,000/- (Rupees Nine Lack Sixty Four Thousand Only) (Valid Up to 30.11.2027)
5.	HSN Code	90181990
6.	Make	Clarity
	Brand/Model	Spectra Smart +
7.	Packing Unit	Each
8.	R/C No. & Validity	R.C. No. 767 (Valid Up to 30.06.2024)
9.	Indicative Quantity	500
10.	Rates for supply to RMSC from Suppliers	
	(i) Basic Rate (in Rs.)	68786.00
	(ii) IGST @ 12% (in Rs.)	8254.32
	(iii) CGST (in Rs.)	0.00
	(iv) SGST (in Rs.)	0.00
	(v) Approved Rate From Supplier To RMSC {(i)+(ii)+(iii)+(iv)} (in Rs.)	77040.32
11.	Rate For Supply to Consignee From RMSC	
	(i) Basic Rate (in Rs.)	68786.00
	(ii) RMSC Surcharge@ 5% On (i) (in Rs.)	3439.00
	(iii) CGST@ 6% {On (i)+(ii)} (in Rs.)	4333.50
	(iv) SGST@ 6% {On (i)+(ii)} (in Rs.)	4333.50
	(v) Approved Rate for Consignee from RMSC {(i)+(ii)+(iii)+(iv)} (in Rs.)	80892.00
12.	Specifications of item "Multi Para Monitor"	
	1) It should having ECG, SpO2, NIBP, Respiration and two Temperature as standard parameters.	
	2) Should have facility to display 7 lead ECG, RR, HR, SpO2, NIBP, and dual Temperature.	
	3) Should display 6 or more waveforms of selected parameters simultaneously, should have colour coding for different waveforms.	
	4) Should have inbuilt battery backup of at least three hours.	
	5) Display: Color LED/LCD touch screen display size of 12" or more.	
	6) Monitor should have both touch screen and knob to operate.	
	7) Should have dual temperature monitoring either in 0Cecius or 0Fahrenheit.	



- 8) Should have facility for displaying multi - screen configurations.
- 9) Should be able to store & display at least 96 hours of graphical trends of all parameters
- 10) Should be suitable for monitoring adult & pediatric & neonate patients.
- 11) The SpO2 technology should be Masimo/Nellcor/TruSignal and it should also sense in hypotension, shivering & motion.
- 12) Should have oscillometric technology for measurement of NIBP with Auto, START and Manual modes.
- 13) Should have different patient type selection.
- 14) The pulse rate should be displayed either with ECG or SpO2.
- 15) The respiration rate should be calculated through Impedance method.
- 16) Should be able to analyze arrhythmias & ST segment changes.
- 17) Memory should not wipe off when the power is turned off.
- 18) Should be able to give visual & audible alarms with three levels of volume adjustment.
- 19) Should have facility of connecting to the central monitoring station having facility for real time monitoring through Ethernet cable or Wireless Connectivity.
- 20) Monitor should have VGA port for External Display
- 21) Wall mount stand should be provided free of the cost with the monitor.
- 22) Quoted model should be European CE (Notified Body)/USFDA approved.
- 23) Warrantee-Three years on equipment from the date of installation.
- 24) CMC: CMC will be given @ 5 % of Net Rate exclusive of GST (As applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after completion of Warrantee period of three years.
- 25) Equipment should be supplied with compatible 5 lead ECG cable (1 Nos.), SpO2 (Adult,Pediatric,Neonate) probe with extension cable (Each 1No.), NIBP Cuff with extension cable of Large Adult, Adult & Pediatric (2 each), Temperature probe (2 Nos.) & Operating / user Manual (1 No.).

Executive Director (EPM)

RMSC, Jaipur

Dated: 11/7/22


No.F-8(308) RMSC/EPM/M-2/NIB-633/2021-22/ 258

Copy forwarded to the following for information & necessary action:

1. P.S. to Hon'ble Health Minister, Medical & Health Deptt. Govt. of Rajasthan.
2. P.S. to ACS/Principal Secy/Secy. Medical Health & FW Deptt. Rajasthan, Jaipur.
3. PS to Pr. Secy/Secy, Medical Education Department, Rajasthan, Jaipur
4. P.S. to Mission Director NHM Jaipur.
5. PA to Managing Director, RMSCL
6. Director, (PH)/(Aids)/(FW)/(RCH)/(IEC)/(ESI).
7. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar & Udaipur.
8. Principal R.U.H.S. College of Dental Science Jaipur.
9. Principal, RUHS College of Medical Sciences, Jaipur.
10. Superintendent Japuriya Hospital Jaipur (Associated with RUHS College of Medical Sciences, Jaipur).
11. Superintendent, SMS Hospital, Zenana Hospital, MahilaChikitsalya, Mental Hospital, J.K. Lon Hospital, T.B & Chest Hospital, Jaipur.
12. Superintendent Associated Group of Hospital-Ajmer, M.G. Hospital, Jodhpur, Ummaid Hospital-Jodhpur, PBM Hospital-Bikaner, MBS Hospital-Kota & A.G. Hospital-Udaipur.
13. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
14. Addl. Director, Medical & Health Services (PW)/ (RH)/(Hospital Admn.)
15. Joint Director, Medical & Health Services, (Zone) Jaipur, Ajmer, Udaipur, Jodhpur, Bikaner, Kota

&Bharatpur.

16. All P.M.O.'s.
17. All CM&HO's.
18. All Dy. CM&HO's.
19. AGM (IT), RMSC, Jaipur.
20. All Zonal BME are directed to carry out inspection of the supplied equipment as per RMSCL Letter no. 2513 dated 10-07-2018.
21. M/s Clarity Medical Pvt. Ltd., Plot No. 1687-A, JLPL Industrial Park, Sector-82, Mohali-140306 (PB) India Email:-contact@clarity-medical.com
22. Guard file.

  
Executive Director (EPM)  
RMSC, Jaipur