

Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065

E-mail : edepmrmc-rj@nic.in

CIN:U24232RJ2011SGC035067

Website: www.rmhc.health.rajasthan.gov.in

F-8(317)RMSC/EPM/M-3/2017-18/NIB-272/2091

Dated:- 4-05-18


Rate Contract

S.N.	Item	Detail
1.	Name of Equipment / Item	Wheel Chair
2.	Name & Address of Firm	M/s A.K. Surgicals, 25/11, 1st Floor, Kaveri Path, New Sanganer Road, Mansarovar, Jaipur-302 020 (Raj.) Email:-aksurgicals@yahoo.co.in
3.	Brand/ Model / Make	AKS-3015
4.	HSN Code	
5.	Packing Unit	Each
6.	R/C No. & Validity	(R.C. No-475) valid up to 31.05.2020
7.	Net Rate Per packing Unit (Rs.)	8888.00
8.	Taxes	GST As applicable
9.	RMSC Surcharges @ 5%	As applicable
10.	Rates for Supply to RMSC from Suppliers	
	I. New Basic Rate (In Rs.)	8888.00
	II. IGST (In Rs.) 0 %	0.00
	III. CGST @ 2.5% (In Rs.)	222.20
	IV. SGST @ 2.5% (In Rs.)	222.20
	V. Approved Rate For Supplier To RMSC {(i)+(ii)+(iii)+(iv) (In Rs.)	9332.40
11.	Rate For Supply to Consignee From RMSC	
	I. New Basic Rate (In Rs.)	8888.00
	II. RMSC Surcharge @ 5 % On (i) (In Rs.)	444.40
	III. CGST @ 2.5% { On (i)+(ii) } (In Rs.)	233.31
	IV. SGST @ 2.5% { On (i)+(ii) } (In Rs.)	233.31
	V. Approved Rate For Consignee from RMSC {(i)+(ii)+(iii)+(iv) (In Rs.)	9799.02
	Technical Specification of Wheel Chair	
	1. The wheel chair shall be made up of 16 gauge Stainless steel 304 grade tube frames and 16 gauge stainless steel 304 sheet for seat and backrest.	
	2. The dimension of wheel chair: 670mm (w) x 1120mm (d) x 92mm (h)±5mm.	
	3. It should have a fixed arm set with soft polyurethane.	
	4. It should have reticulated and breathable cushion and seat.	

5. It should have retractable aluminium foot rest.
6. It should have push handle with plastic cap.
7. It should have 6" swivel nylon castor front wheel, 24" bicycle type rear wheel with puncture proof flat free type tyre.
8. The back wheel fixing bolt shall be covered with cup type nut.
9. It should have wheel breaking system on both side.

Other terms and conditions:

1. Manufacturer should be ISO- 9001 certified.
2. **Guarantee:** should be one year for all items from the date of successfully installation at consignees address.
3. Final Technical approval after demonstration.
4. Installation will be done by supplier free of cost.
5. The Company should submit technical compliance sheet as per technical specifications mentioning the make and model of quoted item along with catalogue in technical bid.



Executive Director (EPM)
RMSC, Jaipur

F-8(317)RMSC/EPM/M-3/2017-18/NIB-272/

Dated:-

Copy forwarded to the following for information & necessary action:

1. P.S. to Mission Director NHM Jaipur.
2. Director, (PH)/(Aids)/(FW)/(RCH)/(IEC)/(ESI).
3. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar & Udaipur.
4. Superintendent Jaipuriya Hospital Jaipur (Associated with RUHS College of Medical Sciences, Jaipur).
5. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
6. All P.M.O.'s.
7. All CM&HO's.
8. Executive Director (Logistics/Finance).
9. AGM (IT), RMSC, Jaipur.
10. All DPC's DDW Rajasthan.
11. M/s A.K. Surgicals, 25/11, 1st Floor, Kaveri Path, New Sanganer Road, Mansarovar, Jaipur-302 020
(Raj.)Email:-aksurgicals@yahoo.co.in
12. Guard file.


Executive Director (EPM)
RMSC, Jaipur