



मुख्यमं

Rajasthan Medical Services Corporation Limited, Jaipur

'D'-block, swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur

Phone no. 0141-2223887, Fax no. 0141-2228065

CIN : U24232RJ2011SGC035067

E-mail: rmsc@nic.in

Website : www.rmsc.nic.in

No. F-8() RMSC/EPM/M-1/ RC 17-18/NIB-324/ Boyals/ 2018/ 3642 Dated: 21-08-18

All Principal, Medical Colleges,
All the Superintendent of Medical College & Attached Hospitals.
Director (PH/HA/IEC/AIDS/RCH)

Sub:- Approval/Purchase of Item, 'Boyls Apparatus with vaporizer and safety device', for the Rate Contract and supply, Ending on 31.08.2020.

Please find enclosed herewith approved firm and rates etc. in respect of Item '**Boyls Apparatus with vaporizer and safety device**', for the Rate Contract and supply, ending on 31.08.2020. The Rate Contract and its terms & conditions are also available on Rajasthan Medical Services Corporation Ltd., Website www.rmsc.nic.in.

In reference to above Rate Contract No. No. F-8() RMSC/EPM/M-1/ RC 17-18/NIB-324/ Boyals/ 2018/ Dated:..... the supplier has executed the agreement with RMSC, as enclose here for your information & necessary action. This bears an approval of M.D., RMSCL, Jaipur.


Executive Director (EPM)
Rajasthan Medical Services Corporation
Ltd, Jaipur

No. F-8() RMSC/EPM/M-1/ RC 17-18/NIB-324/ Boyals 2018/ 3642

Dated: 21-08-18

Copy forwarded to the following for information & necessary action:

1. P.S. to Hon'ble Health Minister, Medical & Health Deptt. Govt. of Rajasthan.
2. P.S. to PHS, Medical Health & FW Deptt. Rajasthan, Jaipur.
3. P.S. to Mission Director NHM Jaipur.
4. Director, (PH)/(Aids)/(FW)/(RCH)/(IEC)/(ESI).
5. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar & Udaipur.
6. Principal R.U.H.S. College of Dental Science Jaipur.
7. Principal, RUHS College of Medical Scinces, Jaipur.
8. Superintendent Japuriya Hospital Jaipur (Associated with RUHS College of Medical Scinces, Jaipur).
9. Superintendent, SMS Hospital, Zenana Hospital, Mahila Chikitsalya, Mental Hospital, J.K. Lon Hospital, T.B & Chest Hospital, Jaipur.
10. Superintendent Associated Group of Hospital-Ajmer, M.G. Hospital, Jodhpur, Ummaid Hospital-Jodhpur, PBM Hospital-Bikaner, MBS Hospital-Kota & A.G. Hospital-Udaipur.
11. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
12. Addl. Director, Medical & Health Services (PW)/(RH)/(Hospital Admn.)
13. Joint Director, Medical & Health Services, (Zone) Jaipur, Ajmer, Udaipur, Jodhpur, Bikaner, Kota & Bharatpur.
14. All P.M.O.'s.
15. All CM&HO's.
16. All Dy. CM&HO's.



17. AGM (IT), RMSC, Jaipur
18. M/s Allied Medical Limited, New Delhi
19. Guard file

[Handwritten signature]
[Handwritten date: 21.8.18]

Executive Director (EPM)
Rajasthan Medical Services Corporation
Ltd, Jaipur



Rajasthan Medical Services Corporation Limited, Jaipur
Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065

CIN:U24232RJ2011SGC035067

E-mail : edepmrmsc-ri@nic.in

Website: <http://rmsc.health.rajabsthan.gov.in>

No. F-8() RMSC/EPM/M-1/ RC 17-18/NIB-324/ Boyals/ 2018/ 3643

Dated: 21-08-18


Rate Contract and Price Details

| S.N. | Item | Detail |
|------|--|--|
| 1 | Name of Equipment / Item | BOYALS APPARATUS WITH VAPORIZER AND SAFETY DEVICE |
| 2 | Name & Address of Firm | M/s Allied Medical Limited 203, Swastik Bhawan, Ranjeet Nagar Commercial Complex New Delhi-110008 Email: sales@alliedmed.co.in |
| 3 | Brand/ Model / Make | Allied Medical Ltd. |
| 4 | Packing Unit | Each |
| 5 | R/C No. & Validity | R.C. No. 519 (Valid Upto 31.08.2020) |
| 6 | | Rates for supply to RMSC from Suppliers |
| | (i) New Basic Rate (in Rs.) | 177777.00 |
| | (ii) IGST @ 12% (in Rs.) | 21333.24 |
| | (iii) CGST (in Rs.) | 0.00 |
| | (iv) SGST (in Rs.) | 0.00 |
| | (v) Approved Rate From Supplier To RMSC {(i)+(ii)+(iii)+(iv)} (in Rs.) | 199110.24 |
| 7 | | Rate For Supply to Consignee From RMSC |
| | (i) New Basic Rate (in Rs.) | 177777.00 |
| | (ii) RMSC Surcharge@5% On (i) (in Rs.) | 8889.00 |
| | (iii) CGST@ 6 % { On (i)+(ii)} (in Rs.) | 11200.00 |
| | (iv) SGST@ 6 % { On (i)+(ii)} (in Rs.) | 11200.00 |
| | (v) Approved Rate For Consignee From RMSC {(i)+(ii)+(iii)+(iv)} (in Rs.) | 209066.00 |
| 8 | | <u>Technical Specifications of Boyals Apparatus with vaporizer and safety device</u> |
| | 1. The Machine shall be suitable to be used for adult and paediatric patients. | |
| | 2. It should be of perfect workmanship with rigid and robust structure of epoxy powder-coated or stainless steel material. | |

AM



3. It should have rotameter with color-coded knobs for Oxygen and Nitrous Oxide.
4. It should have a back bar which is ISO-pin type to attach vaporizer easily.
5. It should have Close circle system with CO2 absorbent canister should be mounted on the machine. There should be facility to bypass closed system for changing soda lime while the machine is in use. There should be quick change over from open circuit to close circuit.
6. Should have NIST Inlets for piped medical gases for O₂, and N₂O.
7. International quality tubing's and leak proof connections in place of conventional copper tubes for easy serviceability and maintenance. Internal pipe work should be color coded as per ISO Standards.
8. Should have color coded, Gas specific, Pin Indexed yokes with sliding stainless steel clamping bars for Oxygen min. 2 Nos. and Nitrous Oxide min. 2 Nos. and **CENTRAL PIPELINE INLET FOR O₂ & N₂O.**
9. Semi-open, semi-closed or closed circuit system.
10. Oxygen Flush at table level to give 35-75Lt/min.
11. Should have pressure relief valve, with auto reset feature, non return valves & oxygen flush.
12. Should have large roller bearing anti-static castors with front brakes & foot rest.
13. Stainless steel table top for writing, keeping syringes, drugs, instruments etc.
14. Wide Monitor shelf at eye level with securing provision for extra monitor.
15. Should have twin selectatec manifold made out of single piece design without connections to avoid leakages for mounting at least two agent specific vaporizers with interlocking facility.
16. Should be supplied with agent specific vaporizers for **Halothane and Isoflourane** (Calibration certificate should be submitted at the time of Installation for both vaporizers)
17. **Supplier shall be responsible for Vaporizers Calibration.**
18. Should be supplied with necessary attachments for use of breathing circuit like Bains, Jackson-Rees, mgills etc.
19. **Machine should be CE Certified/ ISI MARK.**
20. **Vaporizer should be EU CE certified.**
21. **Manufacturer should be ISO 9001/ISO 13485 certified.**
22. **Safety Features / Mechanisms:**
 - a) Hypoxic guard to ensure min 25% oxygen supply at all O₂-N₂O mixture supply flow.
 - b) Audible oxygen failure warning system with N₂O cut off.
 - c) Self activated electro-pneumatic audio visual OFWD to activate an alarm when supply pressure of oxygen falls to 150 kPa-250 kPa
23. **Guarantee:** Two years on equipment from the date of installation.
24. **CMC:** CMC shall be given @ 5 % of net rate (inclusive of Excise Duty & exclusive of VAT/CST etc.) plus service tax (as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after



Guarantee period of two years.

25. Autoclavable anesthetic circuits in all sizes for adult & pediatric (one each), complete set in all sizes of anesthetic face mask for adult & pediatric (one each) and operational/training manual (1 No.) as per requirement of the equipment should be provided along with equipment at the time of installation / commissioning.
26. Installation will be done by supplier free of cost.
27. Firm should mention all the pre-installation requirements in technical bid.
28. Demonstration of machine is must for final approval.
29. The service engineer should be based in Rajasthan.
30. Maximum Permissible time for breakdown call services is 48 hours and call may be logged through phone, Fax, e-mail, or letter.
31. During two years of Guarantee period, for each year, minimum three preventive visits for repair, maintenance and services are mandatory.
32. The company should mention the make & model name/number of the quoted equipment and submit the technical brochure of the quoted model in the Technical bid along with compliance sheet as per technical specifications.

Executive Director (EPM)
RMSC, Jaipur

Dated:

No. F-8() RMSC/EPM/M-1/ RC 17-18/NIB-324/ Boyals/ 2018/

Copy forwarded to the following for information & necessary action:

1. P.S. to Mission Director NHM Jaipur.
2. Director, (PH)/(Aids)/(FW)/(RCH)/(IEC)/(ESI).
3. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar & Udaipur.
4. Superintendent Jaipuria Hospital Jaipur (Associated with RUHS College of Medical Sciences, Jaipur).
5. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
6. All P.M.O.'s.
7. All CM&HO's.
8. Executive Director (Logistics/Finance)
9. AGM (IT), RMSC, Jaipur
10. All DPC's DDW Rajasthan.
11. BME (R&M), RMSC, Jaipur.
12. M/s Allied Medical Limited, New Delhi
13. Guard file.

Executive Director (EPM)
RMSC, Jaipur