



Rajasthan Medical Services Corporation Ltd.

Gandhi Block, Swasthaya Bhawan, Tilak Marg, Jaipur

Ph. No. 0141-2223887, Fax No. 0141-2228065

E-Mail: rmsc@nic.in



F-8 (334) RMSC/EPM/M-3/2017-18/NIB-321/3624 dated 16.08.18

All Principal, Medical Colleges,
All the Superintendent of Medical College & Attached Hospitals.
Director (PH/HA/IEC/AIDS/RCH)

Sub:- Approval/Purchase of Item, 'BUBBLE CPAP (CONTINUOUS POSITIVE AIRWAY PRESSURE) MACHINE', for the Rate Contract and supply, Ending on 31.08.2020.

Please find enclosed herewith approved firm and rates etc. in respect of Item 'BUBBLE CPAP (CONTINUOUS POSITIVE AIRWAY PRESSURE) MACHINE', for the Rate Contract and supply, ending on 31.08.2020. The Rate Contract and its terms & conditions are also available on Rajasthan Medical Services Corporation Ltd., Website www.rmsc.nic.in.

In reference to above Rate Contract F-8(334)RMSC/EPM/M-3/2017-18/NIB-321/..... Dated: the supplier has executed the agreement with RMSC, as enclose here for your information & necessary action. This bears an approval of M.D., RMSCL, Jaipur.

[Handwritten Signature]
16.8.2018

Executive Director (EPM)
Rajasthan Medical Services Corporation
Ltd, Jaipur

F-8 (334) RMSC/EPM/M-3/2017-18/NIB-321/3624 dated 16.08.18

Copy forwarded to the following for information & necessary action:

1. PS to Hon'ble Health Minister, Medical & Health Deptt. Govt. of Rajasthan.
2. PS to PHS, Medical Health & FW Deptt. Rajasthan, Jaipur.
3. PA to Director, (PH)/(Aids)/(FW)/(RCH)/(IEC).
4. Project Director, NRHM.
5. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar & Udaipur.
6. Superintendent, SMS Hospital, Zanana Hospital, Mahila Chikitsalya, Mental Hospital, J.K. Lon Hospital, T.B & Chest Hospital, Jaipur.
7. Superintendent Associated Group of Hospital-Ajmer, M.G. Hospital, Jodhpur, Ummaid Hospital-Jodhpur, PBM Hospital-Bikaner, MBS Hospital-Kota & A.G. Hospital-Udaipur.
8. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
9. Addl. Director, Medical & Health Services (PW)/(RH)/(Hospital Admn.)
10. Joint Director, Medical & Health Services, (Zone) Jaipur, Ajmer, Udaipur, Jodhpur, Bikaner, Kota & Bharatpur.
11. All Dy. CM&HO's.
12. AGM (IT), RMSC, Jaipur.
13. M/s Fanem Medical Devices India Pvt. Ltd., #279/A, 4th main, Peenya Industrial Area, Bengluru- 560058 Email: - info@fanem.in
14. Guard file.

Received
P.K. Wal

[Handwritten Signature]
16.8.2018

Executive Director (EPM)
Rajasthan Medical Services Corporation
Ltd, Jaipur



Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065

E-mail : edepmrmisc-rj@nic.in


CIN:U24232RJ2011SGC035067

Website: www.rmsc.health.rajasthan.gov.in

F-8 (334) RMSC/EPM/M-3/2017-18/NIB-321/325 dated: 16-08-18

Rate Contract

S.N.	Item	Detail
1.	Name of Equipment / Item	BUBBLE CPAP (CONTINUOUS POSITIVE AIRWAY PRESSURE) MACHINE
2.	Name & Address of Firm	M/s Fanem Medical Devices India Pvt. Ltd., #279/A, 4th main, Peenya Industrial Area, Bengluru- 560058, Email: - info@fanem.in
3.	Brand/ Model / Make	BABY PAP-1150S
4.	HSN Code	
5.	Packing Unit	Each
6.	R/C No. & Validity	(R.C. No-518) valid up to 31.08.2020
7.	Net Rate Per packing Unit (Rs.)	188786.00
8.	Taxes	GST As applicable
9.	RMSC Surcharges @ 5%	As applicable
10.	Rates for Supply to RMSC from Suppliers	
	I. New Basic Rate (In Rs.)	188786.00
	II. IGST (In Rs.) 12 %	22654.32
	III. CGST @% (In Rs.)	0.00
	IV. SGST @% (In Rs.)	0.00
	V. Approved Rate For Supplier To RMSC {(i)+(ii)+(iii)+(iv)} (In Rs.)	211440.32
11.	Rate For Supply to Consignee From RMSC	
	I. New Basic Rate (In Rs.)	188786.00
	II. RMSC Surcharge @ 5 % On (i) (In Rs.)	9439.30
	III. CGST@ 6% { On (i)+(ii)} (In Rs.)	11893.52
	IV. SGST@ 6% { On (i)+(ii)} (In Rs.)	11893.52
	V. Approved Rate For Consignee from RMSC {(i)+(ii)+(iii)+(iv)} (In Rs.)	222012.34
	Technical Specifications of BUBBLE CPAP Machine	
	1. Function: Continuous Positive Airway Pressure (CPAP) system to treat premature / new born babies with respiratory distress. Non-invasive method of delivering CPAP. Flow of Air and blender offering precise gas blending of flow 0~15 lpm.	
	2. Operational Requirements: CPAP generator where PEEP should be set by water column which can be adjusted externally. The device should include a heated humidifier, patient circuit, air / oxygen blender and patient interface of various sizes.	
	3. Humidifier: 1. Should be Servo Controlled 2. Display of Airway Temperature. 3. Should have heater wire adapter, Airway Temperature probe. 4. Have variable temperature depending on requirement. 5. Audio/visual alarms present on screen.	
	4. Humidifier chamber: Should be supplied disposable type auto feed valve mechanism for constant compressible volume to maintain CPAP pressure.	

**5. Patient Circuit and accessories:**

1. Should be supplied with **reusable** heater wire breathing circuit and, infant Bonnets. Each Circuit must have disposable Nasal Prongs, Headgear.
2. Pressure limiting valve (10 CM H₂O) to prevent damage from occlusions.
3. PEEP generator with easy direct setting with CPAP from 2~12CM H₂O.

6. Compressor and Stand:

1. Medical Grade Compressor having low noise level and low vibration on sturdy lockable wheels.
2. Should have pressure gauge, electronic dryer and pressure regulator.

7. Air Oxygen Blender:

1. Blender should provide precise mixing of air and oxygen with adjustable knob for FiO₂ from 21 to 100 %.
2. Outlets - Outlet 1-15 lpm (supply pressure at 50 PSI)
3. Operating temperature range 15 to 25 degree Celsius
4. FiO₂ accuracy : 2~ 3% of full scale

8. Safety:

1. The blender, humidifier should be European CE(By Notified Body) or US FDA approved and the Compressor should be CE certified. (Valid certificate should be submitted in technical bid).
2. Recommended ambient temperature range 15 to 35 degrees Celsius.
3. Power Supply 230 volts, Supply frequency 50 or 60 Hz.

9. Scope of Supply to include with EACH machine:

1. It should be supplied with Air and oxygen hose kit and pole mounting clamp - 1
2. Noise free Compressor on wheels - 1
3. Patient Support Arm - 1
4. Servo Controlled Humidifier -1
5. One set of reusable chamber with heater wire- 1
6. Patient interface one set - 1
7. Disposable Head band 3 sizes from 0.5 kg to 4kg - 5set
8. Disposable Nasal Prongs 5 sizes from 0.5 kg to 4kg - 10set

10. Guarantee three years on complete unit including compressor, Blender and Humidifier after installation.

11. CMC: CMC will be given @ 5 % plus GST (as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after completion of Guarantee period of three years.

12. Four preventive maintenance services are essential annually during guarantee and CMC period.

13. Installation will be done by the firm free of cost.

14. Service Network of Firm in Rajasthan is essential.

15. Demonstration of the Equipment may be taken if needed for final technical approval.

16. The company should submit technical compliance sheet along with make / model of the quoted item along with catalogue in technical bid.

Executive Director (EPM)
RMSC, Jaipur

F-8 (334) RMSC/EPM/M-3/2017-18/NIB-321/

dated:

Copy forwarded to the following for information & necessary action:

1. P.S. to Mission Director NHM Jaipur.
2. Director, (PH)/(Aids)/(FW)/(RCH)/(IEC)/(ESI).
3. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar & Udaipur.
4. Superintendent Jaipuriya Hospital Jaipur (Associated with RUHS College of Medical Sciences, Jaipur).
5. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
6. All P.M.O.'s.
7. All CM&HO's.
8. Executive Director (Logistics/Finance).
9. AGM (IT), RMSC, Jaipur.
10. All DPC's DDW Rajasthan.
11. M/s Fanem Medical Devices India Pvt. Ltd., #279/A, 4th main, Peenya Industrial Area, Bengluru- 560058 Email: - info@fanem.in
12. Guard file.

Executive Director (EPM)
RMSC, Jaipur