



# Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065

E-mail : edepmrmsc-rj@nic.in

CIN:U24232RJ2011SGC035067

Website: www.rmsc.health.rajasthan.gov.in

F-8(322)RMSC/EPM/M-3/2017-18/NIB-285/ 3096

Dated:- 3-07-18

## Rate Contract

S.N.	Item	Detail
1.	Name of Equipment / Item	Oxygen Concentrator
2.	Name & Address of Firm	M/s Moditech International, #6694, Madanpura, Phase-2 near Madanpura Chowk, S.A.S. Nagar, Mohali-160055, Punjab Email : moditech1@gmail.com
3.	Brand/ Model / Make	JAY5
4.	HSN Code	
5.	Packing Unit	Each
6.	R/C No. & Validity	(R.C. No-498 ) valid up to 31.07.2020
7.	Net Rate Per packing Unit (Rs.)	26600.00
8.	Taxes	GST As applicable
9.	RMSC Surcharges @ 5%	As applicable
10.	<b>Rates for Supply to RMSC from Suppliers</b>	
	I. New Basic Rate (In Rs.)	26600.00
	II. IGST (In Rs.) 12 %	3192.00
	III. ) CGST @% (In Rs.)	0.00
	IV. SGST @% (In Rs.)	0.00
	V. Approved Rate For Supplier To RMSC {(i)+(ii)+(iii)+(iv) (In Rs.)	29792.00
11.	<b>Rate For Supply to Consignee From RMSC</b>	
	I. New Basic Rate (In Rs.)	26600.00
	II. RMSC Surcharge @ 5 % On (i) (In Rs.)	1330.00
	III. CGST@ 6% { On (i)+(ii)} (In Rs.)	1675.80
	IV. SGST@ 6% { On (i)+(ii)} (In Rs.)	1675.80
	V. Approved Rate For Consignee from RMSC {(i)+(ii)+(iii)+(iv) (In Rs.)	31282.00
	<b>Technical Specifications of Oxygen Concentrator</b>	
	1. Oxygen concentrator to provide oxygen from ambient air.	
	2. Oxygen concentration measured at the flow meter by oxygen sensing device (OSD).	
	3. Sound level should be < 50dB.	
	4. Superior grade of molecular sieve.	
	5. Maintenance free rotary proppet valve.	
	6. Oxygen purity should be 92% ± 2%.	
	7. Oxygen output, approx: 0 - 5 LPM.	
	8. Pressure range should be 6 - 8 PSI.	



9. Single outlet to maintain output pressure & flow/ Double outlet or flow splitter for oxygen Delivery.
10. Oxygen tube of 2m length must be provided with two humidifier bottles and two cabinet filters.
11. Power requirements: 220 V/50 Hz.
12. Device is produced by ISO 9001 certified manufacturer.
13. Device should be certified ISO 80601-2-69:2014.
14. Bottles should be made of autoclavable polycarbonate.
15. With bilateral oxygen nozzle.
16. Oxygen failure or Low Oxygen Alarm facility should be available for below 80% oxygen.
17. Device is safety certified according EUROPEAN CE / USFDA.
18. **Guarantee:** Two years on equipment from the date of installation.
19. **CMC:** CMC will be given @ 5 % exclusive GST (as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after guarantee period of two years.
20. **Supplied with:-**
  - 1) 1 x spare set of tubing.
  - 2) Two spare sets of humidifier bottles.
  - 3) Two spare sets of cabinet filters.
  - 4) User manual with trouble shooting guidance, in English.
  - 5) Technical manual with maintenance and first line technical intervention instructions, in English.
- 22) Rates of accessories like Reusable oxygen tubing, Oxygen Mask for Adult, Pediatric & neonatal should be provided by bidder in BOQ in e-proc. Portal.
- 23) The service engineer should be based in Rajasthan.
- 24) Demonstration of equipment is must for technical approval.
- 25) Installation will be done by supplier free of cost.
- 26) The company should mention the make & model of the quoted item and submit the technical brochure of the quoted model in the Technical bid along with compliance sheet as per technical specifications.

Executive Director (EPM)  
RMSC, Jaipur

F-8(322)RMSC/EPM/M-3/2017-18/NIB-285/

Dated:-

Copy forwarded to the following for information & necessary action:

1. P.S. to Mission Director NHM Jaipur.
2. Director, (PH)/(Aids)/(FW)/(RCH)/(IEC)/(ESI).
3. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar & Udaipur.
4. Superintendent Jaipuria Hospital Jaipur (Associated with RUHS College of Medical Sciences, Jaipur).
5. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
6. All P.M.O.'s.
7. All CM&HO's.
8. Executive Director (Logistics/Finance).
9. AGM (IT), RMSC, Jaipur.
10. All DPC's DDW Rajasthan.
11. M/s Moditech International, #6694, Madanpura, Phase-2 near Madanpura Chowk, S.A.S. Nagar, Mohali-160055, Punjab, Email : moditech1@gmail.com
12. Guard file.

Executive Director (EPM)  
RMSC, Jaipur