

राजस्थान चिकित्सा सेवा निगम लिमिटेड

गाँधी ब्लॉक, स्वास्थ्य भवन, तिलक मार्ग, सी-स्कीम, जयपुर - 302005 (राज.)

फोन नं: 0141-2228066, 2228064

फैक्स नं 0141-2228065

वेबसाइट : www.rmsh.health.rajasthan.gov.in

सीआईएन. : U24232RJ2011SGC035067

ई-मेल :: edpssrmsh@rajasthan.gov.in

क्रमांक : F.02(123)/RMSCL/Proc/S&S (Non Drug)/NIB-09/2022/

3985

Dated:

29/08/2024

कार्यालय आदेश

सर्जिकल (नॉन-ड्रग आईटम्स) की आपूर्ति एवं दर संविदा के लिये जारी की गई निविदा सूचना क्रमांक: F.02(123)/RMSCL/Proc/S&S (Non Drug)/NIB-09/2022/652 Dated: 29.08.2022 के अन्तर्गत अनुमोदित निम्न आईटम की दर संविदा में कॉलम संख्या 4 के अनुसार दर संविदा समाप्ति दिनांक से 3 माह (कॉलम संख्या 5 में अंकित तिथि तक) या नवीन निविदा (एग्जीमेंट होने तक) की दिनांक तक, जो भी पहले हो तक की वृद्धि की जाती है:-

S. No.	Item code	Item name	Rate Contract Validity	Rate contract Extended up to	Name of Firm
1	2	3	4	5	6
1.	S-5(a)	Disposable Sterile Surgical Rubber Gloves Size 6 ½ Inches <ul style="list-style-type: none"> Made of natural rubber Latex, powdered, without tear, properly folded in a paper Free of holes, with Acceptable Quality Level (AQL) of 1.5 or less Tensile strength as per EN 455-2 Powder should be non-allergenic As per IS 13422 ISI marked / CE certified / FDA approved 	30.09.2024	31.12.2024	Swear Healthcare Pvt Ltd Prince Enterprise
2.	S-5(b)	Disposable Sterile Surgical Rubber Gloves Size 6 ½ Inches <ul style="list-style-type: none"> Made of natural rubber Latex, powder free (Polymer /Silicon Coated), without tear, properly folded in a paper Free of holes, with Acceptable Quality Level (AQL) of 1.5 or less Tensile strength as per EN 455-2 As per IS 13422 ISI marked / CE certified / FDA approved (CE Certification / FDA Approval for item is mandatory for importer firms, that cannot avail IS Standards) 	30.09.2024	31.12.2024	Prince Enterprise Swear Healthcare Pvt Ltd
3.	S-6(a)	Disposable Sterile Surgical Rubber Gloves Size 7 Inches <ul style="list-style-type: none"> Made of natural rubber Latex, powdered, without tear, properly folded in a paper Free of holes, with Acceptable Quality Level (AQL) of 1.5 or less 	30.09.2024	31.12.2024	Stoe Exim Prince Enterprise Swear Healthcare Pvt Ltd

Signature valid



Digitally signed by Harish Kumar Lalwani
 Designation : Executive Director Procurement
 Date: 2024.09.24 10:38:54 IST
 Reason: Approved

S. No.	Item code	Item name	Rate Contract Validity	Rate contract Extended up to	Name of Firm
1	2	3	4	5	6
		<ul style="list-style-type: none"> Tensile strength as per EN 455-2 Powder should be non-allergenic As per IS 13422 ISI marked / CE certified / FDA approved 			
4.	S-6(b)	Disposable Sterile Surgical Rubber Gloves Size 7 Inches <ul style="list-style-type: none"> Made of natural rubber Latex, powder free (Polymer /Silicon Coated), without tear, properly folded in a paper Free of holes, with Acceptable Quality Level (AQL) of 1.5 or less Tensile strength as per EN 455-2 As per IS 13422 ISI marked / CE certified / FDA approved (CE Certification / FDA Approval for item is mandatory for importer firms, that cannot avail IS Standards)	30.09.2024	31.12.2024	Prince Enterprise Swear Healthcare Pvt Ltd
5.	S-7(a)	Disposable Sterile Surgical Rubber Gloves Size 7½ Inches <ul style="list-style-type: none"> Made of natural rubber Latex, powdered, without tear, properly folded in a paper Free of holes, with Acceptable Quality Level (AQL) of 1.5 or less Tensile strength as per EN 455-2 Powder should be non-allergenic As per IS 13422 ISI marked / CE certified / FDA approved 	30.09.2024	31.12.2024	Stoe Exim Prince Enterprise Swear Healthcare Pvt Ltd
6.	S-7(b)	Disposable Sterile Surgical Rubber Gloves Size 7 ¼ Inches <ul style="list-style-type: none"> Made of natural rubber Latex, powder free (Polymer /Silicon Coated), without tear, properly folded in a paper Free of holes, with Acceptable Quality Level (AQL) of 1.5 or less Tensile strength as per EN 455-2 As per IS 13422 ISI marked / CE certified / FDA approved (CE Certification / FDA Approval for item is mandatory for importer firms, that cannot avail IS Standards)	30.09.2024	31.12.2024	Prince Enterprise Swear Healthcare Pvt Ltd

नोट:- दिनांक 09.09.2020 को हुयी बोर्ड मीटिंग में लिये गये निर्णय "In case of extension of rate contract, if new tender has been invited in such case, it shall be

Signature valid

Digitally signed by Harsh Kumar Lalwani
Designation : Procurement Director
Procurement
Date: 2024.09.24 10:38:54 IST
Reason: Approved

mentioned in extension agreement that if rate received in new bid of same firm is lower than previous approved rate, then payment to the previous L1 firm shall be made as per price fall clause.” के अनुसार इस अवधि में सप्लाय किए गए सर्जिकल एवं सूचर्स का भुगतान नवीन निविदा में प्राप्त दरों एवं पुरानी दरों में जो भी कम हो के अनुसार किया जावेगा।

कार्यकारी निदेशक (प्रोक्यू.)
आरएमएससीएल

क्रमांक : F.02(123)/RMSCL/Proc/S&S (Non Drug)/NIB-09/2022/ 3925 Dated: 24/9/2024

प्रतिलिपि:- निम्न को सूचनार्थ एवं आवश्यक कार्यवाही हेतु प्रेषित है:-

1. विशेषाधिकारी, आरएमएससी को प्रेषित कर निवेदन है कि क्रयादेश देने से पूर्व विद्यमान बैंक गारण्टी की वैधता अवधि की सुनिश्चितता तथा मूल दर संविदा से मिलान करा लेवें।
2. कार्यकारी निदेशक, (वित्त/क्यू.सी) आरएमएससी।
3. अतिरिक्त निदेशक, आईटी, आरएमएससी
4. संबंधित फर्म मैसर्स

कार्यकारी निदेशक (प्रोक्यू.)
आरएमएससीएल

Signature valid

Digitally signed by Harish Kumar
Lalwani
Designation : Executive Director
Procurement
Date: 2024.09.24 10:38:54 IST
Reason: Approved