

Rajasthan Medical Services Corporation, Swasthaya Bhawan, C-Scheme, Jaipur

Phone No: 0141-2228065, Fax No: 0141-2228065

E_mail : rpsc@nic.in

CIN: U24232RJ2011SGC035067

Website: WWW.rpsc.health.rajasthan.gov.in

F.02(40)/RMSC/Proc./Surgical (Non Drug)/NIB-01/2017/ 84

Dated: 03/03/2017

Corrigendum

Subject: - Amendment to Bid for the rate contract cum supply and empanelment of surgical (non - drug items)

Ref: - F.02(40)/RMSC/Proc./Surgical (Non Drug)/NIB-01/2017/76 Dated: 17.02.2017

Amendment in Specification of Surgical items given in the bid and BOQ are to be read as follows:-

S. No	Item Code	Existing specification	Amended specification
1	S-5(b)	Disposable Sterile Surgical Rubber Gloves Size 6 ½ Inches <ul style="list-style-type: none"> Made of natural rubber Latex, powdered free, without tear, properly folded in a paper Free of holes, with Acceptable Quality Level (AQL) of 1.5 or less Tensile strength as per EN 455-2 should Conform to IS 13422 ISI marked / CE certified / FDA approved 	Disposable Sterile Surgical Rubber Gloves Size 6 ½ Inches <ul style="list-style-type: none"> Made of natural rubber Latex, powder free (Polymer /Silicon Coated), without tear, properly folded in a paper Free of holes, with Acceptable Quality Level (AQL) of 1.5 or less Tensile strength as per EN 455-2 should Conform to IS 13422 ISI marked / CE certified / FDA approved (CE Certification / FDA Approval for item is mandatory for importer firms, that cannot avail IS Standards)
2	S-6(b)	Disposable Sterile Surgical Rubber Gloves Size 7 Inches <ul style="list-style-type: none"> Made of natural rubber Latex, powdered Free , without tear, properly folded in a paper Free of holes, with Acceptable Quality Level (AQL) of 1.5 or less Tensile strength as per EN 455-2 should Conform to IS 13422 ISI marked / CE certified / FDA approved 	Disposable Sterile Surgical Rubber Gloves Size 7 Inches <ul style="list-style-type: none"> Made of natural rubber Latex, powder free (Polymer /Silicon Coated) without tear, properly folded in a paper Free of holes, with Acceptable Quality Level (AQL) of 1.5 or less Tensile strength as per EN 455-2 should Conform to IS 13422 ISI marked / CE certified / FDA approved (CE Certification / FDA Approval for item is mandatory for importer firms, that cannot avail IS Standards)

(Handwritten signatures and marks)

मुख्यमंत्री निःशुल्क दवा योजना



	<p>S-7(b) Disposable Sterile Surgical Rubber Gloves Size 7½ Inches</p> <ul style="list-style-type: none">• Made of natural rubber Latex, powdered Free, without tear, properly folded in a paper• Free of holes, with Acceptable Quality Level (AQL) of 1.5 or less• Tensile strength as per EN 455-2• should Conform to IS 13422• ISI marked / CE certified / FDA approved	<p>Disposable Sterile Surgical Rubber Gloves Size 7½ Inches</p> <ul style="list-style-type: none">• Made of natural rubber Latex, powder free (Polymer /Silicon Coated), without tear, properly folded in a paper• Free of holes, with Acceptable Quality Level (AQL) of 1.5 or less• Tensile strength as per EN 455-2• should Conform to IS 13422• ISI marked / CE certified / FDA approved (CE Certification / FDA Approval for item is mandatory for importer firms, that cannot avail IS Standards)
--	--	--

Rest terms and conditions will remain the same.

E.D. (Proc.)