

# Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065  
CIN:U24232RJ2011SGC035067

E-mail : edepmrmisc-rj@nic.in  
Website: www.rmssc.nic.in

No. F-8(RMISC/EPM/M-2/2015-16/NIB-145/ 1032

Dated: 4/3/16

## CLARIFICATION/CORRIGENDUM/ADDENDUM

**Sub:- Revised Technical Specifications of item LED OT Light double dome and revised bid submission date under NIB No. F-8( )RMISC/EPM/M-2/2015-16/NIB-145/416 dated 02.02.16**

In Reference to subject cited above and NIB-145, the various representations received from the firms and issues raised by the Bidders are examined by the competent Authorities and technical committee. The following Clarification/Corrigendum/Addendum is issued for inclusion in Bid document & Technical Specification of items as below:-

### **A. Amended Technical Specifications of item LED OT Light double dome:-**

1. Surgical light unit with Electronics control, for homogeneous and shadow less light with minimum temperature rise.
2. The highly counter balanced mechanism to fix the dome at any position, easy maneuverable free up & down & swing movements & freely rotatable at any angle.
3. Light intensity & focusing should be adjustable from control panel available on dome. Should provide LED technology. Focusing should be adjustable manually with the help of sterile handles. Sterile handles two numbers should be supplied with light.
4. Light intensity at a distance of 1 meter for each dome at least 1,60,000 lux
5. Life of LEDS 40,000hrs or more
6. Color temperature should be between 3500 to 4500 degree K
7. Light field diameter shall be above 20 cm or better.
8. Light dome shall be compatible for laminar air flow.
9. Power supply: 230 Volts AC, 50Hz.
10. Should be CE / USFDA certified.
11. Installation will be done free of cost by supplier with all Civil, Mechanical and Electrical works required.
12. **Guarantee:** Three years from the date of installation.
13. **CMC:** comprehensive annual maintenance will be given @ 4 % (of net rate - inclusive of Excise Duty & exclusive of VAT / CST etc.) plus service tax (as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after completion of Guarantee period of three years.
14. The dealership/service provider should be based in Rajasthan.
15. The company should submit technical compliance sheet as per technical specifications mentioning the make & model of quoted item along with catalogue in the Technical bid.
16. Four Preventive maintenance services annually during guarantee and CMC period are essential.
17. Final technical approval after demonstration. (if required).

### **B. The bid is re-scheduled as follows:**

Existing Dates			Extended Dates		
Last Date for Sale of Bid Form	Last Date of Receipt of Bid Form	Date of Opening of Technical Bid	Last Date for Sale of Bid Form	Last Date of Receipt of Bid Form	Date of Opening of Technical Bid
1	2	3	4	5	6
up to 11:00 A.M. 09.03.2016	up to 1.00 P.M. 09.03.2016	from 3.00 P.M. 09.03.2016	up to 11:00 A.M. 21.03.2016	up to 1.00 P.M. 21.03.2016	from 3.00 P.M. 21.03.2016

This corrigendum shall be signed and annexed with bid document. This bears an approval of Managing Director, Rajasthan Medical Services Corporation Limited, Jaipur.

**Executive Director (EPM)**  
**RMSCL, Jaipur**