



RAJASTHAN MEDICAL SERVICES CORPORATION LTD.  
D- Block, Swasthya Bhawan, Tilak Marg, C-Scheme JAIPUR (Raj)



Ph. No. 0141-2223887, Fax No. 0141-2228065

CIN: U24232RJ2011SGC035067

E-Mail -edepmrmc-ri@nic.in

No. F-8(24 ) RMSC/EPM/M-1/R. C. 15-16/NIB-121/2015/215 6

Dated: 27/11/15

**CLARIFICATION/CORRIGENDUM/ADDENDUM**

**Subject:- Revised technical Specifications and other terms and conditions of Bid document of (01) Artery forceps Criles straight box joint 6" and 8" (02) Artery forceps Criles curved on flat box joint 6" , 8", 12" and 15" (03) Sterilizer Forceps Cheatle's Pattern 10" (04) Scissors Straight with blunt point 5" (12.70 cms), 6", 8" and 10" S.S (05) Anal Dilators No. 1 to 5 (06) Malleable (Fistula) Probe (07) First Aid Surgical Kit / Dressing Kit under NIB No. F-8( 24 ) RMSC/EPM/M-1/R. C. 15-16/NIB-121/2015/1727 Dated: 20.10.15**

In Reference to subject cited above and NIB-121, the Technical specification are examined by the competent Authorities. The following Clarification/Corrigendum/Addendum are issued for inclusion in Bid document & Technical Specification of items as below:-

**01 Amended Technical Specifications of Items as below:-**

S. N.	Name Item With full Specification	Packing Unit														
1	2	3														
1.	Artery forceps Criles straight	---														
	A. Artery forceps Criles straight box joint 6"	Each														
	B. Artery forceps Criles straight box joint 8"	Each														
2.	Artery forceps Criles curved	---														
	A. Artery forceps Criles curved on flat box joint 6"	Each														
	B. Artery forceps Criles curved on flat box joint 8"	Each														
	C. Artery forceps Criles curved on flat box joint 12"	Each														
	D. Artery forceps Criles curved on flat box joint 15"	Each														
3.	Sterilizer Forceps Cheatle's Pattern 10"	Each														
4.	Scissors Straight with blunt point	---														
	A. Scissors Straight with blunt point 5" (12.70 cms) S.S	Each														
	B. Scissors Straight with blunt point 6" S.S.	Each														
	C. Scissors Straight with blunt point 8" S.S.	Each														
	D. Scissors Straight with blunt point 10" S.S	Each														
5.	Anal Dilators No. 1 to 5	Each														
6.	Malleable (Fistula) Probe	Each														
7.	First Aid Surgical Kit / Dressing Kit	Single Unit														
	<table border="1"> <thead> <tr> <th>Sr. No.</th> <th>Items</th> </tr> </thead> <tbody> <tr> <td>1</td> <td>Dissecting forceps 6"</td> </tr> <tr> <td>2</td> <td>Dressing Scissors 6"</td> </tr> <tr> <td>3</td> <td>Needle Holder 6"</td> </tr> <tr> <td>4</td> <td>Artery Forceps 6"</td> </tr> <tr> <td>5</td> <td>Mosquito Forceps 5"</td> </tr> <tr> <td>6</td> <td>B.P. Handle No. 4 with blade.</td> </tr> </tbody> </table>	Sr. No.	Items	1	Dissecting forceps 6"	2	Dressing Scissors 6"	3	Needle Holder 6"	4	Artery Forceps 6"	5	Mosquito Forceps 5"	6	B.P. Handle No. 4 with blade.	
Sr. No.	Items															
1	Dissecting forceps 6"															
2	Dressing Scissors 6"															
3	Needle Holder 6"															
4	Artery Forceps 6"															
5	Mosquito Forceps 5"															
6	B.P. Handle No. 4 with blade.															
Sr. No.	Terms & Conditions for above items:															
1	All surgical instruments should be CE certified (valid documentation should be submitted in technical Bid)															
2	All surgical instruments should be manufactured by ISO 9001:2008 & ISO 13485:2003 (valid documentation should be submitted in technical Bid)															
3	Manufacturer should have ISO 7153 certificate. (valid documentation should be submitted in technical bid)															
4	Stainless Steel grade AISI 410 and 420, test report should be submitted.															
5	Laser marking along with code should be on the instrument.															
6	Guarantee: One year against any manufacturing defect.															
7	The company should mention the make & model of the quoted items and submit the technical compliance sheet in technical Bid															
8	Final technical approval after sample demonstration.															

2/12

02 Amended BOQ of Items as below:-

BF-IV

**Financial Bid For Quoted Item**

S. N.	Name Item With full Specification	Brand	Approx Qty.	Packing Unit	Net rate per packing unit (In Rs.)	Rate of RST (VAT) of Concession al CST against C-form, if applicable	Amount of RST (VAT) of Concessional CST against C-Form, if applicable	Rate of Excise duty (It should be included in net rate) (Col.6)	Amount of excise duty workout accordingly Col. No.9	Total Amount In Rs. (6+8)						
1	2	3	4	5	6	7	8	9	10	11						
1.	Artery forceps Criles straight		----	---	<b>Do not quote rates here.</b>											
	A. Artery forceps Criles straight box joint 6"		110	Each												
	B. Artery forceps Criles straight box joint 8"		110	Each												
2.	Artery forceps Criles curved		----	---												
	A. Artery forceps Criles curved on flat box joint 6"		110	Each												
	B. Artery forceps Criles curved on flat box joint 8"		110	Each												
	C. Artery forceps Criles curved on flat box joint 12"		110	Each												
	D. Artery forceps Criles curved on flat box joint 15"		110	Each												
3.	Sterilizer Forceps Cheattle's Pattern 10"		110	Each												
4.	Scissors Straight with blunt point		----	---												
	A. Scissors Straight with blunt point 5" (12.70 cms) S.S		110	Each												
	B. Scissors Straight with blunt point 6" S.S.		110	Each												
	C. Scissors Straight with blunt point 8" S.S.		110	Each												
	D. Scissors Straight with blunt point 10" S.S		110	Each												
5.	Anal Dilators No. 1 to 5		110	Each												
6.	Malleable (Fistula) Probe		55	Each												
7.	First Aid Surgical Kit / Dressing Kit		55	Single Unit												
<b>Rates shall only be filled in BOQ <a href="https://eproc.rajasthan.gov.in">https://eproc.rajasthan.gov.in</a></b>																

Note:- The rates of consumables shall not be included/considered for determining the lowest Bidder.

Signature

Name in Capitals  
Company /Firm Seal

Date

Note: -

1. The concessional CST against c- form shall be applicable.
2. The rate quote should be inclusive of excise duty but exclusive of sales tax./vat
3. Excise component should be separately shown in column no.10 for further reference
4. Rate should be quoted on separate sheets for each item.
5. Rate should be quoted only for packing units as mentioned in the bid
6. No quantity or cash discounts should be offered.
7. Read all the terms & conditions before filling the BF-IV.
8. Please quote rates in absolute amount only.

It is also clarified that above revised information shall be communicated to all participating bidders on the website [www.rmhc.nic.in](http://www.rmhc.nic.in) and [sppp.raj.nic.in](http://sppp.raj.nic.in). Please note that individual bidder will not be intimated."

**Note:** Please note that all above amendments/corrigendum in bid conditions is the integral part of (Section-V, Schedule of Supply, Point no. 03) and the bid document. This corrigendum/ addendum should be signed and annexed with bid document. All other terms & conditions remains the same. This bears the approval of M.D., RMSCL, Jaipur.

  
**Executive Director (EPM)**  
**RMSCL, Jaipur**