



RAJASTHAN MEDICAL SERVICES CORPORATION LTD.

D- Block, Swasthya Bhawan, Tilak Marg, C-Scheme JAIPUR (Raj)

Ph. No. 0141-2223887, Fax No. 0141-2228065

CIN:U24232RJ2011SGC035067

E-Mail -edepmrmssc-ri@nic.in

No. F-8() RMSC/EPM/M-3/15-16/NIB-123/

1932

Dated: 5-11-15



CLARIFICATION/CORRIGENDUM/ADDENDUM

Subject: -Revised technical Specifications and other terms and conditions of Bid document the rate contract for Equipment Mother Milk Bank NIB No. F-8()RMSC/EPM/M-3/15-16/NIB-123/1683 Dated: 16.10.2015.

In Reference to subject cited above and NIB-123, the various representations received from the firms and issues raised by the Bidders are examined by the competent Authorities and technical committee. The following Corrigendum/Addendum is issued for inclusion in bid document & Technical Specification of items (Section-V, Schedule of Supply, Point no. 3) as below:-

Revised Technical Specification

Item 1. : Manual Pump

1. Should be BPA Free.
2. Should have soft round Funnel properly Air tight seal.
3. Should be one hand system.
4. Should have parts can be easily dismantled to make sterilization easy & quick.
5. **should be supplied with 12 bottles with cap (compatible with Pump.)**

Item 2. : Electric Breast Pump (Professional/Hospital Use)

1. Should be Light weighted & portable.
2. Should be silent motor to make noise free operation (allowed sound 58 db)
3. Should have soft round Funnel properly Air Tight seal .
4. Should have multi level suction strength controller (up to 250 mmHg).
5. Should have parts that can be easily dismantled to make sterilization easy & quick.
6. Should be AC & Battery Operated (Maximum 360 minutes) .
7. LED monitor - Easy to use and view.
8. Bottle stand Secures bottle to prevent spills.
9. **Should be supplied with 190mL (Minimum), 100 NO's Plastic Bottles with cap (compatible with Pump.)** .

Item 3. : Electric Breast Pump (Domestic use)

1. Should be Light weighted & Portable.
2. Should be silent motor to make noise free operation (allowed sound 58 db)
3. Should have soft round Funnel properly Air Tight seal .
4. Should have multi level suction strength controller (up to 250 mmHg) .
5. Should have parts that can be easily dismantled to make sterilization easy & quick.
6. Should be AC & Battery Operated (Maximum 360 minutes).
7. Bottle stand Secures bottle to prevent spills.
8. **should be supplied with 12 bottles with cap (compatible with Pump.)**

Item 4. : Multi Functional Sterilization Machine

1. Steam Sterilizer is carefully designed for effective sterilization of baby bottles and feeding accessories.
2. Sterilize bottles, breast pump and other accessories
3. Suitable for all sizes and shapes of feeding bottles

4. Large reversible tray for smaller accessories
5. Proper locks to secure the cover
6. Effective sterilization without use of chemicals
7. Turns off automatically when sterilization is done
8. Sterilizing time of 12 to 15 minutes.
9. Contents remain sterile up to 12 hours

Terms & Conditions:

1. Manufacturer should have ISO 9001 / ISO 13485 certified Certificate.
2. Equipment should be European CE/USFDA.
3. Manufacturer should provide Certificate of calibration with All Pumps.
4. Should be supplied with following Spares Parts / accessories like :
 - a) Membranes/ diaphragms - 4 per unit
 - b) Valve(If any) - 4 per unit
 - c) Breast shield (different sizes) - 10 per unit
 - d) Silicone tubing - 10 per unit
 - e) Tubing adapter (If any)- 4 per unit
 - f) Locking ring (If any) - 4 per unit
 - g) Locking adapter (If any) - 4 per unit.
5. **Guarantee:** Two years on equipment from the date of installation.
6. **CMC:** CMC will be given @ 5 % (of net rate - inclusive of Excise Duty & exclusive of VAT / CST etc.) Plus service tax (as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for three years (on yearly basis) after completion of Guarantee period of two years.
7. Installation & Training will be done by supplier free of cost.
8. The service engineer should be based in Rajasthan.
9. The company should mention the make & model name / number of the quoted equipment and submit the technical brochure of the quoted model in the Technical bid along with compliance sheet as per technical specifications.


The last date for sale of bid, receipt of bid & date of opening of technical bid is hereby extended as below:-

Extended dates		
Last Date for Sale of Bid Form	Last Date of Receipt of Bid Form	Date of Opening of Technical Bid
27.11.2015 11:00 AM	27.11.2015 01:00 PM	27.11.2015 03:00 PM

Please note that all clarification/amendment/corrigendum in technical specifications/bid conditions is the integral part of the bid document. This corrigendum/ addendum should be signed and annexed with bid document.

All other terms & conditions remains the same.

This bears the approval of M.D., RMSCL, Jaipur.


 Executive Director (EPM)
 RMSC, Jaipur