



RAJASTHAN MEDICAL SERVICES CORPORATION LTD.

D- Block, Swasthya Bhawan, Tilak Marg, C-Scheme JAIPUR (Raj)



Ph. No. 0141-2223887, Fax No. 0141-2228065

CIN: U24232RJ2011SGC035067

E-Mail -edepmrmsc-ri@nic.in

No. F-8()RMSC/EPM/M-1/NIB-81/2015/ 382

Dated: 18/6/15

CLARIFICATION/CORRIGENDUM/ADDENDUM

Subject: - Revised Bid Schedule for NIB No. F-8()RMSC/EPM/M-1 /NIB-81/2014-15/ 9906

Dated: 17.03.15 & Clarification/Corrigendum/Addendum No. F-8()RMSC/EPM/M-1/NIB-81/2014-15/114 Dated: 15.05.15

In Reference to subject cited above and NIB-81, The following Clarification/Corrigendum/Addendum are issued for inclusion in Bid document revised bid schedule as follows:-

(01) Revised Technical Specification of Item Centrifuge Machine (36 Tubes) as below:-

Point No. 02 of Technical Specification of Item- Speed range 500 to 8000 rpm on load. Which should be controlled by variable speed regulator is replaced by Speed range 500 to 7000 rpm or more on load. Which should be controlled by variable speed regulator.

(02) Revised bid schedule:-

E-bids are invited as per following revised time schedule:-

Last date for sale of bid up to	Last date of receipt of bid up to	Date of opening of technical bid from
14.07.2015 11.00 AM	14.07.2015 01.00 PM	14.07.2015 03.00 PM

It is also clarified that information of award of contract shall be communicated to all participating bidders on the website www.rmsc.nic.in and sppp.raj.nic.in. Please note that individual bidder will not be intimated."

Note: Please note that all above amendments/corrigendum in bid conditions is the integral part of (Section-V, Schedule of Supply, Point no. 3) and the bid document. This corrigendum/addendum should be signed and annexed with bid document. All other terms & conditions remains the same. This bears the approval of M.D., RMSCL, Jaipur.

**Executive Director (EPM)
RMSCL, Jaipur**