



**RAJASTHAN MEDICAL SERVICES CORPORATION LTD.**  
D- Block, Swasthya Bhawan, Tilak Marg, C-Scheme JAIPUR (Raj)



Ph. No. 0141-2223887, Fax No. 0141-2228065  
No. F-8( ) RMSC/EPM/M-3/15-16/NIB-126/

CIN:U24232RJ2011SGC035067

E-Mail -[edepmrmisc-rj@nic.in](mailto:edepmrmisc-rj@nic.in)

1935

Dated: 5-11-15

**CLARIFICATION/CORRIGENDUM/ADDENDUM**

**Subject: -Revised technical Specifications and other terms and conditions of Bid document the rate contract for Equipment Mother Milk Bank NIB No. F-8( )RMSC/EPM/M-3/15-16/NIB-126/1684 Dated: 16.10.2015.**

In Reference to subject cited above and NIB-126, the various representations received from the firms and issues raised by the Bidders are examined by the competent Authorities and technical committee. The following Corrigendum/Addendum is issued for inclusion in bid document & Technical Specification of items (Section-V, Schedule of Supply, Point no. 3) as below:-

**Revised Technical Specification**

1. Analyzer must be portable & ready to use model for screening for **Human Milk content**.
2. Equipment should be based on **Mid IR-technology (Mid Infrared transmission spectroscopy)** and all the nutritional components are analyzed in a single run.
3. The analyzer should be capable of direct determination of the nutritional content of breast milk without using any chemicals/ reagents.
4. With a small sample size of 3-12 ml it should analyze the composition of breast milk for the content of fat, protein, carbohydrates, energy and dry matter.
5. The analyzer should exclude non-protein compounds while determining protein level i.e. rather than total nitrogen or crude (total) protein, it should determine true protein level
6. The analyzer should produce the following units of measurements: Fat g/100 ml, Carbohydrates g/100ml, Protein g/100 ml, Energy kcal/100 ml, Dry matter g/100 ml
7. The analyzer should not use any chemical additive
8. It should give quick results in digital LCD display.
9. The data should be easily transferable to PC for permanent data storage.
10. The analyzer should be standardized for high precision and accuracy.
11. It should have sufficient internal storage capacity and also some method of back up storage of measurement data in case of power cuts.
12. Should have provision to attach a printer for quick printing of the result.
13. Should be Capable of working in power supply of 220-240VAc, 50Hz.
14. User/Technical/Maintenance manuals/**Guide Book** to be supplied - should be in English/**Hindi**.
15. Should be supplied with following Accessories :
  - a) Thermal Printer Paper Roll – 10 No's
  - b) Sample Cup – 4 No's
  - c) Waste Container – 4 No's
  - d) Sample Tube / Hose – 4 No's
  - e) Waste Container Hose - 4 No's
  - f) Cleaning Solution – 4 Packs
  - g) Containers for Flushing Solution – 4 No's
  - h) CD with Software with all connection Cables to PC – 2 No's

**Terms & Conditions:**

1. Manufacturer should have ISO 9001 / ISO 13485 certified Certificate.
2. Equipment should be European CE/USFDA.
3. Manufacturer should provide Certificate of calibration with Analyzer .
4. List of important spare parts and accessories with their part number and costing .
5. **Guarantee:** Two years on equipment from the date of installation.

6. **CMC:** CMC will be given @ 5 % (of net rate - inclusive of Excise Duty & exclusive of VAT / CST etc.) plus service tax (as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for three years (on yearly basis) after completion of Guarantee period of two years.
7. Installation & Training will be done by supplier free of cost.
8. The service engineer should be based in Rajasthan.
9. The company should mention the make & model name / number of the quoted equipment and submit the technical brochure of the quoted model in the Technical bid along with compliance sheet as per technical specifications.

The last date for sale of bid, receipt of bid & date of opening of technical bid is hereby extended as below:-

Extended dates		
Last Date for Sale of Bid Form	Last Date of Receipt of Bid Form	Date of Opening of Technical Bid
27.11.2015 11:00 AM	27.11.2015 01:00 PM	27.11.2015 03:00 PM

Please note that all clarification/amendment/corrigendum in technical specifications/bid conditions is the integral part of the bid document. This corrigendum/ addendum should be signed and annexed with bid document.

All other terms & conditions remains the same.

This bears the approval of M.D., RMSCL, Jaipur.

  
**Executive Director (EPM)**  
**RMSC, Jaipur**