

Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065

E-mail : edepmrmisc-ri@nic.in

CIN:U24232RJ2011SGC035067

Website:www.rmssc.nic.in

No. F-8(18) RMSC/EPM/M-5/2015-16/NIB-155/1883

Dated: 13/4/16

CLARIFICATION/CORRIGENDUM/ADDENDUM

Subject :- Amended technical Specifications and date extension of Bid document for Pulse Oxymeter under NIB No. F-8(18) RMSC/EPM/M-5/2015-16/NIB-155/1114 Dated 10.03.2016.

In Reference to above cited subject and NIB-155, The technical specification of items (Section V, Clause No. 3) are here by amended as below:-

Specification of Pulse Oxymeter

1. Suitable for all types of patients: adult, pediatric, neonatal.
2. Compact Table Top Pulse Oxymeter with colour monitor at least 3.5" TFT/ LCD Screen.
3. SpO₂ measurement range at least 40-70 and 70 to 99%, minimum gradation 1%.
4. Accuracy of SpO₂ better than $\pm 1\%$ for range 40-70 and better than $\pm 3\%$ for range 70-99.
5. Pulse rate range at least **25 to 250bpm**, minimum gradation 1 bpm.
6. Accuracy of pulse rate better than ± 5 bpm.
7. Audiovisual alarms required: high and low SpO₂ and pulse rate (operator variable settings), sensor disconnected, sensor failure, low battery.
8. Plethysmograph (may be in form of bar) display is mandatory.
9. Should have minimum 24 hrs trend memory for SpO₂ & PR.
10. Display shows SpO₂ (%), HR (bpm) and signal strength bar.
11. Should have clinically proven track record to work during motion and very low perfusion conditions.
12. Large display readable from distance, display covers of durable plastic.
13. User preset of high/low alarms on SpO₂ and pulse rate monitoring.
14. Silencing feature for audio alarm.
15. Display reports system errors, probe failure and built-in battery status.
16. Automatic switch from mains to batteries in case of power failure.
17. Power requirements: 220 V/50 Hz and internal re-chargeable battery, approx **3 hrs (minimum)** back up, automatic recharge.
18. Manufacturer should have ISO 13485 Certificate.
19. Equipment should be CE certified.
20. Guarantee: Two years on equipment from the date of installation.
21. **CMC:** CMC will be given @ 5 % (of net rate - inclusive of Excise Duty & exclusive of VAT / CST etc.) plus service tax (as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for three years (on yearly basis) after completion of Guarantee period of two years.
22. **Table Top Pulse Oxymeter should be supplied with standard set of:**
 1. Reusable SpO₂ sensors for Adult including connection cable.
 2. Reusable SpO₂ sensors for neonate, Wrap around type.
 3. User manual with trouble-shooting guidelines should be provided by supplier.
23. Installation will be done by supplier free of cost.
24. The service engineer should be based in Rajasthan.

25. Demonstration of equipment is must for technical approval.
26. The company should mention the make & model name / number of the quoted equipment and submit the technical brochure of the quoted model in the Technical bid along with compliance sheet as per technical specifications.

In Reference to above cited subject and NIB-155, the last date for sale of bid, receipt of bid and date of opening of technical bid is hereby extended as below:-

Existing Date			Extended date		
Last Date for Sale of Bid Form	Last Date of Receipt of Bid Form	Date of Opening of Technical Bid	Last Date for Sale of Bid Form	Last Date of Receipt of Bid Form	Date of Opening of Technical Bid
21.04.2016 11:00 AM	21.04.2016 01:00PM	21.04.2016 03:00 PM	09.05.2016 11:00 AM	09.05.2016 01:00PM	09.05.2016 03:00 PM

Note: Please note that all above amendments/corrigendum in technical specifications/bid conditions is the integral part of the bid document. This corrigendum/ addendum should be signed and annexed with bid document.

All other terms and conditions of bid shall remain the same.

This bears the approval of M.D., RMSCL, Jaipur

20
2
Executive Director (EPM)

RMSCL, Jaipur