

Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065

CIN:U24232RJ2011ISGC035067

E-mail : edepmrm-sc-rj@nic.in

Website: www.rm-sc.nic.in

No. F-8(5) RMSC/EPM/M-2/R.C.15-16/NIB-109/2015/ 1677

Dated: 15/10/18

CLARIFICATION/CORRIGENDUM/ADDENDUM

Sub:- Revised Technical Specifications and revised bid submission date of item IUCD Kit and PPIUCD Forceps under NIB No. F-8()RMSC/EPM/M-2/R.C.15-16/NIB-109/2015/1277 dated 09.09.15

In Reference to subject cited above and NIB-109, the various representations received from the firms and issues raised by the Bidders are examined by the competent Authorities and technical committee. The following Clarification/Corrigendum/Addendum is issued for inclusion in Bid document & Technical Specification of items as below:-

A. Amended Technical Specifications of item IUCD KIT:-

S.No.	Name of Item	Qty. in a set
1	Cheetal Forceps	1 No.
2	Cusco speculum 110 mm (L) X 35 mm (B)	1 No.
3	Sponge Forceps	1 No.
4	Artery Forceps 8" inch (Curved)	1 No.
5	Tenaculam (Single tooth) Double Curve 285 mm	1 No.
6	Uterine sound sims 330 mm	1 No.
7	Thread Cutting Scissor	1 No.
8	Vaginal Speculum Sims 15 cm (L) {34mm (B) X 30 mm (B)}	1 No.
General Terms & Conditions:		
9	All surgical instruments should be CE certified (valid documentation should be submitted in technical Bid)	
10	All surgical instruments should be manufactured by ISO 9001:2008 & ISO 13485:2003 (valid documentation should be submitted in technical Bid)	
11	Manufacturer should have ISO 7153 certificate. (valid documentation should be submitted in technical bid)	
12	Stainless Steel grade AISI 410 and 420, test report should be submitted.	
13	Laser marking along with code should be on the instrument.	
14	Guarantee: One year against any manufacturing defect.	
15	The company should mention the make & model of the quoted items and submit the technical compliance sheet in technical Bid	
16	Final technical approval after sample demonstration.	


B. Amended Technical Specifications of item PPIUCD FORCEPS:-

S.No.	Name of Item	Qty. in a set
1	Placental forcep 315 mm	1 No.
2	Sponge Holding Forceps	1 No.
General Terms & Conditions:		
3	All surgical instruments should be CE certified (valid documentation should be submitted in technical Bid)	
4	All surgical instruments should be manufactured by ISO 9001:2008 & ISO 13485:2003 (valid documentation should be submitted in technical Bid)	
5	Manufacturer should have ISO 7153 certificate. (valid documentation should be submitted in technical bid)	
6	Surgical instruments made by steel grade AISI 410, test report should be submitted.	
7	Laser marking along with code should be on the instrument.	
8	Guarantee: One year against any manufacturing defect.	
9	The company should mention the make & model of the quoted items and submit the technical compliance sheet in technical Bid	
10	Final technical approval after sample demonstration.	

C. The bid is re-scheduled as follows:

Existing Dates			Extended Dates		
Last Date for Sale of Bid Form	Last Date of Receipt of Bid Form	Date of Opening of Technical Bid	Last Date for Sale of Bid Form	Last Date of Receipt of Bid Form	Date of Opening of Technical Bid
1	2	3	4	5	6
up to 11:00 A.M. 27.10.2015	up to 1.00 P.M. 27.10.2015	from 3.00 P.M. 27.10.2015	up to 11:00 A.M. 29.10.2015	up to 1.00 P.M. 29.10.2015	from 3.00 P.M. 29.10.2015

This corrigendum shall be signed and annexed with bid document. This bears an approval of Managing Director, Rajasthan Medical Services Corporation Limited, Jaipur.


 Executive Director (EPM)
 RMSCL, Jaipur ✓