

# Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065  
CIN:U24232RJ2011SGC035067

E-mail : edepmrm-sc-rj@nic.in  
Website: www.rm-sc.nic.in

No. F-8 (7) RMSC/EPM/M-2/NIB-110/15-16/1995

Dated: 16/11/15

## CLARIFICATION/CORRIGENDUM/ADDENDUM

**Sub:- Revised Technical Specifications and revised bid submission date of item Open Newborn Care System On Trolley With Drawers, With Radiant Warmer, Oxygen-Provision under NIB No. F-8 ( )RMSC/ EPM/M-2/R.C.15-16/NIB-110/2015/1291 dated 11.09.15**

In Reference to subject cited above and NIB-110, the various representations received from the firms and issues raised by the Bidders are examined by the competent Authorities and technical committee. The following Clarification/Corrigendum/Addendum is issued for inclusion in Bid document & Technical Specification of items as below:-

### **A. Amended Technical Specifications of item Open Newborn Care System On Trolley With Drawers, With Radiant Warmer, Oxygen-Provision:-**

1. Mobile newborn resuscitation table, table should be tiltable  $\pm$  10-12 degrees in 6-8 steps.
2. Antistatic castors, 2 with breaks.
3. Table surface should be of Plastic / Aluminium / ABS.
4. Mattress-padding: foam density approx. 21 - 25 kg/m<sup>3</sup>
5. Mattress cover should be sealed and waterproof.
6. Side boards should be transparent with thickness between 6-8 mm & 110 - 120 mm height.
7. Under table 2 storage drawers.
8. Side rails allow for mounting of accessories.
9. Hood suspended above the table integrates heating element and overhead light.
10. Overhead light: LED Light or Halogen Light.
11. Provision for integration of one B-type (10.2 Ltr.) oxygen cylinder.
12. Inlet must be able to handle high pressure.
13. Control unit should have T-piece Resuscitator and meconium aspirator suction.
14. 10 nos. T-pieces with 10 circuits and 15 nos. masks (5 each in three sizes) should be supplied with unit at the time of installation.
15. Heating element: quartz heating element with lifelong guarantee.
16. Control unit allow skin temperature preset indicator and drives radiant heater output in servo and manual mode.
17. Unit should have pre-heating mode. Pre-heating mode should be between 25-30% of power.
18. Integrated timer: 1 to 59 min with count-up feature.
19. Temperature range, skin: 34 to 38°C (user pre-settable)
20. Monitoring of skin temperature by means of sensor, range: 30 to 42°C
21. Heater output: 0 to 100% in increments of 5%
22. Control unit: audiovisual alarms according to timer and temperature presets avoiding overheating.
23. Audio-visual alarms for systems errors & sensor failure.
24. Power requirement: 220 V/ 50 Hz
25. Power consumption not more than 800 W.
26. Equipment should be produced by ISO 13485 certified manufacturer (Valid documentation should be submitted in technical bid).
27. Equipment should be European CE or FDA approved and T-piece Resuscitator should be European CE certified. (Valid documentation should be submitted in technical bid).
28. Equipment should be supplied with:
  - a. Mattress (1No.)
  - b. Skin temperature probe including connection cable (1No.)
  - c. Spare skin temperature probe including connection cable (02 No.)
  - d. Oxygen cylinder (10.2 Ltr.) with regulator (1No.)
  - e. Spare set of fuses (5 Set)
  - f. Operating manual / user training material (1Set)
  - g. Voltage Stabilizer of suitable rating for Voltage Protection (01 No)
29. **Guarantee:** Two years on equipment from the date of installation.
30. **CMC:** CMC will be given @ 5 % (of net rate - inclusive of Excise Duty & exclusive of VAT / CST etc.) plus service tax (as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after completion of Guarantee period of two years.
31. Installation will be done by supplier free of cost.
32. The service engineer should be based in Rajasthan.
33. Demonstration of equipment is must for technical approval.
34. The company should mention the make & model name / number of the quoted equipment and submit the technical brochure of the quoted model in the Technical bid along with compliance sheet as per technical specifications.

*2/12*

**B. The bid is re-scheduled as follows:**

Existing Dates			Extended Dates		
Last Date for Sale of Bid Form	Last Date of Receipt of Bid Form	Date of Opening of Technical Bid	Last Date for Sale of Bid Form	Last Date of Receipt of Bid Form	Date of Opening of Technical Bid
1	2	3	4	5	6
up to 11:00 A.M. 17.11.2015	up to 1:00 P.M. 17.11.2015	from 3:00 P.M. 17.11.2015	up to 11:00 A.M. 01.12.2015	up to 1:00 P.M. 01.12.2015	from 3:00 P.M. 01.12.2015

This corrigendum shall be signed and annexed with bid document. This bears an approval of Managing Director, Rajasthan Medical Services Corporation Limited, Jaipur.

**Executive Director (EPM)  
RMSCL, Jaipur**