

राजस्थान चिकित्सा सेवा निगम

स्वास्थ्य भवन, सी-स्कीम, जयपुर

फोन नः 0141-2223887, फैक्स नः 0141-2228065

ई-मेल : rmsc@nic.in

F-8()RMSC/EPM/M-2/NPCDCS/12-14/NIT-12/12/ 270

Dated : 15.01.2013

16.01.13

CLARIFICATION

In reference to F-8()RMSC/EPM/M-2/NPCDCS/12-14/NIT-12/12/142 Dated 22.11.2012 and Corrigendum No 26D dated 09.01.2013

It is clarified that:-


(A) In item No 3-Multi Para Monitor Point No 11 is as under:

"The SpO2 technology (masimo / equivalent) should be sense SPO2 in hypotensive, shivering & motion condition." shall be substituted as "The SpO2 technology (masimo) should be sense SPO2 in hypotensive, shivering & motion condition."

(B) In item No 5-Ventilator (Adult, Paediatric and Neonatal) Point No 1.12 is as under :

"should have integrated or synchronized ultrasonic nebuliser with capability to deliver fine particle On line" shall be substituted as "should have integrated (in-built) nebuliser or synchronized ultrasonic nebuliser with capability to deliver fine particle On line"

The above clarification shall be the integral part of the terms and conditions of tender document.


Executive Director (EPM)
RMSC, Jaipur

F-8()RMSC/EPM/M-2/NPCDCS/12-14/NIT-12/12/

Dated : 15.01.2013

Copy to:-

1. M.D, RISL, Yojna Bhawan, Jaipur
2. PA to M.D RMSC, Jaipur
3. AGM (IT) RMSC, Jaipur
4. Guard File.

11
Executive Director (EPM)
RMSC, Jaipur