

RAJASTHAN MEDICAL SERVICES CORPORATION LTD.
D- Block, Swasthya Bhawan, Tilak Marg, C-Scheme JAIPUR (Raj)



Ph. No. 0141-2223887, Fax No. 0141-2228065

E-Mail - edepmrmsc-rj@nic.in

No. F-8()RMSC/EPM/M-5/14-15/NIB-55/ 6283

Dated: 21/9/14

CORRIGENDUM/ADDENDUM

Subject: - Amended date extension, technical specifications of bid document for the equipments of Digital Thermometer and ARI Timer under NIB No. F-8()RMSC/EPM/M-5/14-15/NIB-55/4003 Dated: 21.07.2014.

1. In reference to subject cited above NIB No. F-8()RMSC/EPM/M-5/14-15/NIB-55/4003 Dated: 21.07.2014, the last date for sale of bid, receipt of bid & date of opening of technical bid is hereby extended as below:-

| Existing Dates | | | Extended Dates | | |
|--------------------------------|----------------------------------|----------------------------------|--------------------------------|----------------------------------|----------------------------------|
| Last date for Sale of Bid Form | Last date of Receipt of Bid Form | Date of Opening of Technical Bid | Last date for sale of Bid Form | Last date of Receipt of Bid Form | Date of Opening of Technical Bid |
| 1. | 2. | 3. | 4. | 5. | 6. |
| 02-Sept-2014 1:00 PM | 03-Sept-2014 1:00 PM | 03-Sept-2014 3:00 PM | 18-Sept-2014 1:00 PM | 19-Sept-2014 1:00 PM | 19-Sept-2014 3:00 PM |

2. In reference to above cited subject and NIB, the technical specification of items (Annexure-C) are here by amended as below:-

Technical Specifications of Items for Digital Thermometer :

Digital Thermometer:- Digital LCD Display / at least 3 1/2 digit, Celsius Fahrenheit scale, accurate measurement $\pm 0.1^{\circ}\text{C}$ between 89.6°F - 107.6°F / 32°C - 42°C should be safe to use, no glass, no mercury. Should have liquid crystal easy to read, should be water proof for ease of cleaning, there should be battery powered with low battery indicator, it should with clear instructions for use or preventive maintenance.

Registration and Model approval certificate from Department of Legal Metrology (Weights and Measures, Ministry of Consumer Affairs Government of India) is mandatory.

The terms and conditions of the Bid remain the same.
This is issued with the approval of MD, RMSC.

Executive Director (EPM)
RMSC, Jaipur