

Rajasthan Medical Services Corporation Limited, Jaipur

E.D. (EPM), D-Block, Swasthya Bhawan, C-Scheme, Jaipur-302005

Ph. No. 0141-2223887, Fax No. 0141-2228065

E-Mail – edepmrmisc-rj@nic.in

No. F-8 () /RMSC/EPM / M-5/2014-15/NIB-55/ 6757

Dated: 22/10/14

OFFICE ORDER

The technical bids of participating bidders opened on 19.09.2014 are examined and evaluated. The bidders declared responsive / non responsive as per documentary & technical evaluation of NIB No. F-8 () /RMSC/EPM /M-5/2014-15/NIB-55/4003, Dated: 21.07.2014 are given below:-

S. N.	Name of Item	Name of Firm	Responsive / Non Responsive	Remarks
1.	Digital Thermometer	M/s Microgene Diagnostic System Pvt. Ltd., 806-Vikram Tower, Rajendra Place, New Delhi -110008. E-mail: contact@microgenediagnostic.com	Responsive	Nil
2.	Digital Thermometer	M/s Medicare Products Inc., C-53A, Mansarover Garden, New Delhi - 110015. E-mail: mcpeti@vsnl.com, mail@medicareproduct.com	Non Responsive	Due to non submission of model approval certificate from department of Legal Metrology (weight and measures, Ministry of consumer affairs Government of India).
3.	ARI Timer	M/s Medicare Products Inc., C-53A, Mansarover Garden, New Delhi - 110015. E-mail: mcpeti@vsnl.com, mail@medicareproduct.com	Responsive	Nil

Online financial bids of Digital Thermometer and ARI Timer will be opened on dated: 27.10.2014 at 4:00 P.M. of the responsive bidders for this bid.

This bears the approval of MD, RMSC Rajasthan, Jaipur.

Executed Director (EPM)
Rajasthan Medical Services Corporation Ltd.,
Jaipur