



RAJASTHAN MEDICAL SERVICES CORPORATION LTD.

D- Block, Swasthya Bhawan, Tilak Marg, C-Scheme JAIPUR (Raj)



Ph. No. 0141-2223887, Fax No. 0141-2228065

CIN:U24232RJ2011SGC035067

E-Mail -edepmrmisc-ri@nic.in

क्रमांक No. F-8() RMSC/EPM/M-3/14-15/NIB-59/

7552

Dated: 30-1-2015

CLARIFICATION/CORRIGENDUM/ADDENDUM

Subject: -Revised technical Specifications and other terms and conditions of Bid document Single, 3 & 12 Channel ECG Machine under NIT No. F-8() RMSC/EPM/M-3/14-15/NIB-59/7247 Dated: 23.12.2014

In Reference to subject cited above and NIB-59, the various representations received from the firms and issues raised by the Bidders are examined by the competent Authorities and technical committee. The following Corrigendum/Addendum is issued for inclusion in bid document & Technical Specification of items (Section-V, Schedule of Supply, Point no. 3) as below:-

1. Amended Technical Specifications of Single, 3 & 12 Channel ECG Machine

Specification for Single Channel ECG Machine

(Fully automatic, microprocessor controlled, Single-channel resting electrocardiograph, mains and battery operated)
Standard unit including :

1. 12 lead simultaneous ECG data acquisition
2. High-resolution built-in thermal printer
3. Sensitivity : 2.5, 5, 10, 20 mm/mV
4. Should have Membrane keypad for operating
5. Printout on 50 mm or more Z fold paper or roll type paper
6. Battery should last minimum up to 200 recordings
7. ECG measurements, Copy Function
8. Auto & Manual Mode
9. Printing Mode: Auto simultaneous acquisition, with parameter calculation and Manual Real Time Should be supplied with USB & Software for storage of ECG on Computer.
10. Guarantee: Three years on equipment & minimum 06 months on Consumables.
11. CMC: CMC shall be given @ 5 % of net rate (inclusive of Excise Duty & exclusive of VAT/CST etc.) plus service tax (as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after Guarantee period of three years.
12. Equipment should be supplied with compatible standard paper rolls (quantity 10 nos.), additional Clamp Electrodes (1 set), Chest Electrodes (1 set.), ECG Jelly (250 ml X 2 bottles), Earth Cable-1, Guarantee Card & Operating Manual.
13. Consumable: Battery, Lead Set, Chest Electrodes, Limb Electrodes, ECG Jelly and ECG Paper. Rates of Battery, Lead Set, Chest Electrodes and Limb Electrodes should be provided by bidder in BOQ (BF-IV).
14. The rates of consumables and spare parts should be provided and that will be freezed for minimum three years.
15. The machine should be supplied with a good quality carry bag to carry ECG Machine along with ECG Cables, Power Cable & ECG Jelly etc.
16. Details of Service Network of Firm in Rajasthan.
17. The company should mention the model name/number and submit the technical brochure of the quoted model in the Technical Bid.

32

Specification for 3 Channel ECG Machine

1. 3 channel ECG Machine with LCD display, size 4" or more
2. Automatic, Manual Recording Modes.
3. Soft touch key Board.
4. Paper size: 60 mm and more.
5. Mains and Rechargeable battery operated.
6. Inbuilt Rechargeable battery for at least 2 hours.
7. Portable, lightweight < 3 kg.
8. Measurement facility with interpretation.
9. Memory for 25 ECG or more.
10. Automatic Lead Marking, Thermal printing.
11. Recording Sensitivity: 2.5, 5, 10, 20 mm/mV.
12. Defibrillation Protection: Internal.
13. Frequency response: 0.05-150 Hz.
14. Filters: Mains and Muscle interference.
15. Safety standards : Class II, type CF.
16. European CE or USFDA approved.
17. Guarantee: Three years on equipment from the date of installation & minimum 06 months on Consumables.
18. CMC: CMC shall be given @ 5 % of net rate (inclusive of Excise Duty & exclusive of VAT/CST etc.) plus service tax (as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after Guarantee period of three years.
19. Equipment should be supplied with compatible standard paper rolls (quantity 10 nos.), additional Clamp Electrodes (1 set.), Chest Electrodes (1 set.), ECG Jelly (250 ml X 2 bottles), Guarantee Card & Operating Manual.
20. Consumable: Battery, Lead Set, Chest Electrodes, Limb Electrodes, ECG Jelly and ECG Paper. Rates of Battery, Lead Set, Chest Electrodes and Limb Electrodes should be provided by bidder in BOQ (BF-IV).
21. The rates of consumables and spare parts should be provided and that will be freezed for minimum three years.
22. The machine should be supplied with a good quality carry bag to carry ECG Machine along with ECG Cables, Power Cable & ECG Jelly etc.
23. Detail of Service Network of Firm in Rajasthan.
24. The company should mention the model name/number and submit the technical brochure of the quoted model in the Technical Bid.

Specification for 12 Channel ECG Machine

1. 12 Channel ECG Machine.
2. In-built Thermal Printer.
3. Automatic, Manual & Pre Programmed Recording Modes.
4. 3/6/12 Channel Print Format on A-4 Size Thermal Paper.
5. 12 Lead Monitoring Simultaneously on LCD display, size 7" or more.
6. Full QWERTY Alpha Numeric Key Board or Touch Screen Keyboard.
7. Mains/ Battery Operated.
8. In-built Rechargeable Battery for at least 100 recordings.
9. Memory for 50 ECG or more.
10. Time & Date indication.
11. Safety standards : Class 1, Type CF.
12. Paper Speed 5/25/ 50 mm/ sec.
13. Built in Defibrillation Protection.
14. Sensitivity: 2.5, 5, 10, 20 mm/ mV.
15. Measurement facility with interpretation.
16. European CE or USFDA approved.
17. Guarantee: Three years on equipment from the date of installation & minimum 06 months on Consumables.
18. CMC: CMC shall be given @ 5 % of net rate (inclusive of Excise Duty & exclusive of VAT/CST etc.) plus service

tax (as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after Guarantee period of three years.

19. Equipment should be supplied with compatible standard paper rolls (quantity 10 nos.), additional Clamp Electrodes (1 set.), Chest Electrodes (1 set.), ECG Jelly (250 ml X 2 bottles), Guarantee Card & Operating Manual.
20. Consumable: Battery, Lead Set, Chest Electrodes, Limb Electrodes, ECG Jelly and ECG Paper. Rates of Battery, Lead Set, Chest Electrodes and Limb Electrodes should be provided by bidder in BOQ (BF-IV).
21. The rates of consumables and spare parts should be provided and that will be frozen for minimum three years.
22. The machine should be supplied with a good quality carry bag to carry ECG Machine along with ECG Cables, Power Cable & ECG Jelly etc.
23. Detail of Service Network of Firm in Rajasthan.
24. The company should mention the model name/number and submit the technical brochure of the quoted model in the Technical Bid.

(i) Bidders are requested to send with bid, printed descriptive literature of the quoted items.


(ii) If Bidder supplied to or have rate contract of quoted items with any other Govt. institutions within one year, he should provide copies of purchase orders, invoices and rate contract, if asked for.

2. Section IV technical bid submission sheet (Cover-A) point no. 12 (v) is deleted and added at Section III.
3. Section III, evaluation and qualification criteria, point no. 2, clause no.8 Tax clearance certificates:-The VAT/Sales Tax and other taxes clearance or declaration is to be submitted by the bidder. Bidders shall have to submit a valid 'VAT' clearance certificate from the concerned Commercial Taxes Officer or Affidavit and the 'PAN' issued by Income Tax Department.
4. Section II Bid Data Sheet (BDS) point no. 6.3 is read as " The Performance security shall be required as per GCC-10 (iii) @ 5% of the value of the indicative quantity in the bid subject to a maximum Rs. 10.00 lacs for each item in favor of MD, Rajasthan Medical Services Corporation Ltd., Jaipur."
5. It is clarified that Information of award of contract shall be communicated to all participating bidders on the website www.rmssc.nic.in and sppp.raj.nic.in. Please note that individual bidder will not be intimated.
6. It is also further clarified that, if any amendment/clarification is carried out in the technical specifications and bid terms & conditions following pre-bid meeting or any other information, the same will also be uploaded on the Corporation website www.rmssc.nic.in, sppp.raj.nic.in and <https://eproc.rajasthan.gov.in> and will not be published in news papers. It will not be intimated to individual bidder. In case any inconvenience is felt, please contact over telephone number i.e. 0141-2223887 or queries may be e-mailed on address "edepmrssc-rj@nic.in".

Please note that all clarification/amendment/corrigendum in technical specifications/bid conditions is the integral part of the bid document. This corrigendum/ addendum should be signed and annexed with bid document.

All other terms & conditions remains the same.

This bears the approval of M.D., RMSCL, Jaipur.



Executive Director (EPM)
RMSCL, Jaipur

