



Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065

E-mail : edepmrmssc-rj@nic.in

CIN:U24232RJ2011SGC035067

Website:www.rmsc.nic.in

No. F-8 () RMSC/EPM/M-5/15-16/NIB-76/ 756

Dated: 29/7/15

CLARIFICATION

Subject :- Clarification Technical Specifications, date extension and other terms and conditions of Bid document for CR System - Single Plate Reader Type for Digital Radiography, Sonography Color Dopler System & B/W Ultrasound System under NIB No. F-8 () RMSC /EPM/M-5/15-16/NIB-76/10168 Dated 10.04.2015

In equipment **Sonography Color Dopler System** at technical specifications point no. 20 shall be read as:

The system should have the following documentation devices.

- Laser color printer for color image printing of latest technology of Dye sublimation technology with features of water proof prints & image persist for atleast 5 years with multiple print formats as per user requirements.
- Sony / Fuji make B/W Thermal printer of latest model.
- Thermal paper Rolls 50 no. & 5 rim of Glossy paper sheet.

In Reference to above cited subject and NIB-76, the last date for sale of bid, receipt of bid and date of opening of technical bid is hereby extended as below:-

Existing Date			Extended date		
Last Date for Sale of Bid Form	Last Date of Receipt of Bid Form	Date of Opening of Technical Bid	Last Date for Sale of Bid Form	Last Date of Receipt of Bid Form	Date of Opening of Technical Bid
27-07-2015 11:00 AM	27-07-2015 01:00PM	27-07-2015 03:00 PM	10-08-2015 11:00 AM	10-08-2015 01:00PM	10-08-2015 03:00 PM

Note: Please note that all above clarification/amendments/corrigendum in technical specifications /bid conditions is the integral part of the bid document. This clarification/corrigendum/addendum should be signed and annexed with bid document.

All other terms and conditions of bid shall remain the same.

This bears the approval of M.D., RMSCL, Jaipur

Executive Director (EPM)
RMSCL, Jaipur