

RAJASTHAN MEDICAL SERVICES CORPORATION LTD.

D- Block, Swasthya Bhawan, Tilak Marg, C-Scheme JAIPUR (Raj)



Ph. No. 0141-2223887, Fax No. 0141-2228065

CIN: U24232RJ2011SGC035067

E-Mail - edepmrmisc-ri@nic.in

No. F-8(36) RMSC/EPM/M-1/R. C. 15-16/NIB-149/2016/ 974

Dated: 2/2/16

CLARIFICATION/CORRIGENDUM/ADDENDUM

Subject: -Revised Bid Schedule & Amended Technical Specifications of Bid document of item Electrolyte Analyzer, under NIB No. F-8(36) RMSC/EPM/M-1/R. C. 15-16/NIB-149/2016/ 474 Dated:03.02.16

In reference to subject cited above and NIB-149, the representations received from the firm and issues raised by the Bidders, are examined by the competent Authorities and technical committee. The following Clarification/Corrigendum/Addendum is issued for inclusion in Bid document & Technical Specification of items as below:-

1. Amended Technical Specifications of Items Electrolyte Analyzer as below:-

<u>Amended Technical Specifications of Item Electrolyte Analyzer</u>			
1.	Analyzer should measure Na+, K+, CL status in blood, serum, plasma, urine.		
2.	Analyzer must be compact and single unit.		
3.	Analyzer should accept sample from sample, tube/cup.		
4.	Printer should give quality control data, daily and weekly summaries, calibration.		
5.	Calibration must be automatic as well as on demand.		
6.	Please specify automatic calibration/wash cycle time during sleep/standby mode.		
7.	Analyzer must store reasonable number of quality control and patients results.		
8.	Interval tubing must be minimum and should be service friendly, any accidental bubble in the system should easily be dislodged.		
9.	Service engineer should be Rajasthan based & available within 48 hrs.		
10.	The life of electrodes should be at least 1 year.		
11.	Comprehensive guarantee period (with spare parts) for a period of three years from the date of installation and demonstration of equipment which excludes consumable.		
12.	Instrument should be based on ISE (Direct Potentiometer).		
13.	Equipment should be USFDA & European CE Approved. (Valid documentation should be submitted in technical bid).		
14.	Company should quote Comprehensive Maintenance Charges (CMC) for 5 years in BOQ.		
15.	Four Preventive Maintenance services annually during Guarantee and CMC period are essential.		
16.	Firm will be supply suitable rating UPS of minimum half an hour backup along with machine.		
17.	The company should submit technical compliance sheet as per technical specifications mentioning the make & model of quoted product along with catalogue in the Technical bid.		
18.	Installation will be done by supplier free of cost.		
19.	Demonstration of the quoted equipment is must for final technical approval.		

(02) Revised bid schedule:-

E-bids are invited as per following revised time schedule:-

Bid Submission Start	Last date & time for sale of bid form	Last date & time of receipt of bid form	Date & time of opening of technical bid
<u>02.03.2016</u> 11.00 AM	<u>16.03.2016</u> 01.00 PM	<u>16.03.2016</u> 01.00 PM	<u>16.03.2016</u> 03.30 PM

It is also clarified that information of award of contract shall be communicated to all participating bidders on the website www.rmssc.nic.in and sppp.raj.nic.in. Please note that individual bidder will not be intimated."

Note: Please note that all above amendments/corrigendum in technical specifications/bid conditions is the integral part of (Section-V, Schedule of Supply, and Point no. 3) and the bid document. This corrigendum/ addendum should be signed and annexed with bid document. All other terms & conditions remains the same. This bears the approval of M.D., RMSCL, Jaipur.

(Signature)

Executive Director (EPM)
RMSCL, Jaipur