

RAJASTHAN MEDICAL SERVICES CORPORATION LTD.

D- Block, Swasthya Bhawan, Tilak Marg, C-Scheme JAIPUR (Raj)



 CIN:U24232RJ2011SGC035067

E-Mail -edepmrmsc-rj@nic.in

Dated: 14/9115

CLARIFICATION/CORRIGENDUM/ADDENDUM

Subject: -Revised technical Specifications and other terms and conditions of Bid document BIPAP MASK VENTILATOR WITH WELLCUSHIONED FACE MASK WITH HARNESS under NIT No. F-8()RMSC/EPM/M-3/15-16/NIB-90/992 Dated: 11.8.2015.

In Reference to subject cited above and NIB-90, the various representations received from the firms and issues raised by the Bidders are examined by the competent Authorities and technical committee. The following Corrigendum/Addendum is issued for inclusion in bid document & Technical Specification of items (Section-V, Schedule of Supply, Point no. 3) as below:-

Amended Technical Specifications of BIPAP Ventilator

- 1. Non-invasive ventilation support system to ensure proper ventilation to COPD patients.
- 2. It should be suitable for Adult and Paediatric patients.
- 3. Modes of operation: CPAP, Spontaneous, S/T, Time Mode & Pressure Control.
- 4. It should have display to monitoring facility for Tidal Volume, RR, Minute ventilation, I: E ratio.
- 5. Operating pressure range of IPAP, EPAP & CPAP: 4 to 30 cm H₂O
- 6. Unit should be capable of delivering Tidal Volume up to 1500 ml when operated in AVAPS or equivalent mode.
- 7. Breath rate should be able to set up to 30 BPM with Ti of 0.3 to 3 sec. and rise time 100ms to 600ms.
- 8. Sensitivity settings: should have trigger and cycle setting.
- 9. Unit should have integrated alarm for power, Pressure (High/Low), Low Minute ventilation, Apnea, Respiratory rate (High/Low), Circuit disconnected etc. With digital technology for improved leak compensation during inspiration & exhalation.
- 10. Complete unit should be suppled in good quality bag which can carry at least one patient circuit and one mask along with equipment.
- 11. Unit should be European CE or FDA approved & manufacturer / supplier should be ISO 9001 certified.
- 12. Unit should be supplied with reusable patient circuit and reusable full face mask with harness Adult & Paediatric (1 no. each), operational / training manual (1 no.).

- 13. The company should mention the make & model name/number of the quoted equipment and submit the technical brochure of the quoted model in the Technical bid along with compliance sheet as per technical specifications.
- 14. Installation will be done by supplier free of cost.
- 15. Demonstration of Equipment is must for final approval.
- 16. Company should have Service Network in Rajasthan.
- 17. Guarantee: Three years from the date of installation.
- 18. CMC: CMC shall be given @ 5% of net rate (inclusive of Excise Duty & exclusive of VAT/CST etc.) plus service tax (as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after Guarantee period of three years.

Please note that all clarification/amendment/corrigendum in technical specifications/bid conditions is the integral part of the bid document. This corrigendum/ addendum should be signed and annexed with bid document.

All other terms & conditions remains the same.

This bears the approval of M.D., RMSCL, Jaipur.

am and

Executive Director (EPM) RMSC, Jaipur