

राजस्थान चिकित्सा सेवा निगम

स्वास्थ्य भवन, सी-स्कीम, जयपुर

फोन नः 0141-2223887, फैक्स नः 0141-2228065

ई-मेल : rmsc@nic.in

क्रमांक : F-8 (M-2)/RMSC/EPM/ /12-14/NIT-10/12/251

Date: 28.12.12

CORRIGENDUM

Subject: - CORRIGENDUM for NIT No. F-8 (M-1/)/RMSC/EPM/12-14/NIT-10/12/135 dated 16.11.2012 and Amended vide corrigendum No. F-8 (M-1/) RMSC/ EPM/12-14/NIT-10/12/229 dated 20.12.2012

In reference to above NIT and corrigendum the Technical Specification of Items: - CR SYSTEM- SINGLE PLATE READER TYPE FOR DIGITAL RADIOGRAPHY & DENTAL CHAIR WITH DENTAL X-RAY MACHINE is here by amended as below:-

S. No.	Existing Technical Specification	Amended Technical Specification
1.	<p><u>Item No. 2 point no. 10</u></p> <p>Please provide cassette for CR with PSP Plate (IP)</p> <p>14" X 17" } 11" X 14"/ 10" X 14" } Two cassette of each size should be supplied 10" X 12" } 8" X 10" }</p>	<p><u>Item No. 2 point no. 10</u></p> <p>Please provide cassette for CR with PSP Plate (IP)</p> <p>14" X 17" } 14" X 14" } Two cassette of each size should be supplied 10" X 12" } 8" X 10" }</p>
2	<p><u>Item No. 8- Dental Chair with Dental X-Ray Machine</u></p> <p><u>Suction Apparatus (additional apart from main unit)</u></p> <p>Should motorized suction with auto flush and auto drain system, be made up of steel epoxy coated fitted with oil immersed motorized vacuum pump with two graduated suction jars each of 1.5 Litre capacity with synthetic rubber Lids, long pressure tubing, vacuum control, automatic overflow cut off device, should be mounted on rubber wheels.</p>	<p><u>Item No. 8- Dental Chair with Dental X-Ray Machine</u></p> <p><u>Suction Apparatus (additional apart from main unit)</u></p> <p>Motorized suction, should be made up of steel epoxy coated fitted with oil immersed motorized vacuum pump with two graduated suction jars each of 1.5 Litre capacity with synthetic rubber Lids, long pressure tubing, vacuum control, automatic overflow cut off device, should be mounted on rubber wheels.</p>

Executive Director (EPM)
RMSC, Jaipur