

RAJASTHAN MEDICAL SERVICES CORPORATION LTD.

(A Government of Rajasthan undertaking)

GANDHI BLOCK, SWASTHYA BHAWAN, TILAK MARG, JAIPUR

PHONE NO: 141-2223887, Fax No. 0141-5111040, CIN:U24232RJ2011SGC035067

E-mail Id- edepmrmisc-rj@nic.in

No. F-8() RMSC/EPM/M-4/15-16/NIB-153/ Sec.E/ 9151

Dated: 4/9/16

CLARIFICATION/CORRIGENDUM/ADDENDUM

Subject:- Amendment under Bid document for the procurement of Operating Microscope, A-Scan and Auto Ref-Keratometer under NIB No. F-8(RM)RMSC/EPM/M-4/15-16/NIB-153/Sec.E/1004 Dated: 02.03.2016.

1. In reference to above cited subject and NIB No.-153, **Auto Ref-Keratometer shall be read as Auto Refractometer.**
2. In reference to above cited subject and NIB No.-153, The Technical Specification of items (Section-V, Schedule of Supply, Point no. 3) are here by amended as below:-

Technical Specifications of Operating Microscope, A-Scan and Auto Refractometer

(1) Technical Specification for Operating Microscope

1. It should have at least 3 step manual / continuous magnifications.
2. It should have eye piece of 12.5 xs or 10x.
3. The binocular tubes should be 45° inclined and the working distance should be 175mm / 200mm.
4. Foot control motorized focus.
5. It should have a coaxial fiber optic illumination system 15V/150W or 12V/120W halogen lamp with reflector or LED light.
6. The light source should be continuously variable.
7. The floor stand should have good quality lockable castor wheels.
8. The vertical movement should be 300 mm. at least.
9. It should operate on standard power supply of 220V-240V, 50-60 Hz.
10. It should be supplied with suitable online UPS with at least half an hour backup.
11. It should be supplied with standard spare halogen bulbs (6 Nos.), if LED (one spare).
12. The equipment should be European CE with notified body / USFDA.
13. Guarantee: Three Years on equipment from the date of installation.
14. CMC: CMC will be given @ 4 % (of net rate- inclusive of Excise Duty & exclusive of VAT/CST etc.) plus service tax (as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after guarantee period of three years.
15. Installation will be done by supplier free of cost.
16. Firm should mention all the pre-installation requirements in technical bid.
17. Final technical approval after demonstration.
18. The company should mention the make & model name/number of the quoted equipment and submit the technical brochure of the quoted model in the Technical bid along with compliance sheet as per technical specifications.
19. Consumable: halogen bulb or LED light.

2M

cyz

22. Rates of halogen bulb or LED light should be provided by bidder in BOQ in e-proc. portal.

Other Terms and Conditions

1. The company should submit technical compliance sheet as per technical specifications mentioning the make & model of quoted item along with catalogue in the Technical bid.
2. Installation will be done by supplier free of cost at consignee place.
3. Demonstration of equipment is must for technical approval.
4. In Instruments, wherever S. S. Used should be 410/420 garde.

(2) Technical Specification for A-Scan

1. Contact scan facility
2. Probe : at least 10 MHz with fixation light
3. Modes : Normal, cataract, dense cataract, aphakia, pseudophakia
4. Auto sensing and freezing
5. Multiple reading detection and standard deviation display
6. Inbuilt IOL formula : SRK-II, SRK -T (regression), Holladay, Hoffer, post-refractive surgery formula etc.
7. Accuracy : Clinival – error of 0.1 mm or less ; electronic : error of 0.05 mm or less
8. Programmable user profiles
9. Standard power supply (220-240 volts 50-60 HZ)
10. Compatible CVT
11. Standard accessories (Extra paper rolls (6), calibration block)
12. The machine should have safety certificate CE with notified Body/ FDA (US). Copy of the certificate shall be produced along with the technical bid.
13. Guarantee shall be three years from the date of successful installation at consignee place.
14. CMC will be given @ 4 % (of net rate - inclusive of Excise Duty & exclusive of VAT / CST etc.) plus service tax (as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after completion of Guarantee period of three years.

Other Terms and Conditions

1. The company should submit technical compliance sheet as per technical specifications mentioning the make & model of quoted item along with catalogue in the Technical bid.
2. Installation will be done by supplier free of cost at consignee place.
3. Demonstration of equipment is must for technical approval.
4. In Instruments, wherever S. S. Used should be 410/420 garde.

(3) Technical Specification for Auto Refractometer

1. Measurement –Spherical -20 to +20 d (min 0.25 D steps)
 - i. Cylinder at least -7 to + 7 D (min 0.25 D steps)
 - ii. Axis 0 to 180 degrees (min 1 degrees steps)
 - iii. Minimum pupil size : 3.0 mm or less
 - iv. Vertex distance adjustable
 - v. Retroillumination mode

2. LED/LCD Display
3. Printer with standard thermal paper
4. Standard 220-240 V supply (50-60 Hz)
5. Standard accessories.
6. Spare thermal paper rolls (20) and spare bulbs (4)
7. Motorized table
8. Compatible online UPS.
9. The product should have safety certificate from a competent authority CE / FDA (US). Copy of the certificate / test report shall be submitted in technical bid.
10. Guarantee shall be three years from the date of successful installation at consignee place.
11. CMC will be given @ 4 % (of net rate - inclusive of Excise Duty & exclusive of VAT / CST etc.) plus service tax (as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after completion of Guarantee period of three years.

Other Terms and Conditions


1. The company should submit technical compliance sheet as per technical specifications mentioning the make & model of quoted item along with catalogue in the Technical bid.
2. Installation will be done by supplier free of cost at consignee place.
3. Demonstration of equipment is must for technical approval.
4. In Instruments, wherever S. S. Used should be 410/420 garde.

3. E-tenders are invited as per following time schedule:-

Last Date for Sale of Bid Form	Last Date of Receipt of Bid Form	Date of Opening of Technical Bid
27.05.2016 Up to 11.00 A.M.	27.05.2016 Up to 1.00 P.M.	27.05.2016 From 3.00 P.M.

Note: This corrigendum/addendum should be signed and annexed with bid document.

All other terms and conditions of bid shall remain the same.
This bears an approval of M.D., RMSCL, Jaipur.


Executive Director (EPM)
RMSCL, Jaipur

Dated:

No. F-8() RMSC/EPM/M-4/15-16/NIB-153/ Sec.E/

Copy to:-

1. M.D.RISL, Yojna Bhawan, Jaipur
2. PA to M.D., RMSC, Jaipur
3. AGM (IT), RMSC, Jaipur
4. Guard File


Executive Director (EPM)
RMSCL, Jaipur

