

Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065

E-mail : edepmrmse-rj@nic.in

CIN:U24232RJ2011SGC035067

Website: www.rmssc.nic.in

No. F-8(3) RMSC/EPM/M-2/NIB-103/15-16/1626

Dated: 9/10/15

CLARIFICATION/CORRIGENDUM/ADDENDUM

Sub:- Revised technical Specifications and revised bid submission date of item labour Table under NIB No. F-8() RMSC/EPM/M-2/15-16/NIB-103/997 dated 11.08.15

In Reference to subject cited above and NIB-103, the various representations received from the firms and issues raised by the Bidders are examined by the competent Authorities and technical committee. The following Clarification/Corrigendum/Addendum is issued for inclusion in Bid document & Technical Specification of items as below:-

A. Amended Technical Specifications of item Labour Table:-

1. Smoothly finished complete stainless steel labour table. (Stainless Steel grade 304).
2. Should have three section top with rexine covered minimum 25mm thick sectional mattresses. The density of mattress foam should not be less than 40.
3. Should have a dimension of L 180 X W 90 X H 80cm. (Tolerance±5%).
4. Should be able to hold minimum of 150 Kg.
5. Back rest should be adjustable on ratchet.
6. Removable SS Bowls under the U cut, size proportionate to the U cut size provided with SS head end side railing made from solid SS rods, SS head grips.
7. Leg section should be telescoped under the mainframe completely for lithotomy positions.
8. Should have S.S. head & side railing steel hand grips, a pair of rubber padded stainless steel lithotomy rods and lithotomy crutches with 2 pairs of rexine ankle straps.
9. Should provide a detachable saline stand having a provision for fixing four positions (Head end and middle position) on the table 2 Nos. of saline rod of SS 304 grade should be provided.
10. Base should be mounted on four heavy duty PVC/rubber shoes.
11. Guarantee: Three years from the date of Installation.
12. CMC: CMC will be given @5% (of net rate-inclusive of Excise Duty & exclusive of VAT/CST etc.) plus service tax (as applicable) and yearly escalation of 5% on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after completion of guarantee period of three years.
13. Manufacturer should be ISO 13485 certified. (Valid documentation should be submitted in technical bid).
14. Product should be CE certified. (Valid documentation should be submitted in technical bid).
15. Final technical approval after demonstration.
16. Company should have Service Network in Rajasthan.
17. Installation will be done by supplier free of cost.
18. The company should submit technical compliance sheet as per technical specifications mentioning the make & model of quoted item along with catalogue in the technical bid.

B. The bid is re-scheduled as follows:

Existing Dates			Extended Dates		
Last Date for Sale of Bid Form	Last Date of Receipt of Bid Form	Date of Opening of Technical Bid	Last Date for Sale of Bid Form	Last Date of Receipt of Bid Form	Date of Opening of Technical Bid
1	2	3	4	5	6
up to 11:00 A.M. 20.10.2015	up to 1.00 P.M. 20.10.2015	from 3.00 P.M. 20.10.2015	up to 11:00 A.M. 26.10.2015	up to 1.00 P.M. 26.10.2015	from 3.00 P.M. 26.10.2015

This corrigendum shall be signed and annexed with bid document. This bears an approval of Managing Director, Rajasthan Medical Services Corporation Limited, Jaipur.


Executive Director (EPM)
RMSCL, Jaipur