

Government of Rajasthan
Medical & Health Department

Rajasthan Medical Services Corporation, Swasthaya Bhawan, C-Scheme, Jaipur

Phone No: 0141-2228065, Fax No: 0141-2228065

E_mail : rmse@nic.in

F. 02 () / RMSCL/PROCUREMENT/S&S/NIT-4/2013/ 294

Dated: 22.11.2013

CORRIGENDUM

Subject: - Amendment to Bid for annual rate contract cum supply and empanelment for supply of surgical and consumables (Non Drug Items)

Ref: - NIB No. F. 02 () / RMSCL/PROCUREMENT/S&S/NIT-4/2013/249

Dated 04.10.2013

The following amendments are made:

1. The date of bid submission is extended to 03.12.2013 up to 1.00 PM
Accordingly dates are revised as follows:
Last date and time for
Downloading the Bid document 02.12.2013 up to 6.00 PM
Date and Time of opening of
online technical bids 03.12.2013 at 2.30 PM

2. Quantity of items given in the bid (Annexure VII) and BOQ are to be read as follows :-

Annexure- VII

Clause 5.1 (b),10.2,10.3

SURGICAL ITEMS LIST with SPECIFICATIONS

S.No.	Code No.	Name of Item	Pack Size	Quantity
1	S-3	Asepto Syringe with Transparent Bulb Sterile, 60 ml	Piece	2086
2	S-40	Urine Collecting Bag, Disposable, 2000 ml <ul style="list-style-type: none"> • Transparent Sheet • Kink resistant flexible tubing not less than 90 cm in length • Should have non-return valve • Top drainage outlet • Graduated Bag 	Unit	504699

		<ul style="list-style-type: none"> Moulded handle for easy handling 		
3	S-48	Corrugated Drainage Sheet, Sterile, Multichannel, with radio Opaque line, Single Use – All sizes	Each piece	3000
4	S-81	Temporary Cardiac Pacing Wire (Electrode) Sterile ½ Cir, Tapercut, 26 mm needle with self retaining system	Sachet	9886
5	S-82	<p>(a) Skin Grafting Knife Blade (Sterile) made of carbon steel or stainless steel material 158 mm long individually wrapped in wrapper corrosion inhibitor paper in single packet, in packs of 10. The edge must be sharp enough to cut the skin in a single shave and should snugly fit in the handle</p> <p>(b) Skin Grafting Knife Handle (Watson modification of Humby's Knife) Stainless Steel, CE certified, in which the blade specified in (a) above should fit snugly.</p>	One pack each	2554
6	S-85	<p>Face Mask, Disposable</p> <ul style="list-style-type: none"> Should be manufactured from non woven poly prop fabric Should be 3 ply construction Should have high Bacterial filtration efficiency Should be heat sealed to keep 3 layers together Standard Size 17.5 x 9 cm Colour Green / Blue <p>There should be a string each at all four corners, length of string should be 40 cm</p>	Piece	1576721
7	S-86(a)	<p>Surgical Cap, Disposable (for Surgeons)</p> <ul style="list-style-type: none"> Should be manufactured from non woven fabric Elastic strip stitched on the back for proper grip on the forehead Green colour Ultrasonically stitched Air permeable / breathable Should retain skin and hair particles 	Piece	600000
8	S-86(b)	<p>Surgical Cap, Disposable (for Nurses)</p> <ul style="list-style-type: none"> Should be manufactured from non woven fabric Blue / Green colour Round upon wearing, with elastic Air permeable / breathable Should retain skin and hair particles 	Piece	664267
9	S-89	<p>Disposable Sterile Surgical Rubber Gloves Size 8 Inches</p> <ul style="list-style-type: none"> Made of natural rubber Latex, powdered, 	Pair	776417

		without tear, properly folded in a paper <ul style="list-style-type: none"> • Free of holes, with Acceptable Quality Level (AQL) of 1.5 or less • Tensile strength as per EN 455-2 • Powder should be non-allergenic • Conforms to IS 13422 • ISI marked / CE certified / FDA approved 		
10	S-90(a)	Latex examination gloves made of natural rubber latex. Non-Sterile, Ambidextous, AQL 1.5 Micro Textured Powdered with absorbable dusting powder USP. Extra Small	Dispenser Box of 100 Gloves	717145
11	S-30(a)	Surgical Blade Sterile, Size-11	100 Blades/ Packet	222729
12	S-30(b)	Surgical Blade Sterile, Size-15	100 Blades/ Packet	178660
13	S-30(c)	Surgical Blade Sterile, Size-22	100 Blades/ Packet	434972

39
22/11/20
ED (P)