


SPECIFICATIONS OF ECG MONITOR

1. Should have at least 5" Color TFT Display
2. Should have ECG as standard parameter
3. Should monitor Lead I, II, III, aVr, aVl, aVf & V with help of 5 lead ECG cable
4. Should have built-in rechargeable battery
5. Should have battery backup of at least 90 minutes
6. Should have 24 hours graphical and tabular trends
7. Should have arrhythmia detection
8. Should have facility to store ECG waveforms
9. Should be CE Europe or USFDA approved
10. Optional : Quote for SpO2 & NIBP parameter



Executive Director (EPM)
Rajasthan Medical Services Corporation Ltd.,
Jaipur