



RAJASTHAN MEDICAL SERVICES CORPORATION LTD.

D- Block, Swasthya Bhawan, Tilak Marg, C-Scheme JAIPUR (Raj)



Ph. No. 0141-2223887, Fax No. 0141-2228065

CIN:U24232RJ2011SGC035067

E-Mail -edepmrmssc-ri@nic.in

No. F-8()RMSC/EPM/M-3/15-16/NIB-91/ 1862

Dated: 14/9/15

CLARIFICATION/CORRIGENDUM/ADDENDUM

**Subject: -Revised technical Specifications and other terms and conditions of Bid document
12 Channel ECG Machine under NIT No. F-8()RMSC/EPM/M-3/15-16/NIB-91/991
Dated: 11.8.2015.**

In Reference to subject cited above and NIB-91, the various representations received from the firms and issues raised by the Bidders are examined by the competent Authorities and technical committee. The following Corrigendum/Addendum is issued for inclusion in bid document & Technical Specification of items (Section-V, Schedule of Supply, Point no. 3) as below:-

Amended Technical Specifications for 12 Channel ECG Machine


1. 12 Channel ECG Machine.
2. In-built Thermal Printer.
3. Automatic & Manual Recording Modes.
4. 3/6/12 Channel Print Format on A-4 Size Thermal Paper.
5. 12 Lead Monitoring Simultaneously on LCD display, size 5" or more.
6. Full QWERTY Alpha Numeric Key Board or Touch Screen Keyboard.
7. Mains & Battery Operated.
8. In-built Rechargeable Battery for at least 100 recordings.
9. Memory for 30 ECG or more.
10. Paper Speed 12.5/25/ 50 mm/ sec.
11. Built in Defibrillation Protection.
12. Sensitivity: 2.5, 5, 10, 20 mm/ mV.
13. Measurement facility with interpretation.
14. Product should be European CE or USFDA approved. (Valid certificate should be submitted in technical bid).
15. Guarantee: Three years on equipment from the date of installation & minimum 06 months on Consumables.
16. CMC: CMC shall be given @ 5 % of net rate (inclusive of Excise Duty & exclusive of VAT/CST etc.) plus service tax (as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after Guarantee period of three years.
17. The machine should be supplied with a good quality carry bag to carry ECG Machine along with compatible standard paper rolls (quantity 10 nos.), Complete ECG Lead Set (1 No.), Clamp Electrodes (1Set), Chest bulb Electrodes (1Set), ECG Jelly (250 ml X 2 bottles), Guarantee Card & Operating Manual.

18. Consumable: Battery, Complete Lead Set, Chest bulb Electrodes, Limb Electrodes, ECG Jelly and ECG Paper. Rates of Battery, Complete Lead Set, Chest bulb Electrodes and Limb Electrodes, ECG Jelly, ECG Paper should be provided by bidder in BOQ.
19. The company should submit technical compliance sheet along with catalogue as per amended technical specifications mentioning the make & model of quoted item in the Technical bid .
20. Installation will be done by the firm free of cost.
21. Service Network of Firm in Rajasthan is essential.

Please note that all clarification/amendment/corrigendum in technical specifications/bid conditions is the integral part of the bid document. This corrigendum/ addendum should be signed and annexed with bid document.

All other terms & conditions remains the same.

This bears the approval of M.D., RMSCL, Jaipur.


Executive Director (EPM)
RMSC, Jaipur