


SPECIFICATIONS OF ICE PACKS

1. Ice packs suitable for vaccine carriers and cold boxes in Health Department.
 2. It should be HDPE material and suitable for -15°C to 25°C.
 3. It should have volume of 0.30 liters to 0.40 liters and external dimension 161X91X33mm (±2mm)
 4. It should be robust and pass the leakage test.
 5. Its empty weight should be 75 to 80 grams approx.
 6. It should have removable caps and cork for water filling and have a filling line mark clearly indicated.
 7. It should have 2 holes for keeping vaccine vials.
 8. It should confirm to WHO-PQS performance / UNICEF specification.
 9. The manufacture must provide a certificate of compliance of WHO test procedures by and independent testing laboratory, accredited for type testing of vaccine carriers by WHO/UNISEF/STQC/NABL.
- Supplier should provide two years guarantee as per specification.



Executive Director (EPM)
Rajasthan Medical Services Corporation Ltd.,
Jaipur

2