

List of approved and not approved reagents kit for tender in RMSC (2013)

S.No.	Code	Item	Approved	Not approved	Reason
1	LR001	Multiparameter Urine Strips	1. Span 2. Piramal	1. Sd Bio standard	Glucose sensitivity is less and leucocyte sensitivity is less
				2. Transasia	Certificate of sensitivity not provided in the browser
2	LR002	Urine Strip albumin	1. Piramal	1. Agappe	Glucose sensitivity is less or not in desired unit asked in technical specification
				2. Transasia	Certificate of sensitivity not provided in the brochure
				3. Sd BioStandard	Glucose sensitivity is less
3	LR006	Leishman's stain	1. Span 2. Sd Fine Chem		
5	LR007	Giemsa Stain	1. Span	1. Sd Fine Chem	Wrong Chemical composition
6	LR008	Not Quoted			
7	LR009	Anti-A (Monoclonal)	1. Tulip 2. Span	1. Agappe	Brochure stating the technique is wrong
8	LR010	Anti-B (Monoclonal)	1. Tulip 2. Span	1. Agappe	Brochure stating the technique is wrong
9	LR011	Anti-D (Monoclonal)	1. Tulip 2. Span	1. Agappe	Brochure stating the technique is wrong
10	LR012	Trisodium Citrate	1. Span 2. Sd Fine Chem		
11	LR014	PT Test Kit	1. Transasia 2. Tulip		
12	LR015	APTT Test Kit	1. Transasia 2. Tulip		
13	LR016	Reagent for five part hematology - SYSMEX	1. Transasia (Condition applied)		

14	LR017	Reagent for three part hematology - SYSMEX	1. Transasia (Condition applied) 2. Agappe (Condition		
15	LR043	CSF Diluting Fluid	1. Span		
16	LR044	WBC Diluting Fluid (Turkck's)	1. Span 2. Sd Fine Chem		
17	LR046	TEC Fluid	No Bidder		

Note: 1. M/s Transasia Biomedical Limited usually make reagents for sale counter of its own instrument (SYSMEX) close system.
2. M/S Agappe in our opinion makes reagents for both open and close system. Before using open system the company making the instrument should be informed and consent of using his reagents in open system should be taken with regards to maintenance of the instrument.

Dr. G.C. Meena

Dr. Ranjana
Solanki

Dr. Ajay Yadav