


SPECIFICATIONS OF FETAL DOPPLER

1. Should be complete system with a permanently fixed 2 MHz transducer.
2. Handy and attractive Cabinet.
3. Highly sensitive ultrasonic beam permits hearing the fetal heart beats clearly from 10-12th week of pregnancy. Fetal heart beats can be heard either by inbuilt speaker or earphone.
4. Machine should have battery and mains operation.
5. Should be supplied with rechargeable battery.
6. Unit should be designed as per IEC-601 -1 (certificate to be submitted)
7. Unit should be CE marked and from ISO 13485 certified manufacturers (certificate to be submitted) should submit a test report from national or state Govt. approved laboratory.
8. List of spare parts and accessories to be supplied.
Should be supplied with Ultrasound gel bottles (10 No.)



Executive Director (EPM)
Rajasthan Medical Services Corporation Ltd.,
Jaipur

2