

SPECIFICATIONS OF GLUCOMETER

5 Second test result chemical bio Sensor Technology with Capillary System having No Coding System (Automatic Calibration) with at least 450 test memory. Memory Measuring range Should be 10-500 mg/dl. Should be standard (AAA) alkaline battery/li ion battery operated with LCD size to be at least 50mm (diagonally). With ± 5 variance.

Strip should have shelf of minimum 12 month at time of delivery
Glucometer should be ISI/USFDA/CE certified/ approved product.
One step testing strip should be freely available at least at district headquarters. Pack of 50 test strip price should be quoted along with Glucometer as well as in item no. H/12. A free pack of 50 test strip, pricker & 50 lancets for pricker should be supplied along with Glucometer.

The price of test strip (Each Pkt. Of 50 strips) shall be taken into account while evaluating the price bid. One strip bottle 1x 50 test should be provided with each equipment.

Warrantee for two year.

Final Technical approval after demonstration

Executive Director (EPM)

Rajasthan Medical Services Corporation Ltd.,

Jaipur