


SPECIFICATIONS OF DRY BATH INCUBATOR

- Main Input : 230V/50Hz / 110V/60Hz A.C. Supply
- WxDxH : 168(W)x232 (D)x93(H) \pm 10%
- Body : Metal
- Body Weight : 2.3kg. \pm 10%
- Temp. Controller : Digital Type
- Temp. Range : 1°C to 110°C
- Accuracy : 0.3°C at the given range
- Heating Block : 24 Holes of 12.7 mm Diameter or as selected
- Timer : 1 to 60 min. (digital Type)
- The system should be FDA/CE approved.
- On sight annual calibration and validation.
- **CAMC for five years** after Expiry of Two years Guarantee period



Executive Director (EPM)
Rajasthan Medical Services Corporation Ltd.,
Jaipur

