

SPECIFICATIONS OF FULLY AUTOMATED DELUXE DONOR COUCH

- (i) It should be equipped with following electrically operated functions operated by Remote Control.
- (ii) It should have back section adjustment.
- (iii) It should have leg section adjustment.
- (iv) It should be easily converted into Bed.
- (v) Should have lifting capacity at least 110-120 kg.
- (vi) The frame of the Recliner is constructed of High grade Steel duly epoxy powder-coated.
- (vii) The base is covered with smooth and elegant finish polymer molded which is rust-free, scratch resistant and easy to clean and should have wheels to move the couch.
- (viii) The Recliner is fitted on heavy duty manoeuvring heavy duty 125mm PU Castors with brakes.
- (ix) It should has a three sectional top duly covered with Elegant upholstered high density, max. Comfort Foam Padding.
- (x) Fitted with auto adjusted Arm supports which automatically takes its positions during the adjustment of Backrest.
- (xi) Unit should have accessories- blood Collection Monitors Tray, BP. Apparatus Tray and Tray for other accessories and telescopic IV Stand.
- (xii) It should have one year guarantee.
- (xiii) It should be ISO 9001:2008 TUV/CE Certified.

Executive Director (EPM)
Rajasthan Medical Services Corporation Ltd.,
Jaipur