


**SPECIFICATION OF OPEN NEWBORN CARE SYSTEM ON TROLLEY WITH  
DRAWERS, WITH RADIANT WARMER, OXYGEN-PROVISION**

1. Mobile newborn resuscitation table with fixed-height radiant warmer
  2. Antistatic castors, 2 with breaks
  3. Table surface with mattress with infant head/shoulder support
  4. Mattress-padding: foam density approx. 21 - 25 kg/m<sup>3</sup>
  5. Mattress cover should be sealed and waterproof
  6. Side boards transparent acryl, drop down and lockable
  7. Under table 2 storage drawers
  8. Side rails allow for mounting of accessories
  9. Hood suspended above the table integrates heating element and overhead light
  10. Overhead light: 2 x 50W halogen spot, with dimming function
  11. Integrated support for two 10 L oxygen bottles
  12. Control unit has flow meter and displays pressure
  13. Heating element: with lifelong guarantee
  14. Control unit allows air and skin temperature preset (LED indicator) and drives radiant heater output (servo and manual)
  15. Integrated timer: 1 to 59 min, with count-up and count-down feature
  16. Temperature range, skin: 34 to 38°C (user pre-settable)
  17. Monitoring of skin temperature by means of sensor, range: 30 to 42°C
  18. Heater output: 0 to 100% in increments of 5%
  19. Control unit: audiovisual alarms according to timer and temperature presets avoiding overheating
  20. Display reports systems errors, sensor failure
  21. Power requirement: 220 V/50 Hz
  22. Power consumption: 800 W
  23. Device is produced by ISO 9001 certified manufacturer (Certificate to be submitted, further details see "Technical Provisions")
  24. Device is safety certified according ISO & CE 93/42/ FDA 510k
- Supplied with:**
25. 1 x mattress
  26. 1 x skin temperature probe (including connection cable)
  27. 1 x spare skin temperature probe (including connection cable)
  28. No additional (spare) heating element heating element is required
  29. 2 x empty 10 L oxygen cylinders
  30. 1 x spare set of fuses
  31. User manual with trouble shooting guidance, in English. Technical manual with maintenance and first line technical intervention instructions, in English
  32. List of priced accessories
  33. List of priced spare parts
  34. List with name and address of technical service providers in India
  35. Training and installation at end-user site
  36. Proposal for full service AMC, for five years after completion of guarantee period, covering (i) 2 preventive maintenances per year
  37. on-call technical interventions, spare parts and travel



**Executive Director (EPM)**  
**Rajasthan Medical Services Corporation Ltd.,**  
**Jaipur**