

## Alphabetical List of Surgical & Sutures NIT-1/2013/04 Dated 15/04/2013 (R.C.-VIII)

S. No.	Code No.	Name of approved item (s) with specification	Packing Unit	Name of Approved Firm	Approved Rate exclusive CST /VAT	Percent of excise duty inclusive in approved rate	Rate contract validity up to	Applicable rate of performance guarantee & amount of performance guarantee deposited by the firm/remark	R.C. No.												
1	S-99(a)	<p>Sanitary Napkin, Beltless</p> <p>1. Covering – Covering of the absorbent filler shall be good quality knitted sleeve or non-woven fabric which has sufficient porosity to permit the assembled napkin to meet the absorbency requirements. The napkins shall have a non absorbent barrier on one side which shall have an identifying mark indicating the side of the barrier.</p> <p>2. Absorbent Filler – The filler material shall consist of cellulose pulp/ wadding, and shall be free from lumps, oil spots, dirt or foreign material, etc.</p> <p>3. Back Strip – A back strip for sticking the sanitary napkin onto the underwear should be there using good quality adhesive material.</p> <p>4. Size - The size of absorbent section / complete sanitary napkin shall be as follows:( in mm)</p> <table style="margin-left: 40px; border: none;"> <thead> <tr> <th></th> <th>Absorbent section</th> <th>Total</th> </tr> </thead> <tbody> <tr> <td>Pad Length</td> <td>210 +_ 10</td> <td>230 +_ 10</td> </tr> <tr> <td>Width</td> <td>65 to 75</td> <td>70 to 85</td> </tr> <tr> <td>Thickness</td> <td>8 +_ 2</td> <td></td> </tr> </tbody> </table> <p>5. Weight : Not more than 10 gm</p>		Absorbent section	Total	Pad Length	210 +_ 10	230 +_ 10	Width	65 to 75	70 to 85	Thickness	8 +_ 2		6 napkins per pack	M/S Johnson & Johnson Ltd., 30, Forjett Street, Post Box No. 9301, Mumbai – 400 036	Rs.10.43 (Ten and paisa forty three only) + Tax Nil	2.23	15.01.2015	EMD Rs. 200000 adjusted in SD	VIII
	Absorbent section	Total																			
Pad Length	210 +_ 10	230 +_ 10																			
Width	65 to 75	70 to 85																			
Thickness	8 +_ 2																				

2	S-99(b)	<p>Sanitary Napkin, Belt-type</p> <ol style="list-style-type: none"> <li>1. Covering – Covering of the absorbent filler shall be good quality knitted sleeve or non-woven fabric which has sufficient porosity to permit the assembled napkin to meet the absorbency requirements. The napkins shall have a non absorbent barrier on one side which shall have an identifying mark indicating the side of the barrier.</li> <li>2. Absorbent Filler – The filler material shall consist of cellulose pulp/ wadding, and shall be free from lumps, oil spots, dirt or foreign material, etc.</li> <li>3. Size - The size of absorbent section / complete sanitary napkin shall be as follows:( in mm) <ul style="list-style-type: none"> <li>Absorbent section</li> <li>Total</li> <li>Pad Length 220 +_ 10</li> <li>Width 70+_ 5</li> <li>Thickness 17+_3</li> </ul> </li> <li>4. Weight : 12+_3 gm</li> <li>5. Pack – Six Napkins in a pack</li> <li>6. Elastic belt with loops shall be provided in each pack.</li> <li>7. Absorbency: The napkin should be able to absorb not less than 30ml of normal saline or colored water or test fluid when poured on to the centre of the napkin at the rate of 15ml per minute.</li> </ol>	6 napkins per pack	M/S Johnson & Johnson Ltd, Mumbai – 400 036	Rs.14.88 (Fourteen and paisa eighty eight only) + Tax Nil	4.12	15.01.2015		VIII
---	---------	---	--------------------	---	---	------	------------	--	------